

8-Plex Multi-Pathway Signaling Kit – Phosphoprotein

Cat. # 48-680

MILLIPLEX[®] MAP

8-Plex Multi-Pathway Signaling Kit – Phosphoprotein (Erk/MAP Kinase 1/2, STAT3, JNK, p70 S6 kinase, IκBα, STAT5A/B, CREB, and p38/SAPK)


#48-680

TABLE OF CONTENTS	PAGE
Introduction	2
Principle	3
Storage Conditions Upon Receipt	3
Reagents Supplied	4
Materials Required But Not Provided	5
Safety Precautions	5
Technical Guidelines	6
Sample Collection And Storage	7
Preparation Of Reagents For Immunoassay	8
Immunoassay Protocol	10
Instrument Settings	11
Supplemental Protocols	12
Replacement Reagents	12
Representative Data	13
Troubleshooting Guide	14
Ordering Information	16
Well Map	17

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

INTRODUCTION

Cells respond to their environment in many different ways through intracellular signaling. Signaling through RTKs (Receptor Tyrosine kinases) often promote increased metabolism and cell growth. Activation of ERK/MAP kinase is one of the key Ser/Thr kinases activated via RTK signaling, promoting increased activity of p70 S6 kinase, Msk1, STAT3 (Ser727), CREB, and many other signaling intermediates. Other signaling pathways induced via stress (ex. heat shock/arsenite treatment) or death receptors (ex. TNF α) promote p38, JNK, and I κ B signaling pathways.



The MILLIPLEX[™] MAP 8-plex Multi-Pathway Signaling kit, phosphoprotein, is used to detect changes in phosphorylated ERK/MAP kinase 1/2 (Thr185/Tyr187), STAT3 (Ser727), JNK (Thr183/Tyr185), p70 S6 kinase (Thr412), IκBα (Ser32), STAT5A/B (Tyr694/699), CREB (Ser133), and p38 (Thr180/Tyr182) in cell lysates using the Luminex® system. The detection assay is a rapid, convenient alternative to Western Blotting and immunoprecipitation procedures. Each kit has sufficient reagents for one 96 well plate assay.

PRINCIPLE

MILLIPLEX MAP is based on the Luminex® xMAP® technology — one of the fastest growing and most respected multiplex technologies offering applications throughout the life sciences and capable of performing a variety of bioassays including immunoassays on the surface of fluorescent-coded beads known as microspheres.

- Luminex® uses proprietary techniques to internally color-code microspheres with two fluorescent dyes. Through precise concentrations of these dyes, 100 distinctly colored bead sets can be created, each of which is coated with a specific capture antibody.
- After an analyte from a test sample is captured by the bead, a biotinylated detection antibody is introduced.
- The reaction mixture is then incubated with Streptavidin-PE conjugate, the reporter molecule, to complete the reaction on the surface of each microsphere.
- The microspheres are allowed to pass rapidly through a laser which excites the internal dyes marking the microsphere set. A second laser excites PE, the fluorescent dye on the reporter molecule.
- Finally, high-speed digital-signal processors identify each individual microsphere and quantify the result of its bioassay based on fluorescent reporter signals.

The capability of adding multiple conjugated beads to each sample results in the ability to obtain multiple results from each sample. Open-architecture xMAP® technology enables multiplexing of many types of bioassays reducing time, labor and costs over traditional methods.

This kit is for research purposes only.

Please read entire protocol before use.

It is important to use same assay incubation conditions throughout your study.

By purchasing this product, which contains fluorescently labeled microsphere beads authorized by Luminex Corporation ("Luminex"), you, the customer, acquire the right under Luminex's patent rights, if any, to use this product or any portion of this product, including without limitation the microsphere beads contained herein, only with Luminex's laser based fluorescent analytical test instrumentation marketed under the name of Luminex100, 200, HTS.

STORAGE CONDITIONS UPON RECEIPT

- Recommended storage for kit components is 2 8 °C.
- DO NOT FREEZE Antibody-Immobilized Beads, Detection Antibody, and Streptavidin-Phycoerythrin.

REAGENTS SUPPLIED	CATALOG NUMBER	VOLUME	QUANTITY
MILLIPLEX MAP 8-plex Multi- Pathway-Phosphoprotein, Beads (20x)	42-680	131 μL	1 tube
MILLIPLEX MAP 8-plex Multi- Pathway-Phosphoprotein, Biotin (20x) (Detection Antibody)	44-680	131 μL	1 tube
MILLIPLEX MAP Lysis Buffer	43-040	55 mL	1 bottle
MILLIPLEX MAP Assay Buffer 2	43-041	55 mL	1 bottle
MILLIPLEX MAP HeLa Cell Lysate: HS/Ars	47-211		1 vial
MILLIPLEX MAP HeLa Cell Lysate: TNF α	47-212		1 vial
MILLIPLEX MAP A431 Cell Lysate: EGF	47-210		1 vial
MILLIPLEX MAP HeLa Cell Lysate: Unstimulated	47-205		1 vial
MILLIPLEX MAP Streptavidin- Phycoerythrin	45-001D	115 μL	1 tube
MILLIPLEX MAP Amplification Buffer (1X)	43-024A	3 mL	1 bottle
MILLIPLEX MAP 96-well Filter Plate	MX-PLATE		1 plate
Empty mixing vials			3 vials

Protein	Bead Set
Erk/MAP Kinase 1/2	9
STAT3	14
JNK	16
p70 S6 kinase	19
ΙκΒα	33
STAT5A/B	35
CREB	37
p38	54

MATERIALS REQUIRED BUT NOT PROVIDED

Reagents

- Protease inhibitors (recommend Millipore Catalog #20-201 or similar product)
- Coomassie or BCA-based total protein assay or an assay normalization control, such as the GAPDH MAPmates (catalog #46-667)

Instrumentation / Materials

- Adjustable Pipettes with Tips capable of delivering 25 μL to 1000 μL
- Multichannel Pipettes capable of delivering 25 µL to 200 µL
- Reagent Reservoirs
- Polypropylene Microfuge Tubes
- Rubber Bands
- Absorbent Pads
- Laboratory Vortex Mixer
- Sonicator
- Titer Plate Shaker
- Vacuum Filtration Unit (Millipore Vacuum Manifold Catalog #MSVMHTS00 or equivalent with Millipore Vacuum Pump Catalog #WP6111560 or equivalent)
- Luminex 100[™] IS, 200[™], or HTS by Luminex Corporation
- Plate Stand (Millipore Catalog # MX-STAND)
- Filter devices for clearing lysates
 - 2 mL or greater, Millipore Catalog # SLHVX13NL
 - 0.5 2 mL, Millipore Catalog # UFC40DV25
 - Less than 0.5 mL, Millipore Catalog # UFC30DV25
 - For 96 well plates, Millipore Catalog # MSBVN1210

SAFETY PRECAUTIONS

- All tissue components and biological materials should be handled as potentially hazardous. Follow universal precautions as established by the Centers for Disease Control and Prevention and by the Occupational Safety and Health Administration when handling and disposing of infectious agents.
- Sodium azide or Proclin has been added to some reagents as a preservative. Although the concentrations are low, sodium azide may react with lead and copper plumbing to form highly explosive metal azides. On disposal, flush with a large volume of water to prevent azide build up.

TECHNICAL GUIDELINES

To obtain reliable and reproducible results, the operator should carefully read this entire manual and fully understand all aspects of each assay step before running the assay. The following notes should be reviewed and understood before the assay is set up.

- FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.
- Do not use beyond the expiration date on the label.
- Do not mix or substitute reagents with those from other lots or sources.
- The Antibody-Immobilized Beads are light sensitive and must be protected from light at all times. Cover the assay plate containing beads with an opaque plate lid or aluminum foil during all incubation steps.
- Mix only the required amount of beads and biotin-labeled detection antibody prior to assay setup. Discard any unused premixed beads and biotin-labeled detection antibodys.
- It is important to allow all reagents to warm to room temperature (20-25°C) before use in the assay.
- The bottom of the Microtiter Filter Plate should not be in direct contact with any surface during assay set-up or incubation times. The plate can be set on a plate stand or on the non-flat side of the plate cover or any other plate holder to raise the plate from the surface. A plate stand can be purchased separately from Millipore (Millipore Catalog #MX-STAND).
- Incomplete washing can adversely affect the assay outcome. All washing must be performed with the appropriate Assay Buffer provided.
- After the wash steps, keep the bottom of the Microtiter Filter Plate clean by blotting on paper towels or absorbent pads to prevent any leakage due to capillary action.
- Keep the vacuum suction on the plate as low as possible. It is recommended to have a vacuum setting that will remove 200 μL of buffer in ≥ 5 seconds (equivalent to < 100 mmHg).
- The plate should be read immediately after the assay is finished. If, however, the plate cannot be read immediately, seal the plate, cover with aluminum foil or an opaque lid, and store the plate at 2-8°C for up to 24 hours. Prior to reading, agitate the plate on the plate shaker at room temperature for 10 minutes. Delay in reading a plate may result in decreased sensitivity.
- The titer plate shaker should be set at a speed to provide maximum orbital mixing without splashing of liquid outside the wells. For the recommended plate shaker, this would be a setting of 5-7 which is approximately 500-800 rpm.
- Ensure that the needle probe is clean. This may be achieved by sonication and/or alcohol flushes. Adjust probe height according to the protocols recommended by Luminex to the kit filter plate using 3 alignment discs prior to reading an assay.
- Vortex all reagents well before adding to plate.

SAMPLE COLLECTION AND STORAGE

A. Considerations for Cell Stimulation.

- Treating cells with growth factors (ex. EGF), cytokines (ex. TNFα), or other compounds (ex. Arsenite) induce a multitude of signaling cascades. The duration of stimulation in addition to the concentration of the respective factor/compound should be considered since they influence the degree of phosphorylation of any given analyte.
- Cellular responses to growth factors are typically improved when cells have been serum starved prior to treatment.
- Cell lines will differ in the robustness of their signaling response for any given stimulation.
- The suggested working range of protein concentration for the assay is 1 to $25 \ \mu g$ of total protein/well ($25 \ \mu L$ /well at 40 to 1000 μg /mL). A total protein amount of 10 μg / well is generally a good starting point for lysates for which target protein expression levels are unknown.

B. Preparation of cell lysates

MILLIPLEX MAP Lysis Buffer is supplied as **1X** stock solution. The lysis buffer contains phosphatase inhibitors *including* 1 mM sodium orthovanadate (Na₃VO₄) but does **NOT** contain protease inhibitors. It is recommended that protease inhibitors (Millipore catalog #20-201 or a similar product) be added immediately before use.

Suggested cell lysis protocol for adherent cells

- After treatments, wash cells with ice cold Phosphate Buffered Saline (PBS) and drain off PBS.
- Add ice cold **1X** MILLIPLEX MAP Lysis Buffer with freshly added protease inhibitors to cells (0.6 mL per 150 mm dish, 0.3 mL per 100 mm dish, or 0.1 mL per well of 24-well plate).
- Scrape adherent cells off the dish with a cell scraper. Transfer the cell suspension into a centrifuge tube and gently rock for 10-15 minutes at 4 °C.
- Remove particulate matter by filtration. Suggested Millipore filters:
 - 2 mL or greater, Millipore Catalog # SLPBDZ5NZ
 - o 0.5 2 mL, Millipore Catalog # UFC 0DV 25
 - Less than 0.5 mL, Millipore Catalog # UFC30DV00
- Aliquot and store the lysate at -70 ℃. The lysate should be stable for several months.
- It is recommended that the lysate be diluted at least 1:10 with PBS for determining the protein concentration with Coomassie-based assays or 1:4 for BCA assays. Alternatively, protein quantification may be omitted if an assay normalization control, such as the GAPDH MAPmates (catalog #46-667), is used.

Suggested cell lysis protocol for non-adherent cells

- Pellet the cells by centrifugation (500 1000 x g) in a tabletop centrifuge for 5 minutes.
- Wash the cells in ice cold TBS.

- Add ice cold **1X** MILLIPLEX MAP Lysis Buffer containing freshly prepared protease inhibitors to cells (1 mL per 1 x 10⁷ cells).
- Gently rock the lysate for 10-15 minutes at 4 °C.
- Remove particulate matter by filtration (See above). Aliquot and store the lysate at -70 °C. The lysate should be stable for several months.
- It is recommended that the lysate be diluted at least 1:10 in PBS for determining the protein concentration with Coomassie-based assays or 1:4 for BCA assays. Alternatively, protein quantification may be omitted if an assay normalization control, such as the GAPDH MAPmates (catalog #46-667), is used.

Cell lysis protocol for cells in sterile 96-well tissue culture plates

Adherent or non-adherent cells seeded or grown in sterile 96-well tissue culture grade plates (See supplemental protocols) can be washed, treated, and lysed in the same plate, but need to be filtered in a separate 96-well filter plate. Wash the cells by centrifugation in a microplate carrier 2 minutes at 500 x g.

- Remove the supernatant via aspiration and add 100 μ L of ice cold PBS.
- Centrifuge and remove supernatant via aspiration.
- Add 30-50 μL/well of ice cold 1X MILLIPLEX MAP Lysis Buffer containing freshly prepared protease inhibitors.
- Place the plate on an orbital shaker (600 800 rpm) for 10-15 minutes at 4 °C.
- Transfer the lysate to a 96-well filter plate that has been pre-wetted with 1X lysis buffer.
- Place a low protein binding, 96-well round bottom or V-bottom plate underneath the filter plate.
- Centrifuge the plates in a microplate carrier for 5 minutes at 500 x g.
- Store the filtered lysate at -70 °C until ready for use.
- It is recommended that the lysate be diluted at least 1:10 in PBS for determining the protein concentration with Coomassie-based assays or 1:4 for BCA assays. Alternatively, protein quantification may be omitted if an assay normalization control, such as the GAPDH MAPmates (catalog #46-667), is used.

PREPARATION OF REAGENTS FOR IMMUNOASSAY

A. Preparation of antibody Multi-pathway phosphoprotein beads

MILLIPLEX MAP capture beads are provided as a **20X** stock solution and should be protected from light.

- Sonicate **20X** stock capture beads for 15 seconds, then vortex for 30 seconds.
- Dilute the beads to **1X** by combining 0.125 mL beads with 2.375 mL of MILLIPLEX MAP Assay Buffer 2. Use the empty mixing vial provided.
- Vortex the **1X** capture beads for 15 seconds.
- For use, transfer 1X capture beads with a pipette, do not pour from mixing vial.
- Please note that multiplexing phospho-specific and total MAPmate pairs is not recommended due to cross-reactivity

B. Preparation of Biotin-Labeled Detection Antibody and Streptavidin-PE

MILLIPLEX MAP Detection Antibody is provided as a **20X** stock solution.

- Vortex the **20X** Detection Antibody stock for 10 seconds, it may be necessary to centrifuge briefly after vortexing for complete recovery of contents.
- Dilute the Detection Antibody to **1X** by combining 0.125 mL of Detection Antibody with 2.375 mL of MILLIPLEX MAP Assay Buffer 2. Use the empty bead-mixing vial provided.
- Vortex the MILLIPLEX MAP Streptavidin-Phycoerythrin 1:25 (SAPE) for 10 seconds.
- Dilute SAPE by combining 0.1ml of Streptavidin-Phycoerythrin with 2.4 mL of Milliplex MAP Cell Signaling Assay Buffer 2. Use the empty mixing vial provided.
- Transfer 1X biotinylated detection antibody and SAPE with a pipette do not pour from mixing vial.

C. Multiplexing additional MILLIPLEX MAP Cell Signaling MAPmates with the 8plex Multi-Pathway Signaling Kit, phosphoprotein.

Additional Cell Signaling Phospho-MAPmates may be combined with this kit, up to a maximum of 8 additional MAPmates.

Please note that Total MAPmate pairs should not be multiplexed with the 8-plex Multi-Pathway Signaling Kit, phosphoprotein.

- For each additional MAPmate, sonicate **20X** stock capture beads for 15 seconds, then vortex for 30 seconds.
- Add 125 µL 8-plex Multi-Pathway beads to the mixing vial
- For <u>each additional MAPmate</u>, add 125 μL from each antibody bead vial to the Mixing Bottle and bring final volume to 2.5 mL with Assay Buffer 2. Vortex the mixed beads well.
- Use the same preparation volumes for the Detection Antibody

Example 1: When using 2 additional MAPmates, add 125 μ L 8-plex Multi-Pathway Beads/ Detection Antibody and 125 μ L of each additional MAPmate Beads/ Detection Antibody to the mixing vial. Then add 2.125 mL Assay Buffer 2, for a final volume of 2.5 mL.

Example 2: When using 5 additional MAPmates, add 125 μ L 8-plex Multi-Pathway Beads/ Detection Antibody and 125 μ L of each additional MAPmate Beads/ Detection Antibody to the mixing vial. Then add 1.75 mL Assay Buffer 2, for a final volume of 2.5 mL.

D. Preparation of lyophilized MILLIPLEX MAP Cell Lysates (Catalog # 47-205, 47-210, 47-211, 47-212).


MILLIPLEX MAP HeLa Cell Lysate: Unstimulated (#47-205) is provided as a lyophilized stock of cell lysate prepared from unstimulated HeLa cells and is used as a negative control. MILLIPLEX MAP HeLa Cell Lysate: HS/Ars (#47-211) is provided as a lyophilized stock of cell lysate prepared from HeLa cells heat shocked for 30 min at 42°C then 16 hrs at 37°C followed by 200 μ M arsenite for 30 min. MILLIPLEX MAP HeLa Cell Lysate: TNF α (#47-212) is provided as a lyophilized stock of cell lysate prepared from HeLa cells stimulated with 100 ng/mL TNF- α (7 min). MILLIPLEX MAP A431 Cell Lysate: EGF (#47-210) is provided as a lyophilized stock of cell lysate prepared from A431 cells stimulated with 50 ng/mL of EGF (5 min). Each of the cell lysates were prepared in MILLIPLEX MAP Lysis Buffer containing protease inhibitors and lyophilized for stability. The lysates can be used as positive and negative control samples or alternatively, to create calibration curves for relative quantification of different phosphoprotein analytes.

MILLIPLEX MAP Cell Lysates as a positive and negative control

- Reconstitute each of the lyophilized cell lysates in 100 μ L of ultrapure water, for each vial this will yield 100 μ L of lysate at 2 mg/mL total protein.
- Gently vortex and incubate the reconstituted lysates for 5 min at RT (store on ice).
- Pipette 150 μL of MILLIPLEX MAP Assay Buffer 2 to each cell lysate vial. The cell lysate is now prepared for use in the MILLIPLEX MAP 8-plex Multi-Pathway Cell Signaling Assay.
- If desired, unused lysate may be stored in its original container at -80 °C for up to one month.

IMMUNOASSAY PROTOCOL

- 1. Dilute filtered lysates <u>at least</u> 1:1 in MILLIPLEX MAP Assay Buffer 2. The suggested working range of protein concentration for the assay is 1 to 25 μ g of total protein/well (25 μ L/well at 40 to 1,000 μ g/mL).
- Pre-wet filter plate with 25 μL/well of MILLIPLEX MAP Assay Buffer 2. Remove by vacuum filtration by placing the filter plate over a vacuum manifold and gently applying vacuum. Gently blot the bottom of the filter plate on a paper towel to remove excess liquid.
- 3. Vortex the **1X** bead suspension for 10 seconds. Add 25 μ L of 1X bead suspension to each well.
- Add 25 μL of diluted cell lysate (or reconstituted HeLa or A431 Cell Lysate Control) to each well and incubate overnight at 4 °C (or 2 hours RT) on a plate shaker (600-800rpm) protected from light.
- 5. Remove the lysate by vacuum filtration.
- Add 100 μL/well of MILLIPLEX MAP Assay Buffer
 Remove buffer by vacuum filtration and gently blot the bottom of the filter plate on a paper towel. Repeat this step again for a total of two washes.
- Add 25 μL/well of **1X** MILLIPLEX MAP Detection Antibody.
- 8. Incubate on a plate shaker for 1 hour at room temperature, protected from light.
- 9. Remove Detection Antibody by vacuum filtration and gently blot the bottom of the filter plate on a paper towel.
- 10. Add 25 μL of 1X MILLIPLEX MAP Streptavidin-Phycoerythrin (SAPE).
- 11. Incubate on a plate shaker for 15 minutes at room temperature, protected from light.
- 12. **DO NOT REMOVE** SAPE. Add 25 μL of MILLIPLEX MAP Amplification Buffer to each well.



- 13. Incubate on a plate shaker for 15 minutes at room temperature, protected from light.
- 14. Remove MILLIPLEX MAP SAPE /Amplification buffer by vacuum filtration and gently blot the bottom of the filter plate on a paper towel.
- 15. Resuspend beads in 150 μL of MILLIPLEX MAP Assay Buffer 2, and mix on plate shaker for 5 minutes.
- 16. Analyze using the Luminex[®] system.



Incubate 15 min at RT with shaking; dark

Remove Streptavidin-PE/ Amplification buffer and resuspend beads in 150 µL assay buffer. Read results using appropriate Luminex® instrument.

INSTRUMENT SETTINGS

These specifications are for Luminex100/200 instruments. Luminex instruments with other software (e.g. MasterPlex, StarStation, LiquiChip, Bio-Plex, LABScan100) would need to follow instrument instructions for gate settings and additional specifications from the vendors.

Events	50 per bead					
Sample Size	100 μL					
Gate Settings	8,000 to	0 15,000				
Reporter Gain	Default (I	_ow PMT)				
Time Out	60 seconds					
Bead Set						
	Erk/MAP Kinase 1/2	9				
	STAT3	14				
	JNK	16				
	p70 S6 kinase	19				
	ΙκΒα	33				
	STAT5A/B	35				
	CREB	37				
	p38 54					

SUPPLEMENTAL PROTOCOLS

A. Analysis of viscous cell lysates

Some cell lysates may not flow through the filter plate efficiently due to high viscosity or the formation of particulate matter from long-term storage. For these samples, the initial capture and wash steps can be done in microcentrifuge tubes. The beads are then transferred into 96-well filter plates for the rest of the assay.

- Add 25 µL/assay point of 1X beads to a 500 µL centrifuge tube.
- Next, add lysate diluted in MILLIPLEX MAP Assay Buffer 2 to a final volume of 100 μL or higher.
- Vortex the mixture at high speed for 15 seconds then sonicate for an additional 15 seconds.
- Rotate the mixture overnight at 4°C, protected from light.
- Centrifuge the beads for 1 min at 2,000 x g and carefully remove the supernatant to minimize bead loss.
- Resuspend the pelleted beads in 25 $\mu L/assay$ point of MILLIPLEX MAP Assay Buffer 2.
- Transfer 25 μL of the bead mixture to pre-wet filter plate wells and proceed to step 4 of the Immunoassay protocol.

REPLACEMENT REAGENTS

MILLIPLEX MAP Lysis Buffer	43-040
MILLIPLEX MAP Assay Buffer 2	43-041
MILLIPLEX MAP HeLa Cell Lysate: HS/Ars	47-211
MILLIPLEX MAP HeLa Cell Lysate: TNF α	47-212
MILLIPLEX MAP A431 Cell Lysate: EGF	47-210
MILLIPLEX MAP HeLa Cell Lysate: Unstimulated	47-205
MILLIPLEX MAP Streptavidin-Phycoerythrin	45-001D
MILLIPLEX MAP Amplification Buffer (1X)	43-024A
MILLIPLEX MAP 96-well Filter Plate	MX-PLATE

REPRESENTATIVE DATA

8-plex Multi-Pathway analysis of TNF- α treated HeLa cells

8-plex Multi-Pathway analysis of EGF treated HeLa cells




Figure 1. Multiplex analysis of HeLa cells treated with EGF, TNF- α , or heat shock/arsenite. HeLa cells stimulated with 50 ng/mL of EGF (5 min), 100 ng/mL TNF- α (7 min), or heat shocked for 30 min. at 42°C then 16 hrs at 37°C followed by 200 μ M arsenite for 30 min were assayed. The cells were lysed in MILLIPLEX MAP Lysis Buffer containing protease inhibitors. 20 μ g total protein of each lysate diluted in MILLIPLEX MAP Assay Buffer 2 were analyzed according the Assay protocol (lysate incubation at 4°C overnight). The Median Fluorescence Intensity (MFI) was measured with the Luminex[®] system. The figures represent the average and standard deviation of three replicate wells.

JNK

S6K

lkB

STAT5

CREB

p38

Erk

STAT3




Figure 2. Immunoprecipitation/Western Blot analysis of multiplexed analytes in HeLa cells. 10 μ g of lysates (described in Figure 1) were mixed with capture antibodies to immunoprecipitate each respective protein. The immunoprecipitated proteins were separated on SDS-PAGE, transferred to nitrocellulose, and probed with biotin labeled phospho-specific detection antibodies. The proteins were imaged using Streptavidin-HRP and chemiluminescence.

TROUBLESHOOTING GUIDE

Situation	Possible Problem	Solution			
	Mechanical	<u>Mechanical</u>			
Data acquisition time exceeds 30 seconds per well and/or "Sample Empty" occurs	 Needle height is not correct (common problem due to variation in well depth between plate types). Sample needle is clogged. Air in the system. Low pressure in the system. 	 Adjust needle height using 2 disks using "options; XY setup function". Remove needle (see user's manual) and sonicate it to remove obstruction. Perform alcohol flush and then wash with alcohol; followed by a wash with sheath fluid and a prime. <u>Tighten</u> tops on the sheath fluid container. Loosen the top on the waste container. 			
	Sample Related	Sample Related			
	 Lysate concentration too high. Particulate matter in lysate. No buffer in well. 	 Refer to lysate preparation protocol. Centrifuge or filter lysate to remove particulate matter. 			
Readings are lower than expected	 Gate is not set properly. Calibration is not correct. Cells responded poorly to stimulation. 	 Right click within the "doublet discriminator" and create gate. Move lines to 8,000 and 13,500 Run a machine calibration with calibration beads available from Luminex[®] Corp. Check stimulation conditions. 			

Bead pattern is diffuse and missing the bead target (white oval)	 Precipitate buildup in system. Calibration is not correct. 	 Drain the system, followed by a backflush, proceed with solution for air in the system. 			
	3. Incompatible buffer used to resuspend beads for Luminex [®] analysis.	 Run a machine calibration with calibration beads available from Luminex[®] Corp. Vacuum plate and resuspend beads in MILLIPLEX MAP Assay Buffer 2. 			
Applying vacuum to filter plate does not remove liquid from wells	 Wells not in use are empty. Particulate matter in lysate. 	 Place tape over the top of empty wells that are not in use. Pre-filter lysate before use. 			
Aggregation of beads during analysis	 Particulate matter in lysate. Incompatible lysis buffer. Beads not sonicated for step 3 of main protocol. 	 Pre-filter the lysate. Use recommended lysis buffer. Sonicate <i>resuspended</i> 1X beads prior to adding to filter plate. 			
Liquid wicking out from 96- well filter plate	 Plate was not blotted on paper towel prior to adding next reagent. The bottom of the plate is in contact with absorbent material such as paper towel or bench paper. 	 After vacuum step, gently blot the bottom of the plate using a paper towel to remove excess liquid. Place the plate on a flat, non-absorbent surface during loading steps. 			

ORDERING INFORMATION

To place an order:

FAX: (636) 441-8050

Include:

- Your name, telephone and/or fax number
- Customer account number
- Shipping and billing address
- Purchase order number
- Catalog number and description of product
- Quantity of kits

Toll-Free US: (800) MILLIPORE Mail Orders: Millipore Corp. Attn: Customer Service 6 Research Park Drive St. Charles, Missouri 63304 U.S.A.

For International Customers:

To best serve our international customers in placing an order or obtaining additional information about MILLIPLEX MAP products, please contact your multiplex specialist or sales representative or email our European Customer Service at <u>customerserviceEU@Millipore.com</u>.

Conditions of Sale

All products are for research use only. They are not intended for use in clinical diagnosis or for administration to humans or animals. All products are intended for *in vitro* use only.

Material Safety Data Sheets (MSDS)

Material Safety Data Sheets for Millipore products may be ordered by fax or phone or through our website at www.millipore.com/techlibrary/index.do

Technical Services

For product technical assistance call or write.

Toll-Free US: (800) MILLIPORE

Email: techserv_dd@millipore.com

WELL MAP

	1	2	3	4	5	6	7	8	9	10	11	12
A	HeLa unstim negative control											
в	HeLa unstim negative control											
с	HeLa: HS/Ars positive control											
D	HeLa: HS/Ars positive control											
E	HeLa: TNFα positive control											
F	HeLa: TNFα positive control											
G	A431: EGF positive control											
н	A431: EGF positive control											