



Peninsula Laboratories, LLC

A Member of the Bachem Group

305 Old County Road, San Carlos, CA 94070

Tel: (800) 922-1516 • (650) 592-5392

Fax: (650) 595-4071

www.bachem.com



Cat. No. S-1152

EIA for Somatostatin-14

(for human serum and plasma samples)

SPECIFICATIONS: Extraction-free kit for human serum/plasma


STANDARD: 1 µg H-1490 H-Ala-Gly-Cys-Lys-Asn-Phe-Phe-Trp-Lys-Thr-Phe-Thr-Ser-Cys-OH, (Disulfide bond)

PROTOCOL: V (Ab1hr.Std2hr.BtON)

RANGE: 0-25 ng/ml

AVERAGE IC₅₀: 0.8 ng/ml

SHELF-LIFE: One year from production date



Suggested Preparation of Standards

	ng/ml	Range: 0.02 to 25 ng/ml
Stock	1000	
S1	25.00	Add 25 µl stock + 975 µl diluent
S2	6.25	Add 200 µl S1 + 600 µl diluent
S3	1.56	Add 200 µl S2 + 600 µl diluent
S4	0.39	Add 200 µl S3 + 600 µl diluent
S5	0.10	Add 200 µl S4 + 600 µl diluent
S6	0.02	Add 200 µl S5 + 600 µl diluent
S0	0.00	500 µl diluent

CROSS REACTION:

PEPTIDE	%
Somatostatin-14	100
Somatostatin-28	100
Somatostatin-25	100
[Des-Ala ¹]-Somatostatin-14	100
[D-Trp ⁸]-Somatostatin-14	0.05
Prosomatostatin (1-32) (porcine)	0.002
Somatostatin Analog, RC-160	0
H-D-Phe-Cys-Tyr-D-Trp-Orn-Thr-Pen-Thr-NH ₂	0
Substance P	0
NPY (porcine)	0
VIP (human, bovine, rat, porcine)	0
Insulin (human)	0
Glucagon (1-29) (human)	0
Amylin (human)	0