



Peninsula Laboratories, LLC
A Member of the Bachem Group
 305 Old County Road, San Carlos, CA 94070
 Tel: (800) 922-1516 • (650) 592-5392
 Fax: (650) 595-4071
 www.bachem.com




Cat. No. S-1330
Angiotensin I/II (1-7)

SPECIFICATIONS: Formulated for extracted samples

STANDARD: 1 µg H-1715 H-Asp-Arg-Val-Tyr-Ile-His-Pro-OH

PROTOCOL: III (Std.Ab1hr.Bt)
RANGE: 0-10 ng/ml
AVERAGE IC₅₀: 0.25 ng/ml
SHELF-LIFE: One year from production date



Suggested Preparation of Standards		
	ng/ml	Range: 0.01 to 10 ng/ml
Stock	1000	
S1	10.00	Add 10 µl stock + 990 µl diluent
S2	2.50	Add 200 µl S1 + 600 µl diluent
S3	0.63	Add 200 µl S2 + 600 µl diluent
S4	0.16	Add 200 µl S3 + 600 µl diluent
S5	0.04	Add 200 µl S4 + 600 µl diluent
S6	0.01	Add 200 µl S5 + 600 µl diluent
S0	0.00	500 µl diluent

CROSS REACTION:

PEPTIDE	%
Angiotensin I/II (1-7)	100
Angiotensin I	0
Angiotensin II	0
Angiotensin III	0
Angiotensin A	0