

Peninsula Laboratories, LLC

A Member of the Bachem Group

305 Old County Road, San Carlos, CA 94070 Tel: (800) 922-1516 • (650) 592-5392 Fax: (650) 595-4071 www.bachem.com

S-5000 Extraction kit

Contents:

Buffer A (x 2) - 1% trifluoroacetic acid (Y-1040) Buffer B (x 1) - 60% acetonitrile with 1% trifluoroacetic acid (Y-1045) Sep-Column (50 columns) - C18 sorbent (Y-1000) The kit components should be enough to extract 50 samples.

Storage:

Kit components can be stored at room temperature. The Sep-columns can be kept in its original package and must be stored in desiccators.

Recommended Use:

Extraction of Peptide from Plasma

- 1. Add an equal amount of Buffer A to the plasma. For example, if you are using 1 ml of plasma, acidify with 1 ml of Buffer A. Centrifuge at 6,000xg to 17,000xg for 20 minutes at 4°C. Discard any pellet that may be present.
- 2. Equilibrate a SEP-COLUMN containing 200 mg of C18 (Cat No. Y-1000) by washing with Buffer B or Buffer D (100% Acetonitrile) (1 ml, once) followed by Buffer A (3 ml, three times).
- 3. Load the plasma solution onto the pre-treated C-18 SEP-COLUMN.

Note: For steps 4 and 5, a light vacuum (10 sec/drop) may be applied to the column.

- 4. Slowly wash the column with Buffer A (3 ml, twice) and discard the wash.
- 5. Elute the peptide slowly with Buffer B (3 ml, once) and collect eluant in a polypropylene tube.
- 6. Evaporate eluant to dryness using a centrifugal concentrator or by a suitable method;
 - Ex: Lyophilizer. Recommended to freeze eluant with dry ice/methanol for fastest freeze.
- 7. Dissolve the residue in a suitable volume of assay (EIA or RIA) buffer such that the concentration of the substance of interest will fall close to the IC_{50} (within the measuring range).
- 8. The total time for extraction should be 0.5 days for extraction and 0.5 days for lyophilization. Once the sample is lyophilized, it can be stored at -70°C before assaying, but should be assayed as soon as possible.

NOTE: This is a generic extraction protocol, which can be used for multiple types of samples. If a more suitable extraction method is found in literature, we encourage the researcher to use it.