

**DRG<sup>®</sup> 17- $\alpha$ -OH Progesterone CLIA (CLA-4665)**

REVISED 17 OCT. 2008 (VERS. 2.1)

**1 INTRODUCTION****1.1 Intended Use**

The **DRG 17- $\alpha$ -OH Progesterone Chemiluminescence Immunoassay Kit** provides materials for the quantitative determination of 17- $\alpha$ -OH Progesterone in serum.

**This assay is intended for in vitro diagnostic use only.**

**1.2 Summary and Explanation**

The steroid 17- $\alpha$ -Hydroxyprogesterone (17- $\alpha$ -OHP) is produced by both the adrenal cortex and gonads. Even though 17- $\alpha$ -OHP has relatively little progestational activity, it is of intense clinical interest because it is the immediate precursor to 11-desoxycortisol (Cpd-S). Because Cpd-S is produced by 21-hydroxylation of 17- $\alpha$ -OHP, measurement of 17- $\alpha$ -OHP is a useful indirect indicator of 21-hydroxylase activity. In congenital 21-hydroxylase deficiency, the most common variety of congenital adrenal hyperplasia (CAH), 17- $\alpha$ -OHP is secreted in abundant excess. It is moderately elevated in the 11- $\beta$ -hydroxylase deficiency as well. Measurement of 17- $\alpha$ -OHP is therefore valuable in the initial diagnosis of CAH.

**1.3 Clinical Physiology****Adult non-pregnant women:**

In adult non-pregnant women in the childbearing age group, 17- $\alpha$ -OHP concentrations vary over the menstrual cycle with luteal phase concentrations being higher than follicular phase concentrations. This is because 17- $\alpha$ -OHP is secreted parallel with progesterone from maturing follicles or from the corpus luteum. There is also a diurnal variation of 17- $\alpha$ -OHP concentrations.

This rhythm is parallel with adrenal cortisol secretion such that maximum 17- $\alpha$ -OHP concentrations are measured in samples obtained between midnight and 8:00 am.

**Adult males:**

There is little information available on the systematic variability of 17- $\alpha$ -OHP concentration in adult males.

**Pregnant women and newborn children:**

The steroid 17- $\alpha$ -OHP is produced in large amounts by the fetus and the adrenals. It is secreted in abundance into both the fetal and maternal circulation. The maternal concentrations of 17- $\alpha$ -OHP increase very sharply after 32 weeks gestational age to about 4-fold above basal concentrations at term.

**1.4 Clinical Application****Congenital adrenal hyperplasia:**

The principal application of the 17- $\alpha$ -OHP RIA is in the diagnosis of CAH in newborns with ambiguous genitalia and in virilized adolescent girls. Since 17- $\alpha$ -OHP is the immediate precursor to 11-desoxycortisol, basal 17- $\alpha$ -OHP concentrations are sharply elevated in patients with 21-hydroxylase deficiency and to a lesser degree in patients with 11-hydroxylase deficiency.

Because 17- $\alpha$ -OHP concentrations are so markedly elevated in newborns and adolescent girls afflicted with CAH, a single basal measurement is all that is normally required to make the diagnosis.

**Late onset adrenal hyperplasia:**

**DRG<sup>®</sup> 17- $\alpha$ -OH Progesterone CLIA (CLA-4665)****REVISED 17 OCT. 2008 (VERS. 2.1)**

More recently, 17- $\alpha$ -OHP concentrations have been utilized in the evaluation of androgenized women where late onset 21-hydroxylase is suspected. This condition is clinically very subtle and since the presentation is the same as classical polycystic ovarian disease, basal plasma 17- $\alpha$ -OHP concentrations, unlike classical congenital adrenal hyperplasia, are normal. The diagnosis is made by administration of an ACTH stimulation test.

**Other applications:**

Measurement of 17- $\alpha$ -OHP concentrations is also utilized in evaluation of both men and women with acne vulgaris, male pattern baldness and in some subtle forms of infertility. Experiences with these applications are very limited.

**2 PRINCIPLE OF THE TEST**

The DRG 17- $\alpha$ -OH Progesterone CLIA Kit is a chemiluminescence immunoassay (CLIA), based on the principle of competitive binding.

The microtiter wells are coated with a polyclonal antibody directed towards an antigenic site on the 17- $\alpha$ -OHP molecule. Endogenous 17- $\alpha$ -OHP of a patient sample competes with a 17- $\alpha$ -OHP-horseradish peroxidase conjugate for binding to the coated antibody. After incubation the unbound conjugate is washed off.

The amount of bound peroxidase conjugate is inversely proportional to the concentration of 17- $\alpha$ -OHP in the sample. After addition of the substrate solution, the intensity of emitted light is inversely proportional to the concentration of 17- $\alpha$ -OHP in the patient sample.

**3 PRECAUTIONS**

- This kit is for in vitro diagnostic use only.
- For information on hazardous substances included in the kit please refer to Material Safety Data Sheets.
- All reagents of this test kit which contain human serum or plasma have been tested and confirmed negative for HIV I/II, HBsAg and HCV by FDA approved procedures. All reagents, however, should be treated as potential biohazards in use and for disposal.
- Never pipet by mouth and avoid contact of reagents and specimens with skin and mucous membranes.
- Do not smoke, eat, drink or apply cosmetics in areas where specimens or kit reagents are handled.
- Wear disposable latex gloves when handling specimens and reagents. Microbial contamination of reagents or specimens may give false results.
- Handling should be in accordance with the procedures defined by an appropriate national biohazard safety guideline or regulation.
- Do not use reagents beyond expiry date as shown on the kit labels.
- All indicated volumes have to be performed according to the protocol. Optimal test results are only obtained when using calibrated pipettes and microtiter plate luminometer.
- The luminescence substrate reagents (*Reagent A* and *Reagent B*) are sensitive to light and should be stored in the original dark bottle away from direct sunlight.
- Do not mix or use components from kits with different lot numbers. It is advised not to exchange wells of different plates even of the same lot. The kits may have been shipped or stored under different conditions and the binding characteristics of the plates may result slightly different.
- Chemicals and prepared or used reagents have to be treated as hazardous waste according the national biohazard safety guideline or regulation.
- Safety Data Sheets for this product are available upon request directly from DRG International, Inc.  
The Safety Data Sheets fit the demands of: EU-Guideline 91/155 EC.

**DRG<sup>®</sup> 17- $\alpha$ -OH Progesterone CLIA (CLA-4665)****REVISED 17 OCT. 2008 (VERS. 2.1)****4 REAGENTS****4.1 Reagents provided**

1. **Microtiterwells**, 12x8 (break apart) strips, 96 wells;  
Wells coated with a anti-17- $\alpha$ -OHP antibody (polyclonal).
2. **Standard (Standard 0-6)**, 7 vials, 1 mL, ready to use;  
Concentrations: 0; 0.15; 0.5; 1.5; 3; 7.5; 20 ng/mL  
Conversion: 1 ng/mL = 3.0 nmol/L  
contain 0.03% Proclin 300, 0.015% BND and 0.010% MIT as a preservatives.
3. **Enzyme Conjugate**, 1 vial, 12 mL, ready to use;  
17- $\alpha$ -OHP conjugated to horseradish Peroxidase;  
contains 0.03% Proclin 300, 0.015% BND and 0.010% MIT as a preservative.
4. **Chemiluminescence Substrate Solution**,  
**Reagent A**, 1 vial, 2 mL, *Note: light sensitive!*  
**Reagent B**, 1 vial, 2 mL, *Note: light sensitive!*  
**Reagent C**, 1 vial, 10 mL  
see „Preparation of Reagents“.
5. **Wash Solution**, 1 vial, 30 mL (40X concentrated);  
see „Preparation of Reagents“.

\* BND = 5-bromo-5-nitro-1,3-dioxane  
MIT = 2-methyl-2H-isothiazol-3-one

**Note:** Additional *Standard 0* for sample dilution is available upon request.

**4.2 Equipment and material required but not provided**

- A microtiter plate luminometer.
- Calibrated variable precision micropipettes.
- Absorbent paper.
- Distilled or deionized water
- Timer
- Semi logarithmic graph paper or software for data reduction

**4.3 Storage Conditions**

When stored at 2-8°C unopened reagents will retain reactivity until expiration date. Do not use reagents beyond this date. Opened reagents must be stored at 2-8°C. Microtiter wells must be stored at 2-8°C. Once the foil bag has been opened, care should be taken to close it tightly again. Opened kits retain activity for six weeks if stored as described above.

**DRG<sup>®</sup> 17- $\alpha$ -OH Progesterone CLIA (CLA-4665)****REVISED 17 OCT. 2008 (VERS. 2.1)****4.4 Reagent Preparation**

Allow all reagents and required number of strips to reach room temperature prior to use.

***Wash Solution***

Add deionized water to the 40X concentrated Wash Solution.

Dilute 30 mL of concentrated Wash Solution with 1170 mL deionized water to a final volume of 1200 mL.

*The diluted Wash Solution is stable for 2 weeks at room temperature.*

***Chemiluminescence Substrate Solution***

Mix **1 part** of the chemiluminescence **Reagent A** with **1 parts** of **Reagent B** and dilute this mixture 1:4 with **Reagent C**.

This gives the ready to use substrate solution.

The prepared substrate solution is stable for one hour. Prepare fresh before use.

If the whole plate is to be used prepare the substrate solution as follows:

Add 1.5 mL of each **Reagent A** and **Reagent B** into 9 mL **Reagent C**

**4.5 Disposal of the Kit**

The disposal of the kit must be made according to the national regulations. Special information for this product is given in the Material Safety Data Sheets (see chapter 13).

**4.6 Damaged Test Kits**

In case of any severe damage to the test kit or components, DRG has to be informed in writing, at the latest, one week after receiving the kit. Severely damaged single components should not be used for a test run. They have to be stored until a final solution has been found. After this, they should be disposed according to the official regulations.

**5 SPECIMEN COLLECTION AND PREPARATION**

Serum can be used in this assay.

Do not use haemolytic, icteric or lipaemic specimens.

*Please note:* Samples containing sodium azide should not be used in the assay.

**5.1 Specimen Collection****Serum:**

Collect blood by venipuncture (e.g. Sarstedt Monovette # 02.1388.001), allow to clot, and separate serum by centrifugation at room temperature. Do not centrifuge before complete clotting has occurred. Patients receiving anticoagulant therapy may require increased clotting time.

**5.2 Specimen Storage and Preparation**

Specimens should be capped and may be stored for up to 24 hours at 2-8°C prior to assaying.

Specimens held for a longer time should be frozen only once at -20°C prior to assay. Thawed samples should be inverted several times prior to testing.

**DRG<sup>®</sup> 17- $\alpha$ -OH Progesterone CLIA (CLA-4665)****REVISED 17 OCT. 2008 (VERS. 2.1)****5.3 Specimen Dilution**

If in an initial assay, a specimen is found to contain more than the highest standard, the specimens can be diluted with *Standard 0* and reassayed as described in Assay Procedure.

For the calculation of the concentrations this dilution factor has to be taken into account.

Example:

- a) Dilution 1:10: 10  $\mu$ L Serum + 90  $\mu$ L *Standard 0* (mix thoroughly)  
b) Dilution 1:100: 10  $\mu$ L dilution a) 1:10 + 90  $\mu$ L *Standard 0* (mix thoroughly).

**6 ASSAY PROCEDURE****6.1 General Remarks**

- All reagents and specimens must be allowed to come to room temperature before use. All reagents must be mixed without foaming.
- Once the test has been started, all steps should be completed without interruption.
- Use new disposal plastic pipette tips for each standard, control or sample in order to avoid cross contamination.
- Light intensity is a function of the incubation time and temperature. Before starting the assay, it is recommended that all reagents are ready, caps removed, all needed wells secured in holder, etc. This will ensure equal elapsed time for each pipetting step without interruption.
- As a general rule the enzymatic reaction is linearly proportional to time and temperature.

DRG<sup>®</sup> 17- $\alpha$ -OH Progesterone CLIA (CLA-4665)

REVISED 17 OCT. 2008 (VERS. 2.1)

**6.2 Test Procedure**

Each run must include a standard curve.

1. Secure the desired number of Microtiter wells in the holder.
2. Dispense **20  $\mu$ L** of each *Standard*, *Control* and samples with new disposable tips into appropriate wells.
3. Dispense **100  $\mu$ L** *Enzyme Conjugate* into each well.
4. Thoroughly mix for 10 seconds. It is important to have a complete mixing in this step.
5. Incubate for **60 minutes** at room temperature.
6. Briskly shake out the contents of the wells.  
Rinse the wells 5 times with diluted Wash Solution (400  $\mu$ L per well). Strike the wells sharply on absorbent paper to remove residual droplets.  
**Important note:**  
The sensitivity and precision of this assay is markedly influenced by the correct performance of the washing procedure!
7. Add **100  $\mu$ L** of the freshly prepared Substrate Solution to each well. (See "*Preparation of Reagents.*")
8. Incubate for **10 minutes** at room temperature.
9. Read the RLU with a microtiter plate luminometer **within 20 minutes** after incubation time of substrate.

**6.3 Calculation of Results**

1. Calculate the average Relativ Light Units (RLU) values for each set of standards, controls and patient samples.
2. Construct a standard curve by plotting the mean RLU obtained from each standard against its concentration with RLU value on the vertical(Y) axis and concentration on the horizontal (X) axis.
3. Using the mean RLU value for each sample determine the corresponding concentration from the standard curve.
4. Automated method: The results in the IFU have been calculated automatically using a 4 PL (4 Parameter Logistics) curve fit. Other data reduction functions may give slightly different results.
5. The concentration of the samples can be read directly from this standard curve. Samples with concentrations higher than that of the highest standard have to be further diluted. For the calculation of the concentrations this dilution factor has to be taken into account.

**6.3.1 Example of Typical Standard Curve**

The following data\*\* is for demonstration only and **cannot** be used in place of data generations at the time of assay.



**DRG® 17- $\alpha$ -OH Progesterone CLIA (CLA-4665)**

**REVISED 17 OCT. 2008 (VERS. 2.1)**

<b>Standard</b>	<b>RLU (x10<sup>3</sup>)</b>	<b>RLU/RLU<sub>max</sub> (%)</b>
Standard 0 (0 ng/mL)	3783	100
Standard 1 (0.15 ng/mL)	2529	66.8
Standard 2 (0.5 ng/mL)	1395	36.9
Standard 3 (1.5 ng/mL)	631	16.7
Standard 4 (3 ng/mL)	338	8.9
Standard 5 (7.5 ng/mL)	149	3.9
Standard 6 (20 ng/mL)	71	1.9

\*\* It is recommended to use the RLU/RLU<sub>max</sub> values for comparative purposes since luminometers vary considerably between manufacturers. Results from different luminometers will show different RLU values, however, the RLU/RLU<sub>max</sub> values remain consistent.



## DRG® 17- $\alpha$ -OH Progesterone CLIA (CLA-4665)

REVISED 17 OCT. 2008 (VERS. 2.1)

### 7 EXPECTED NORMAL VALUES

It is strongly recommended that each laboratory should determine its own normal and abnormal values.

In a study using the DRG 17- $\alpha$ -OH Progesterone CLIA the following values are observed:

Newborn	5. – 30. day	< 0.7 - 2.5 ng/mL
	31. – 60. day (male)	0.8 - 5.0 ng/mL
	31. – 60. day (female)	0.5 - 2.3 ng/mL
Children	3 - 14 years	0.07 - 1.7 ng/mL
Reproductive aged women	Follicular phase:	0.1 - 0.8 ng/mL
	Luteal phase:	0.6 - 2.3 ng/mL
	Ovulation:	0.3 - 1.4 ng/mL
	Post ACTH:	< 3.2 ng/mL
	Third trimester:	2.0 - 12 ng/mL
	Postmenopausal women	0.13 - 0.51 ng/mL
Normal men		0.5 - 2.1 ng/mL

The results alone should not be the only reason for any therapeutic consequences. The results should be correlated to other clinical observations and diagnostic tests.

### 8 QUALITY CONTROL

Good laboratory practice requires that controls be run with each calibration curve. A statistically significant number of controls should be assayed to establish mean values and acceptable ranges to assure proper performance.

It is recommended to use control samples according to state and federal regulations. The use of control samples is advised to assure the day to day validity of results. Use controls at both normal and pathological levels.

The controls and the corresponding results of the QC-Laboratory are stated in the QC certificate added to the kit. The values and ranges stated on the QC sheet always refer to the current kit lot and should be used for direct comparison of the results.

It is also recommended to make use of national or international Quality Assessment programs in order to ensure the accuracy of the results.

Employ appropriate statistical methods for analysing control values and trends. If the results of the assay do not fit to the established acceptable ranges of control materials patient results should be considered invalid.

In this case, please check the following technical areas: Pipetting and timing devices; photometer, expiration dates of reagents, storage and incubation conditions, aspiration and washing methods.

After checking the above mentioned items without finding any error contact your distributor or DRG directly.

**DRG<sup>®</sup> 17- $\alpha$ -OH Progesterone CLIA (CLA-4665)**

REVISED 17 OCT. 2008 (VERS. 2.1)

**9 PERFORMANCE CHARACTERISTICS****9.1 Assay Dynamic Range**

The range of the assay is between 0.05 – 20 ng/mL.

**9.2 Specificity of Antibodies (Cross Reactivity)**

The following substances were tested for cross reactivity of the assay:

<b>Steroid</b>	<b>% Cross Reaction</b>
17- $\alpha$ -OH Progesterone	100.0
Estriol	< 0.01
Estradiol 17 $\beta$	< 0.01
Testosterone	< 0.01
Dihydrotestosterone	< 0.01
DOC	0.05
11-Desoxicortisol	1.4
Progesterone	1.2
DHEA	< 0.01
DHEAS	< 0.001
Cortisol	< 0.01
Corticosterone	< 0.05
Aldosterone	< 0.01
Androstendione	< 0.01
Dehydroepiandrosten sulfate	< 0.01
Prednison	< 0.01

**9.3 Analytical Sensitivity**

The analytical sensitivity was calculated from the mean minus two standard deviations of twenty (20) replicate analyses of *Standard 0* and was found to be 0.05 ng/mL.

**9.4 Precision****9.4.1 Intra Assay Variation**

The within assay variability is shown below:

Sample	n	Mean (ng/mL)	CV (%)
1	20	2.77	4.62
2	20	3.05	4.51
3	20	3.86	3.60

# DRG® 17- $\alpha$ -OH Progesterone CLIA (CLA-4665)



REVISED 17 OCT. 2008 (VERS. 2.1)

## 9.4.2 Inter Assay Variation

The between assay variability is shown below:

Sample	n	Mean (ng/mL)	CV (%)
1	12	1.84	1.77
2	12	2.29	3.93
3	12	11.42	4.73

## 9.5 Recovery

Recovery of the DRG CLIA was determined by adding increasing amounts of the analyte to three different patient sera containing different amounts of endogenous analyte. Each sample (non-spiked and spiked) was assayed and analyte concentrations of the samples were calculated from the standard curve. The percentage recoveries were determined by comparing expected and measured values of the samples.

Sample	Endogenous 17 $\alpha$ OH-P ng/mL	Added 17 $\alpha$ OH-P ng/mL	Measured Conc. 17 $\alpha$ OH-P ng/mL	Expected Conc 17 $\alpha$ OH-P ng/mL	Recovery ( % )
1	14.31	10	17.79	17,16	103,7
		3.75	9.82	10,91	90,1
		1.5	7.43	8,66	85,8
2	1.71	10	10.12	10,86	93,2
		3.75	4.07	4,61	88,4
		1.5	2.20	2,36	93,4
3	3,10	10	11,68	11,55	101,1
		3,75	4,97	5,30	93,8
		1,5	2,61	3,05	85,6

## DRG<sup>®</sup> 17- $\alpha$ -OH Progesterone CLIA (CLA-4665)



REVISED 17 OCT. 2008 (VERS. 2.1)

### 9.6 Linearity

Sample	Dilution	Measured Conc. 17 $\alpha$ OH-P (ng/mL)	Expected Conc. 17 $\alpha$ OH-P (ng/mL)	Recovery (%)
1	None	1.04	1.04	
	1:2	0.51	0.52	98.1
	1:4	0.25	0.26	96.2
	1:8	0.12	0.13	92.3
	1:16	0.07	0.07	107.7
2	None	3.15	3.15	
	1:2	1.77	1.58	112.4
	1:4	0.85	0.79	107.9
	1:8	0.43	0.39	109.2
	1:16	0.21	0.20	106.7
3	None	2.67	2.67	
	1:2	1.36	1.34	101.9
	1:4	0.72	0.67	107.9
	1:8	0.35	0.33	104.9
	1:16	0.19	0.17	113.9

## 10 LIMITATIONS OF USE

Reliable and reproducible results will be obtained when the assay procedure is performed with a complete understanding of the package insert instruction and with adherence to good laboratory practice.

Any improper handling of samples or modification of this test might influence the results.

### 10.1 Interfering Substances

Haemoglobin (up to 4 mg/mL), Bilirubin (up to 0.5 mg/mL) and Triglyceride (up to 7.5 mg/mL) have no influence on the assay results.

### 10.2 Drug Interferences

Until today no substances (drugs) are known to us, which have an influence to the measurement of 17- $\alpha$ -OH Progesterone in a sample.

### 10.3 High-Dose-Hook Effect

No hook effect was observed in this test.

**DRG<sup>®</sup> 17- $\alpha$ -OH Progesterone CLIA (CLA-4665)****REVISED 17 OCT. 2008 (VERS. 2.1)****11 LEGAL ASPECTS****11.1 Reliability of Results**

The test must be performed exactly as per the manufacturer's instructions for use. Moreover the user must strictly adhere to the rules of GLP (Good Laboratory Practice) or other applicable national standards and/or laws. This is especially relevant for the use of control reagents. It is important to always include, within the test procedure, a sufficient number of controls for validating the accuracy and precision of the test.

The test results are valid only if all controls are within the specified ranges and if all other test parameters are also within the given assay specifications. In case of any doubt or concern please contact DRG.

**11.2 Therapeutic Consequences**

Therapeutic consequences should never be based on laboratory results alone even if all test results are in agreement with the items as stated under point 11.1. Any laboratory result is only a part of the total clinical picture of a patient.

Only in cases where the laboratory results are in acceptable agreement with the overall clinical picture of the patient should therapeutic consequences be derived.

The test result itself should never be the sole determinant for deriving any therapeutic consequences.

**11.3 Liability**

Any modification of the test kit and/or exchange or mixture of any components of different lots from one test kit to another could negatively affect the intended results and validity of the overall test. Such modification and/or exchanges invalidate any claim for replacement.

Claims submitted due to customer misinterpretation of laboratory results subject to point 11.2. are also invalid.

Regardless, in the event of any claim, the manufacturer's liability is not to exceed the value of the test kit. Any damage caused to the test kit during transportation is not subject to the liability of the manufacturer.

**DRG® 17- $\alpha$ -OH Progesterone CLIA (CLA-4665)**



**REVISED 17 OCT. 2008 (VERS. 2.1)**

**12 REFERENCES**

1. Abraham, G.E., R.S. Swerdloff, D. Tulchinsky et al: Radioimmunoassay of plasma 17-hydroxyprogesterone. J. Clin. Endocrinol. Metab. 33:42, 1971
2. Chrousos, G.P., D. L. Loriaux, D.L. Mann, et al: Late onset 21- hydroxylase deficiency mimicking idiopathic hirsutism or polycystic ovarian disease. Annals Intern. Med. 96:143, 1982.
3. Buster, J.E., R.J. Chang, D.L. Preston, et al: Interrelationships of circulating maternal steroids; progesterone, 16 $\alpha$ -hydroxyprogesterone, 17 $\alpha$ -hydroxyprogesterone, 20 $\alpha$ -dihydroprogesterone, gamma-5-pregnenolone, gamma-5-pregnenolone-sulfate, gamma-5-pregnenolone-sulfate and 17-hydroxy gamma-5-pregnenolone, J. Clin. Endocrinol. Metab. 48:133, 1979.
4. New, M.I., B. Dupont, S. Pang, et al: An update on congenital adrenal hyperplasia. Recent Progress in Hormone Research, 37:105, 1981. S. Pang S.
5. J. Hotchkiss, A. Drash, et al: Micro filter paper method for 17 $\alpha$ -hydroxyprogesterone radioimmunoassay: Its application for rapid screening for congenital adrenal hyperplasia. J. Clin. Endocrinol. Metab., 45:1003, 1977.
6. Lobo, R.A., U. Goebelsmann: Adult manifestation of congenital adrenal hyperplasia due to incomplete 21-hydroxylase deficiency mimicking polycystic ovarian disease. Am. J. Obstet. Gynecol., 138:720, 1980.
7. Urban, M.D., P.A. Lee and C.J. Migeon: Adult high infertility in men with congenital adrenalized hyperplasia. N. Engl. J. Med. 299:1392, 1978.
8. Meikle, A.W., R.J. Worley and C.D. West: Adrenal corticoid hyper-response in hirsute women. Fertil. Steril. 41:575, 1984
9. Ueshiba, H., Zerah M., New M. I.: Enzyme-linked Immunosorbent assay (ELISA). Method for screening of non-classical steroid 21-Hydroxylase deficiency. Norm. Metab. Res. 26:43, 1994