

## 1 INTRODUCTION

### 1.1 Intended Use

The **DRG Androstenedione ELISA** is an enzyme immunoassay for the quantitative *in vitro diagnostic* measurement of Androstenedione in serum and EDTA plasma

### 1.2 Summary and Explanation

The steroid hormone Androstenedione is one of the main androgens, besides Testosterone and Dehydroepiandrosterone. Testosterone, the most important biological active androgen, is derived from peripheral enzymatic conversion of Androstenedione.

In males, androgens are secreted primarily by the Leydig cells of the testes, to some degree also in the adrenal cortex. In females, the androgens are secreted mainly in the adrenal glands and in the ovary. Around 10% of the androgens are derived from peripheral conversion, mainly of DHEA. Androstenedione and Testosterone show high diurnal variability. The highest levels are measured in the morning. At the age of puberty serum androstenedione levels rise, after menopause they decline again. High androstenedione levels are measured during pregnancy.

In women, high levels of androstenedione (47-100% above normal) are generally found in hirsutism, mostly in combination with other androgens as testosterone and DHEA-S. Androstenedione overproduction is due to ovarian dysfunction or maybe of adrenal origin. High circulating androstenedione levels are found in women with polycystic ovaries and 21-hydroxylase effect. Significant lower androstenedione levels are found in postmenopausal osteoporosis.

## 2 PRINCIPLE OF THE TEST

The DRG Androstenedione ELISA Kit is a solid phase enzyme-linked immunosorbent assay (ELISA), based on the principle of competitive binding.

The microtiter wells are coated with a polyclonal [rabbit] antibody directed towards an antigenic site of the Androstenedione molecule. Endogenous Androstenedione of a patient sample competes with an Androstenedione-horseradish peroxidase conjugate for binding to the coated antibody. After incubation the unbound conjugate is washed off.

The amount of bound peroxidase conjugate is inversely proportional to the concentration of Androstenedione in the sample. After addition of the substrate solution, the intensity of colour developed is inversely proportional to the concentration of Androstenedione in the patient sample.

## 3 WARNINGS AND PRECAUTIONS

1. This kit is for in vitro diagnostic use only. For professional use only.
2. All reagents of this test kit which contain human serum or plasma have been tested and confirmed negative for HIV I/II, HBsAg and HCV by FDA approved procedures. All reagents, however, should be treated as potential biohazards in use and for disposal.
3. Before starting the assay, read the instructions completely and carefully. Use the valid version of the package insert provided with the kit. Be sure that everything is understood.
4. The microplate contains snap-off strips. Unused wells must be stored at 2 °C to 8 °C in the sealed foil pouch and used in the frame provided.
5. Pipetting of samples and reagents must be done as quickly as possible and in the same sequence for each step.



Revised 24 Mar. 2010 rm (Vers. 6.1)



6. Use reservoirs only for single reagents. This especially applies to the substrate reservoirs. Using a reservoir for dispensing a substrate solution that had previously been used for the conjugate solution may turn solution colored. Do not pour reagents back into vials as reagent contamination may occur.
7. Mix the contents of the microplate wells thoroughly to ensure good test results. Do not reuse microwells.
8. Do not let wells dry during assay; add reagents immediately after completing the rinsing steps.
9. Allow the reagents to reach room temperature (21-26°C) before starting the test. Temperature will affect the absorbance readings of the assay. However, values for the patient samples will not be affected.
10. Never pipet by mouth and avoid contact of reagents and specimens with skin and mucous membranes.
11. Do not smoke, eat, drink or apply cosmetics in areas where specimens or kit reagents are handled.
12. Wear disposable latex gloves when handling specimens and reagents. Microbial contamination of reagents or specimens may give false results.
13. Handling should be done in accordance with the procedures defined by an appropriate national biohazard safety guideline or regulation.
14. Do not use reagents beyond expiry date as shown on the kit labels.
15. All indicated volumes have to be performed according to the protocol. Optimal test results are only obtained when using calibrated pipettes and microtiterplate readers.
16. Do not mix or use components from kits with different lot numbers. It is advised not to exchange wells of different plates even of the same lot. The kits may have been shipped or stored under different conditions and the binding characteristics of the plates may result slightly different.
17. Avoid contact with *Stop Solution* containing 0.5 M H<sub>2</sub>SO<sub>4</sub>. It may cause skin irritation and burns.
18. Some reagents contain Proclin 300, BND and/or MIT as preservatives. In case of contact with eyes or skin, flush immediately with water.
19. TMB substrate has an irritant effect on skin and mucosa. In case of possible contact, wash eyes with an abundant volume of water and skin with soap and abundant water. Wash contaminated objects before reusing them. If inhaled, take the person to open air.
20. Chemicals and prepared or used reagents have to be treated as hazardous waste according to the national biohazard safety guideline or regulation.
21. For information on hazardous substances included in the kit please refer to Material Safety Data Sheets. Material Safety Data Sheets for this product are available upon request directly from DRG.



Revised 24 Mar. 2010 rm (Vers. 6.1)



#### 4 REAGENTS

##### 4.1 Reagents provided

1. **Microtiterwells**, 12x8 (break apart) strips, 96 wells;  
Wells coated with anti-Androstenedione antibody (polyclonal).
2. **Standard (Standard 0-5)**, 6 vials, 1 mL, ready to use  
Concentrations: 0 – 0.1 – 0.3 – 1.0 – 3.0 – 10 ng/mL  
Conversion: ng/mL x 3.492 = nmol/L,  
Contain non-mercury preservative.
3. **Enzyme Conjugate**, 1 vial, 25 mL, ready to use,  
Androstenedione conjugated to horseradish peroxidase;  
Contains non-mercury preservative.
4. **Substrate Solution**, 1 vial, 25 mL, ready to use,  
Tetramethylbenzidine (TMB).
5. **Stop Solution**, 1 vial, 14 mL, ready to use,  
contains 0.5M H<sub>2</sub>SO<sub>4</sub>,  
Avoid contact with the stop solution. It may cause skin irritations and burns.
6. **Wash Solution**, 1 vial, 30 mL (40X concentrated),  
see „Preparation of Reagents“.

**Note:** Additional *Standard 0* for sample dilution is available upon request.

##### 4.2 Materials required but not provided

- A microtiter plate calibrated reader (450 ± 10 nm) (e.g. the DRG Instruments Microtiter Plate Reader).
- Calibrated variable precision micropipettes.
- Absorbent paper.
- Distilled or deionized water
- Timer
- Semi logarithmic graph paper or software for data reduction

##### 4.3 Storage Conditions

When stored at 2 °C to 8 °C unopened reagents will retain reactivity until expiration date. Do not use reagents beyond this date.

Opened reagents must be stored at 2 °C to 8 °C. Microtiter wells must be stored at 2 °C to 8 °C. Once the foil bag has been opened, care should be taken to close it tightly again.

Opened kits retain activity for three months if stored as described above.

##### 4.4 Reagent Preparation

Bring all reagents and required number of strips to room temperature prior to use.

##### **Wash Solution**

Add deionized water to the 40X concentrated Wash Solution.

Dilute 30 mL of concentrated *Wash Solution* with 1170 mL deionized water to a final volume of 1200 mL.

*The diluted Wash Solution is stable for 2 weeks at room temperature.*



Revised 24 Mar. 2010 rm (Vers. 6.1)



#### 4.5 Disposal of the Kit

The disposal of the kit must be made according to the national regulations. Special information for this product is given in the Material Safety Data Sheet.

#### 4.6 Damaged Test Kits

In case of any severe damage to the test kit or components, DRG has to be informed in writing, at the latest, one week after receiving the kit. Severely damaged single components should not be used for a test run. They have to be stored until a final solution has been found. After this, they should be disposed according to the official regulations.

### 5 SPECIMEN COLLECTION AND PREPARATION

Serum or EDTA plasma can be used in this assay.

Do not use Heparin or Citrate plasma. Heparin plasma leads to slightly reduced values. For citrate plasma the results are significant increased.

Do not use haemolytic, icteric or lipaemic specimens.

Please note: Samples containing sodium azide should not be used in the assay.

#### 5.1 Specimen Collection

##### Serum:

Collect blood by venipuncture (e.g. Sarstedt Monovette # 02.1388.001), allow to clot, and separate serum by centrifugation at room temperature. Do not centrifuge before complete clotting has occurred. Patients receiving anticoagulant therapy may require increased clotting time.

##### Plasma:

Whole blood should be collected into centrifuge tubes containing anti coagulant and centrifuged immediately after collection.

(E.g. for EDTA plasma Sarstedt Monovette – red cap - # 02.166.001;

#### 5.2 Specimen Storage and Preparation

Specimens should be capped and may be stored for up to 5 days at 2 °C to 8 °C prior to assaying.

If EDTA plasma is stored at 2 °C to 8 °C, it must be assayed within 48 hours.

Specimens held for a longer time should be frozen only once at -20°C prior to assay. Thawed samples should be inverted several times prior to testing.

#### 5.3 Specimen Dilution

If in an initial assay, a specimen is found to contain more than the highest standard, the specimens can be diluted with *Standard 0* and reassayed as described in Assay Procedure.

For the calculation of the concentrations this dilution factor has to be taken into account.

##### Example:

a) dilution 1:10: 10 µL Serum + 90 µL *Standard 0* (mix thoroughly)

b) dilution 1:100: 10 µL dilution a) 1:10 + 90 µL *Standard 0* (mix thoroughly).



Revised 24 Mar. 2010 rm (Vers. 6.1)



## **6 ASSAY PROCEDURE**

### **6.1 General Remarks**

- All reagents and specimens must be allowed to come to room temperature before use. All reagents must be mixed without foaming.
- Once the test has been started, all steps should be completed without interruption.
- Use new disposal plastic pipette tips for each standard, control or sample in order to avoid cross contamination.
- Absorbance is a function of the incubation time and temperature. Before starting the assay, it is recommended that all reagents are ready, caps removed, all needed wells secured in holder, etc. This will ensure equal elapsed time for each pipetting step without interruption.
- As a general rule the enzymatic reaction is linearly proportional to time and temperature.



Revised 24 Mar. 2010 rm (Vers. 6.1)



## 6.2 Test Procedure

Each run must include a standard curve.

1. Secure the desired number of Microtiter wells in the frame holder.
2. Dispense **20 µL** of each *Standard*, *Control* and **samples** with new disposable tips into appropriate wells.
3. Dispense **200 µL Enzyme Conjugate** into each well.  
Thoroughly mix for 10 seconds. It is important to have a complete mixing in this step.
4. Incubate for **60 minutes** at room temperature.
5. Briskly shake out the contents of the wells.  
Rinse the wells **3times** with diluted *Wash Solution* (400 µL per well). Strike the wells sharply on absorbent paper to remove residual droplets.

### Important note:

The sensitivity and precision of this assay is markedly influenced by the correct performance of the washing procedure!

6. Add **200 µL** of *Substrate Solution* to each well.
7. Incubate for **15 minutes** at room temperature.
8. Stop the enzymatic reaction by adding **100 µL** of *Stop Solution* to each well.
9. Determine the absorbance (OD) of each well at **450 ± 10 nm** with a microtiter plate reader.  
It is recommended that the wells be read **within 10 minutes** after adding the *Stop Solution*.

## 6.3 Calculation of Results

1. Calculate the average absorbance values for each set of standards, controls and patient samples.
2. Using semi-logarithmic graph paper, construct a standard curve by plotting the mean absorbance obtained from each standard against its concentration with absorbance value on the vertical (Y) axis and concentration on the horizontal (X) axis.
3. Using the mean absorbance value for each sample determine the corresponding concentration from the standard curve.
4. Automated method: The results in the IFU have been calculated automatically using a 4 PL (4 Parameter Logistics) curve fit. 4 Parameter Logistics is the preferred method. Other data reduction functions may give slightly different results.
5. The concentration of the samples can be read directly from this standard curve. Samples with concentrations higher than that of the highest standard have to be further diluted or reported as > 10.0 ng/mL. For the calculation of the concentrations this dilution factor has to be taken into account.



Revised 24 Mar. 2010 rm (Vers. 6.1)

**6.3.1 Example of Typical Standard Curve**

The following data is for demonstration only and **cannot** be used in place of data generations at the time of assay.

Standard	Optical Units (450 nm)
Standard 0 (0 ng/mL)	2.01
Standard 1 (0.1 ng/mL)	1.34
Standard 2 (0.3 ng/mL)	0.86
Standard 3 (1.0 ng/mL)	0.47
Standard 4 (3.0 ng/mL)	0.23
Standard 5 (10.0 ng/mL)	0.10

**7 EXPECTED NORMAL VALUES**

It is strongly recommended that each laboratory should determine its own normal and abnormal values.

In a study conducted with apparently normal healthy adults, using the DRG Androstenedione ELISA the following values are observed:

Population	(ng/mL)
Males	0.91 – 3.0
Females	0.57 – 2.63

The results alone should not be the only reason for any therapeutic consequences. The results should be correlated to other clinical observations and diagnostic tests.

**8 QUALITY CONTROL**

Good laboratory practice requires that controls be run with each calibration curve. A statistically significant number of controls should be assayed to establish mean values and acceptable ranges to assure proper performance.

It is recommended to use control samples according to state and federal regulations. The use of control samples is advised to assure the day to day validity of results. Use controls at both normal and pathological levels.

The controls and the corresponding results of the QC-Laboratory are stated in the QC certificate added to the kit. The values and ranges stated on the QC sheet always refer to the current kit lot and should be used for direct comparison of the results.

It is also recommended to make use of national or international Quality Assessment programs in order to ensure the accuracy of the results.

Employ appropriate statistical methods for analysing control values and trends. If the results of the assay do not fit to the established acceptable ranges of control materials patient results should be considered invalid.

In this case, please check the following technical areas: Pipetting and timing devices; photometer, expiration dates of reagents, storage and incubation conditions, aspiration and washing methods.

After checking the above mentioned items without finding any error contact your distributor or DRG directly.



Revised 24 Mar. 2010 rm (Vers. 6.1)



## 9 PERFORMANCE CHARACTERISTICS

### 9.1 Assay Dynamic Range

The range of the assay is between 0.019 – 10 ng/mL.

### 9.2 Specificity of Antibodies (Cross Reactivity)

The following substances were tested for cross reactivity of the assay:

<u>Compound</u>	<u>Crossreactivity %</u>
Androstenedione	100
Androsterone	< 0.01
Cortisol	< 0.01
Dihydrotestosterone	< 0.01
Dihydroepiandrosterone	0.01
Estriol	< 0.01
16-Epiestriol	< 0.01
Estradiol	< 0.01
Estriol-3-glucuronide	< 0.01
Estriol-16-glucuronide	< 0.01
Estriol-16-sulfate	< 0.01
Estrone	< 0.01
17a-Pregnenolone	< 0.01
17a-Progesterone	< 0.01
Progesterone	< 0.01
Testosterone	0.01

### 9.3 Sensitivity

The analytical sensitivity of the DRG ELISA was calculated by subtracting 2 standard deviations from the mean of 20 replicate analyses of the Standard 0 (S0) and was found to be 0.019 ng/mL.

### 9.4 Reproducibility

#### 9.4.1 Intra Assay

The within assay variability is shown below:

Sample	n	Mean (ng/mL)	CV (%)
1	20	0.3	9.1
2	20	2.6	5.6
3	20	4.7	4.7





Revised 24 Mar. 2010 rm (Vers. 6.1)

**9.4.2 Inter Assay**

The between assay variability is shown below:

Sample	n	Mean (ng/mL)	CV (%)
1	12	0.2	9.6
2	12	2.3	12.1
3	12	4.4	8.8

**9.5 Recovery**

Samples have been spiked by adding Androstenedione solutions with known concentrations in a 1:1 ratio.

The expected values were calculated by addition of half of the values determined for the undiluted samples and half of the values of the known solutions. The % Recovery has been calculated by multiplication of the ratio of the measurements and the expected values with 100.

Sample	Added Concentration 1:1 (v/v) (ng/mL)	Measured Conc. (ng/mL)	Expected Conc. (ng/mL)	Recovery (%)
1	-	0.1	0.1	100
	1	0.5	0.6	89
	3	1.5	1.6	96
	10	4.7	5.1	93
2	-	2.0	2.0	100
	1	1.5	1.4	107
	3	2.4	2.4	99
	10	5.0	6.0	92
3	-	5.4	5.4	100
	1	3.4	3.2	106
	3	4.4	4.2	105
	10	7.3	7.7	95



Revised 24 Mar. 2010 rm (Vers. 6.1)

**9.6 Linearity**

Sample	Dilution	Measured Conc. (ng/mL)	Recovery (%)
1	None	0.443	-
	1:2	0.231	104
	1:4	0.105	95
	1:8	0.050	90
	1:16	0.028	101
2	None	1.258	-
	1:2	0.577	92
	1:4	0.292	93
	1:8	0.146	93
	1:16	0.074	94
3	None	4.471	-
	1:2	2.250	101
	1:4	1.124	101
	1:8	0.580	104
	1:16	0.281	101

**10 LIMITATIONS OF USE**

Reliable and reproducible results will be obtained when the assay procedure is performed with a complete understanding of the package insert instruction and with adherence to good laboratory practice.

Any improper handling of samples or modification of this test might influence the results.

**10.1 Interfering Substances**

Haemoglobin (up to 4 mg/mL), Bilirubin (up to 0.125 mg/mL) and Triglyceride (up to 7.5 mg/mL) have no influence on the assay results.

**10.2 Drug Interferences**

Until today no substances (drugs) are known to us, which have an influence to the measurement of Androstenedione in a sample.

**10.3 High-Dose-Hook Effect**

No hook effect was observed in this test.

**11 LEGAL ASPECTS****11.1 Reliability of Results**

The test must be performed exactly as per the manufacturer's instructions for use. Moreover the user must strictly adhere to the rules of GLP (Good Laboratory Practice) or other applicable national standards and/or laws. This is especially

**Revised 24 Mar. 2010 rm (Vers. 6.1)**

relevant for the use of control reagents. It is important to always include, within the test procedure, a sufficient number of controls for validating the accuracy and precision of the test.

The test results are valid only if all controls are within the specified ranges and if all other test parameters are also within the given assay specifications. In case of any doubt or concern please contact DRG.

### **11.2 Therapeutic Consequences**

Therapeutic consequences should never be based on laboratory results alone even if all test results are in agreement with the items as stated under point 11.1. Any laboratory result is only a part of the total clinical picture of a patient.

Only in cases where the laboratory results are in acceptable agreement with the overall clinical picture of the patient should therapeutic consequences be derived.

The test result itself should never be the sole determinant for deriving any therapeutic consequences.

### **11.3 Liability**

Any modification of the test kit and/or exchange or mixture of any components of different lots from one test kit to another could negatively affect the intended results and validity of the overall test. Such modification and/or exchanges invalidate any claim for replacement.

Claims submitted due to customer misinterpretation of laboratory results subject to point 11.2. are also invalid.

Regardless, in the event of any claim, the manufacturer's liability is not to exceed the value of the test kit. Any damage caused to the test kit during transportation is not subject to the liability of the manufacturer.

## **12 REFERENCES / LITERATURE**

1. Kicman, A. T., Bassindale, T., Cowan, D. A., Dale, S., Hutt, A. J., and Leeds, A. R. Effect of androstenedione ingestion on plasma testosterone in young women; a dietary supplement with potential health risks. Clin Chem 2003, 49:167-169.
2. Brown, G.A., Vukovich, M.D., Martini, E.R., Kohut, M.L., Franke, W.D., Jackson, D.A., and King, D.S. Endocrine responses to chronic androstenedione intake in 30- to 56-year-old men. J Clin Endocrinol Metab 2000, 85:4074-4080.
3. Erickson GF 1993 Normal regulation of ovarian androgen production. Seminars in Reproductive Endocrinology 11:307-312
4. Mango D, Scirpa P, Battaglia F, Tartaglia E, Manna P. Diagnostic significance of steroid hormones in patients with ovarian cancer. J Endocrinol Invest. 1986 Aug;9(4):307-14