



Revised 24 Feb. 2010 rm (Vers. 2.0)

**RUO** in the USA*Please use only the valid version of the package insert provided with the kit.***NAME AND INTENDED USE**

Anti-RNP 70 is an indirect solid phase enzyme immunoassay (ELISA) for the quantitative measurement of IgG class autoantibodies against RNP 70 in human serum or plasma. The assay is intended for in vitro use only as an aid in the determination of Systemic Lupus Erythematosus (SLE) and Mixed connective tissue disease (MCTD).

**SUMMARY AND EXPLANATION OF THE TEST**

Overlap syndrome is a generic term coined to combine several inflammatory rheumatic diseases characterized by overlapping symptoms of systemic lupus erythematosus (SLE), progressive systemic scleroderma (PSS), Dermato/Polymyositis (DM/PM), rheumatoid arthritis (RA), or Sjögren's syndrome.

Mixed connective tissue disease (MCTD) is a serologically well defined overlap syndrome characterized by typical clinical manifestations, distinct autoantibodies specificities and characteristic HLA constellation (i.e. DR2 and DR4). Outstanding clinical manifestation of patients suffering from MCTD is Raynaud's phenomenon with swollen fingers and hands. Additionally, the patients exhibit arthritic complaints. During the course of disease different symptoms of the distinct overlapping connective tissue diseases often emerge in patients suffering from MCTD. However, not all MCTD patients suffer those transitions.

Serologically MCTD is characterized by the presence of autoantibodies directed against the 70 kDa protein (also named 60 kDa protein) of the U1-snRNP-Komplex. Up to 100% of MCTD patients manifest high titers of Anti-RNP-70 antibodies. Patients exhibiting only Anti-RNP-70 among the various antinuclear autoantibodies show the classical clinical manifestations of mixed connective tissue disease. Additionally, often SLE typical autoantibodies (e.g. Anti-dsDNA, Anti-Sm, Anti-Histone) come along in the sera of those patients. In these cases classical MCTD will proceed to SLE.

The U1-snRNP molecule (its function is splicing of pre-messenger RNA) consists of a uridine rich RNA complexed with the classical RNP antigens, i.e. 70 kDa protein, protein A (33 kDa) and protein C (22 kDa). Also the Sm specific proteins are associated with this RNA. Anti-Sm antibodies are pathognomonic markers for SLE. In the Anti-RNP-70 ELISA the wells are exclusively coated with the 70 kDa protein of the U1-snRNP complex. Accordingly, this assay provides the highly specific detection of mainly MCTD associated Anti-RNP-70 autoantibodies.

**PRINCIPLE OF THE TEST**

Recombinant RNP 70 is bound to microwells. Antibodies against this antigen, if present in diluted serum or plasma, bind to the respective antigen. Washing of the microwells removes unspecific serum and plasma components. Horseradish peroxidase (HRP) conjugated antihuman IgG immunologically detects the bound patient antibodies forming a conjugate/antibody/antigen complex. Washing of the microwells removes unbound conjugate. An enzyme substrate in the presence of bound conjugate hydrolyzes to form a blue color. The addition of an acid stops the reaction forming a yellow end-product. The intensity of this yellow color is measured photometrically at 450 nm. The amount of colour is directly proportional to the concentration of IgG antibodies present in the original sample.

## DRG® Anti-RNP 70 (EIA-3579)



Revised 24 Feb. 2010 rm (Vers. 2.0)

**RUO** in the USA

### WARNINGS AND PRECAUTIONS

1. All reagents of this kit are strictly intended for in vitro use only.
2. Do not interchange kit components from different lots.
3. Components containing human serum were tested and found negative for HBsAg, HCV, HIV1 and HIV2 by FDA approved methods. No test can guarantee the absence of HBsAg, HCV, HIV1 and HIV2, and so all human serum based reagents in this kit must be handled as though capable of transmitting infection.
4. Avoid contact with the TMB (3,3',5,5'-Tetramethyl-benzidine). If TMB comes into contact with skin, wash thoroughly with water and soap.
5. Avoid contact with the Stop Solution which is acid. If it comes into contact with skin, wash thoroughly with water and seek medical attention.
6. Some kit components (i.e. Controls, Sample buffer and Buffered Wash Solution) contain Sodium Azide as preservative. Sodium Azide ( $\text{NaN}_3$ ) is highly toxic and reactive in pure form. At the product concentrations (0.09%), though not hazardous. Despite the classification as non-hazardous, we strongly recommend using prudent laboratory practices (see 8., 9., 10.).
7. Some kit components contain Proclin 300 as preservative. When disposing reagents containing Proclin 300, flush drains with copious amounts of water to dilute the components below active levels.
8. Wear disposable gloves while handling specimens or kit reagents and wash hands thoroughly afterwards.
9. Do not pipette by mouth.
10. Do not eat, drink, smoke or apply makeup in areas where specimens or kit reagents are handled.
11. Avoid contact between the buffered Peroxide Solution and easily oxidized materials; extreme temperature may initiate spontaneous combustion.

Observe the guidelines for performing quality control in medical laboratories by assaying controls and/or pooled sera. During handling of all kit reagents, controls and serum samples observe the existing legal regulations.

### CONTENTS OF THE KIT

Package size	96 determ.
Qty.1	Divisible <b>microplate</b> consisting of 12 modules of 8 wells each, coated with recombinant RNP 70. Ready to use.
6 vials, 1.5 ml each	Anti-RNP 70 <b>Calibrators</b> (A-F) in a serum/buffer matrix (PBS, BSA, $\text{NaN}_3$ <0,1% (w/w)) containing: IgG 0; 12.5; 25; 50; 100; and 200 U/ml.. Ready to use.
2 vials, 1.5 ml each	Anti- RNP 70 <b>Controls</b> in a serum/buffer matrix (PBS, BSA, $\text{NaN}_3$ <0.1% (w/w)) positive (1) and negative (2), for the respective concentrations see the enclosed QC insert. Ready to use.
1 vial, 20 ml	<b>Sample buffer</b> (Tris, $\text{NaN}_3$ <0.1% (w/w)), yellow, concentrate (5x).

## DRG® Anti-RNP 70 (EIA-3579)



Revised 24 Feb. 2010 rm (Vers. 2.0)

**RUO** in the USA

1 vial, 15 ml	<b>Enzyme conjugate</b> solution (PBS, Proclin 300 <0.5% (v/v)), (light red) containing polyclonal rabbit anti-human IgG, labelled with horseradish peroxidase. Ready to use.
1 vial, 15 ml	<b>TMB substrate solution.</b> Ready to use.
1 vial, 15 ml	<b>Stop solution</b> (contains acid). Ready to use.
1 vial, 20 ml	<b>Wash solution</b> (PBS, NaN <sub>3</sub> <0.1% (w/w)), concentrate (50x).

### STORAGE AND STABILITY

1. Store the kit at 2-8 °C.
2. Keep microplate wells sealed in a dry bag with desiccants.
3. The reagents are stable until expiration of the kit.
4. Do not expose test reagents to heat, sun or strong light during storage and usage.
5. Diluted sample buffer and wash buffer are stable for at least 30 days when stored at 2-8 °C.

### MATERIALS REQUIRED

#### Equipment

- Microplate reader capable of endpoint measurements at 450 nm
- Multi-Channel Dispenser or repeatable pipet for 100 µl
- Vortex mixer
- Pipets for 10 µl, 100 µl and 1000 µl
- Laboratory timing device
- Data reduction software

#### Preparation of reagents

- Distilled or deionized water
- Graduated cylinder for 100 and 1000 ml
- Plastic container for storage of the wash solution

### SPECIMEN COLLECTION, STORAGE AND HANDLING

1. Collect whole blood specimens using acceptable medical techniques to avoid hemolysis.
2. Allow blood to clot and separate the serum by centrifugation.
3. Test serum should be clear and non-hemolyzed. Contamination by hemolysis or lipemia is best avoided, but does not interfere with this assay.
4. Specimens may be refrigerated at 2-8 °C for up to five days or stored at -20 °C up to six months.
5. Avoid repetitive freezing and thawing of serum samples. This may result in variable loss of autoantibody activity.
6. Testing of heat-inactivated sera is not recommended.



Revised 24 Feb. 2010 rm (Vers. 2.0)

**RUO** in the USA**PROCEDURAL NOTES**

1. Do not use kit components beyond their expiration dates.
2. Do not interchange kit components from different lots.
3. All materials must be at room temperature (20-28 °C).
4. Have all reagents and samples ready before start of the assay. Once started, the test must be performed without interruption to get the most reliable and consistent results.
5. Perform the assay steps only in the order indicated.
6. Always use fresh sample dilutions.
7. Pipette all reagents and samples into the bottom of the wells.
8. To avoid carryover contamination change the tip between samples and different kit controls.
9. It is important to wash microwells thoroughly and remove the last droplets of wash buffer to achieve best results.
10. All incubation steps must be accurately timed.
11. Control sera or pools should routinely be assayed as unknowns to check performance of the reagents and the assay.
12. Do not re-use microplate wells.

For all controls, the respective concentrations are provided on the labels of each vial. Using these concentrations a calibration curve may be calculated to read off the patient results semi-quantitatively.

**PREPARATION OF REAGENTS***Preparation of sample buffer*

Dilute the contents of each vial of the sample buffer concentrate (5x) with distilled or deionized water to a final volume of 100 ml prior to use.

Store refrigerated: stable at 2-8 °C for at least 30 days after preparation or until the expiration date printed on the label.

*Preparation of wash solution*

Dilute the contents of each vial of the buffered wash solution concentrate (50x) with distilled or deionized water to a final volume of 1000 ml prior to use.

Store refrigerated: stable at 2-8 °C for at least 30 days after preparation or until the expiration date printed on the label.

*Sample preparation*

Dilute all patient samples **1:100** with sample buffer before assay.

Therefore combine 10 µl of sample with 990 µl of sample buffer in a polystyrene tube. Mix well.

Controls are ready to use and need not be diluted.

**TEST PROCEDURE**

1. Prepare a sufficient number of microplate modules to accommodate controls and prediluted patient samples.
2. Pipet **100 µl** of calibrators, controls and prediluted patient samples in duplicate into the wells.



Revised 24 Feb. 2010 rm (Vers. 2.0)

**RUO** in the USA

	1	2	3	4	5	6
<b>A</b>	SA	SE	P1	P5		
<b>B</b>	SA	SE	P1	P5		
<b>C</b>	SB	SF	P2	P..		
<b>D</b>	SB	SF	P2	P..		
<b>E</b>	SC	C1	P3			
<b>F</b>	SC	C1	P3			
<b>G</b>	SD	C2	P4			
<b>H</b>	SD	C2	P4			

**SA-SF:** standards A to F  
**P1, P2...:** patient sample 1, 2 ...  
**C1:** positive control  
**C2:** negative control

- Incubate for 30 minutes at room temperature (20-28 °C).
- Discard the contents of the microwells and wash 3 times with **300 µl** of wash solution.
- Dispense **100 µl** of enzyme conjugate into each well.
- Incubate for 15 minutes at room temperature.
- Discard the contents of the microwells and wash 3 times with **300 µl** of wash solution.
- Dispense **100 µl** of TMB substrate solution into each well.
- Incubate for 15 minutes at room temperature.
- Add **100 µl** of stop solution to each well of the modules and incubate for 5 minutes at room temperature.
- Read the optical density at 450 nm and calculate the results. Bi-chromatic measurement with a reference at 600-690 nm is recommended.

The developed colour is stable for at least 30 minutes. Read optical densities during this time.

#### Automation

The Anti-RNP 70 ELISA is suitable for use on open automated ELISA processors. The test procedure detailed above is appropriate for use with or without automation.

## INTERPRETATION OF RESULTS

#### Quality Control

This test is only valid if the optical density at 450 nm for Positive Control (1) and Negative Control (2) as well as for the Calibrator A and F complies with the respective range indicated on the Quality Control Certificate enclosed to each test kit !

If any of these criteria is not fulfilled, the results are invalid and the test should be repeated.

#### Calculation of results

For Anti-RNP70 IgG a 4-Parameter-Fit with lin-log coordinates for optical density and concentration is the data reduction method of choice.



**Revised 24 Feb. 2010 rm (Vers. 2.0)**

**RUO in the USA**

### **Recommended Lin-Log Plot**

First calculate the averaged optical densities for each calibrator well. Use lin-log graph paper and plot the averaged optical density of each calibrator versus the concentration. Draw the best fitting curve approximating the path of all calibrator points. The calibrator points may also be connected with straight line segments. The concentration of unknowns may then be estimated from the calibration curve by interpolation.

### *Interpretation of results*

In a normal range study with serum samples from healthy blood donors the following ranges have been established with the Anti-RNP 70 test:

<b>Anti-RNP 70 [U/ml]</b>	
normal:	< 25
elevated:	≥ 25

Positive results should be verified concerning the entire clinical status of the patient. Also every decision for therapy should be taken individually. It is recommended that each laboratory establishes its own normal and pathological ranges of serum Anti-RNP 70.

## **PERFORMANCE CHARACTERISTICS**

### *Parallelism*

In dilution experiments sera with high IgG-antibody concentrations were diluted with sample buffer and assayed in the Anti-RNP 70 kit. The assay shows linearity over the full measuring range.

### *Precision (Reproducibility)*

Statistics for coefficients of variation (CV) were calculated for each of three samples from the results of 24 determinations in a single run for Intra-Assay precision. Run-to-run precision was calculated from the results of 5 different runs with 6 determinations of each sample:

Intra-Assay		
Sample No	Mean [U/ml]	CV [%]
1	12.8	5.8
2	105.0	3.9
3	182.0	5.7

Intra-Assay		
Sample No	Mean [U/ml]	CV [%]
1	14.4	4.9
2	108.7	5.2
3	175.2	5.2

### *Sensitivity*

The lower detection limit for Anti-RNP 70 has been determined at 1.0 U/ml.

### *Specificity*

The microplate is coated with recombinant RNP 70 as antigen. The Anti-RNP 70 test kits recognises only autoantibodies specific to the concerning antigen.

**DRG<sup>®</sup> Anti-RNP 70 (EIA-3579)****Revised 24 Feb. 2010 rm (Vers. 2.0)****RUO in the USA**

No cross reactivities to other autoantigens have been observed.

*Calibration*

Since no international reference preparation for Anti-RNP 70 autoantibodies is available, the assay system is calibrated in relative arbitrary units..

**LIMITATIONS OF PROCEDURE**

The Anti-RNP 70 IgG ELISA is a diagnostic aid. A definite clinical diagnosis should not be based on the results of a single test, but should be made by the physician after all clinical and laboratory findings have been evaluated.

A negative Anti-RNP 70 result does not rule out the presence of SLE or MCTD.

**INTERFERING SUBSTANCES**

No interference has been observed with haemolytic (up to 1000 mg/dL), lipemic (up to 3 g/dL triglycerides) or bilirubin (up to 40 mg/dL) containing sera. Nor have any interfering effects been observed with the use of anticoagulants. However for practical reasons it is recommended that grossly hemolyzed or lipemic samples should be avoided.

**REFERENCES / Literature**

1. Antinuclear Antibodies The Lancet 1984; 15: 611 - 613.
2. Harley, J. B. and Gaither, K. K. Autoantibodies.  
Rheum. Disease Clin. of North America 1988; Vol 14, No 1: 43 - 56.
3. Hartung, K. and Deicher, H. Systemischer Lupus erythematoses. Allergologie 1985; (8) Nr. 7: 275 - 281.
4. Hartung, K. and Deicher, H. Symptome und Symptommuster beim systemischen Lupus erythematoses.  
Die Internistische Welt 1987; 9: 236 - 241.
5. Takeda, J. et al. Antigenic Domains on the U1 Small Nuclear Ribonucleoprotein-Associated 70K Polypeptide: A Comparison of Regions Selectively Recognized by Human and Mouse Autoantibodies and by Monoclonal Antibodies.  
Clin. Immunol. Immunopathol 1991; 61: 55 - 68.
6. Reichlin, M. and Van Venrooij, W. J. Autoantibodies to the URNP particles: relationship to clinical diagnosis and nephritis.  
Clin. Exp. Immunol. 1991; 83: 286 - 290.
7. Hietarinta, M., Lassila, O. and Hietaharju, A. Association of Anti-U1RNP- and Anti-Scl-70-antibodies with Neurological Manifestations in Systemic Sclerosis (Scleroderma).  
Scand. J. Rheumatol. 1994; 23: 64 - 67.











# DRG® Anti-RNP 70 (EIA-3579)













Revised 24 Feb. 2010 rm (Vers. 2.0)

**RUO** in the USA

## Symbols used with DRG Assays

Symbol	English	Deutsch	Français	Español	Italiano
	Consult instructions for use	Gebrauchsanweisung beachten	Consulter les instructions d'utilisation	Consulte las instrucciones de uso	Consultare le istruzioni per l'uso
	European Conformity	CE-Konformitätskennzeichnung	Conformité aux normes européennes	Conformidad europea	Conformità europea
	In vitro diagnostic device	In-vitro-Diagnostikum	Usage Diagnostic in vitro	Para uso Diagnóstico in vitro	Per uso Diagnostica in vitro
	For research use only	Nur für Forschungszwecke	Seulement dans le cadre de recherches	Sólo para uso en investigación	Solo a scopo di ricerca
	Catalogue number	Katalog-Nr.	Numéro de catalogue	Número de catálogo	Numero di Catalogo
	Lot. No. / Batch code	Chargen-Nr.	Numéro de lot	Número de lote	Numero di lotto
	Contains sufficient for <n> tests/	Ausreichend für "n" Ansätze	Contenu suffisant pour "n" tests	Contenido suficiente para <n> ensayos	Contenuto sufficiente per "n" saggi
	Storage Temperature	Lagerungstemperatur	Température de conservation	Temperatura de conservación	Temperatura di conservazione
	Expiration Date	Mindesthaltbarkeits-datum	Date limite d'utilisation	Fecha de caducidad	Data di scadenza
	Legal Manufacturer	Hersteller	Fabricant	Fabricante	Fabbricante
Distributed by	Distributor	Vertreiber	Distributeur	Distribuidor	Distributore
Content	Content	Inhalt	Conditionnement	Contenido	Contenuto
Volume/No.	Volume / No.	Volumen/Anzahl	Volume/Quantité	Volumen/Número	Volume/Quantità

Symbol	Portugues	Dansk	Svenska	Ελληνικά
	Consulte as instruções de utilização	Se brugsanvisning	Se bruksanvisningen	Εγχειρίδιο χρήστη
	Conformidade com as normas europeias	Europaeisk overensstemmelse	Europeisk överensstämmelse	Ευρωπαϊκή Συμμόρφωση
	Diagnóstico in vitro	In vitro diagnostik	Diagnostik in vitro	in vitro διαγνωστικό
				
	Catálogo n.º	Katalognummer	Katalog nummer	Αριθμός καταλόγου
	No do lote	Lot nummer	Batch-nummer	Αριθμός Παρτίδος
		Indeholder tilstrækkeligt til "n" test	Innehåller tillräckligt till "n" tester	Περιεχόμενο επαρκές για «n» εξετάσεις
	Temperatura de conservação	Opbevarings-temperatur	Förvaringstempratur	Θερμοκρασία αποθήκευσης
	Prazo de validade	Udløbsdato	Bäst före datum	Ημερομηνία λήξης
	Fabricante	Producent	Tillverkare	Κατασκευαστής
Content	Conteúdo	Indhold	Innehåll	Περιεχόμενο
Volume/No.	Volume/Número	Volumen/antal	Volym/antal	Όγκος/αριθ..