



REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

Please use only the valid version of the package insert provided with the kit.

1 INTENDED USE

The Resistin (human) ELISA is a sandwich enzyme immunoassay for the quantitative measurement of human resistin.

Features

- European Union: for in vitro diagnostic use.
- Rest of the world: for research use only!
- The total assay time is less than 4 hours.
- The kit measures total resistin in serum and plasma (heparin, citrate or EDTA),.
- Assay format is 96 wells.
- Quality Controls are human serum based. No animal sera are used.
- Standard is recombinant protein based.
- Components of the kit are provided ready to use, concentrated or lyophilized.

2 STORAGE, EXPIRATION

Store the complete kit at 2-8°C. Under these conditions, the kit is stable until the expiration date (see label on the box).

For stability of opened reagents see Chapter 9.

3 INTRODUCTION

Resistin, a product of the RSTN gene, is a peptide hormone belonging to the class of cysteine-rich secreted proteins which is termed the RELM family, and is also described as ADSF (Adipose Tissue-Specific Secretory Factor) and FIZZ3 (Found in Inflammatory Zone). Human resistin contains 108 amino acids as a prepeptide, and its hydrofobic signal peptide is cleaved before its secretion. Resistin circulates in human blood as a dimeric protein consisting of two 92 amino acid polypeptides, which are disulfide-linked via Cys26.

Resistin may be an important link between obesity and insulin resistance. Mouse resistin, specifically produced and secreted by adipocyte, acts on skeletal muscle myocytes, hepatocytes and adipocytes themselves so that it reduces their sensitivity to insulin. Steppan et al. have suggested that resistin suppresses the ability of insulin to stimulate glucose uptake. They have also suggested that resistin is present at elevated levels in blood of obese mice, and is down regulated by fasting and antidiabetic drugs. Way et al., on the other hand, have found that resistin expression is severly suppressed in obesity and is stimulated by several antidiabetic drugs.





2

DRG[®] Resistin (human) ELISA (EIA-4572)

REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

Other studies have shown that mouse resistin increases during the differentiation of adipocytes, but it also seems to inhibit adipogenesis. In contrast, the human adipogenic differentiation is likely to be associated with a down regulation of resistin gene expression. Recent studies have shown that human resistin is expressed also in macrophages and may be a novel link between inflammation and insulin resistance.

Areas of investigation:

Energy metabolism and body weight regulation

4 TEST PRINCIPLE

In the DRG Resistin (human) ELISA, Standards, Quality Controls and samples are incubated in microplate wells precoated with polyclonal anti-human resistin antibody. After 60 minutes incubation and washing, biotin-labelled second polyclonal anti-human resistin antibody is added and incubated with captured resistin for 60 minutes. After another washing, streptavidin-HRP conjugate is added. After 60 minutes incubation and the last washing step, the remaining conjugate is allowed to react with the substrate solution (TMB). The reaction is stopped by addition of acidic solution and absorbance of the resulting yellow product is measured. The absorbance is proportional to the concentration of resistin.

A standard curve is constructed by plotting absorbance values against concentrations of Standards, and concentrations of unknown samples are determined using this standard curve.

5 PRECAUTIONS

For professional use only.

- Wear gloves and laboratory coats when handling immunodiagnostic materials.
- Do not drink, eat or smoke in the areas where immunodiagnostic materials are being handled.
- This kit contains components of human origin. These materials were found nonreactive for HBsAg, HCV antibody and for HIV 1/2 antigen and antibody. However, these materials should be handled as potentially infectious, as no test can guarantee the complete absence of infectious agents.
- 1. This kit contains components of animal origin. These materials should be handled as potentially infectious.
- Avoid contact with the acidic Stop Solution and Substrate Solution, which contains hydrogen peroxide and tetramethylbenzidine (TMB). Wear gloves and eye and clothing protection when handling these reagents. Stop and/or Substrate Solutions may cause skin/eyes irritation. In case of contact with the Stop Solution and the Substrate Solution wash skin/eyes thoroughly with water and seek medical attention, when necessary.
- The materials must not be pipetted by mouth

6 TECHNICAL HINTS

- o Reagents with different lot numbers should not be mixed.
- o Use thoroughly clean glassware.
- o Use deionized (distilled) water, stored in clean containers.

DRG International, Inc. USA Fax: (908) 233 0758 e-mail: corp@drg-international.com





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

- Avoid any contamination among samples and reagents. For this purpose, disposable tips should be used for each sample and reagent.
- Substrate Solution should remain colourless until added to the plate. Keep Substrate Solution protected from light.
- Stop Solution should remain colourless until added to the plate. The colour developed in the wells will turn from blue 0 to yellow immediately after the addition of the Stop Solution. Wells that are green in colour indicate that the Stop Solution has not mixed thoroughly with the Substrate Solution.
- Dispose of consumable materials and unused contents in accordance with applicable national regulatory requirements.

Kit Components State Quantity 96 wells Antibody Coated Microtiter Strips ready to use 13 mL Biotin Labelled Antibody ready to use StreptavidinHRP Conjugate ready to use 13 mL Master Standard lyophilized 1 vial Quality Control HIGH lyophilized 1 vial Quality Control LOW lyophilized 1 vial **Dilution Buffer** ready to use 20 mL Wash Solution Conc. (10x) concentrated 100 mL Substrate Solution ready to use 13 mL Stop Solution 13 mL ready to use Product Data Sheet + Certificate of Analysis 1 pc

REAGENT SUPPLIED 7

8 MATERIAL REQUIRED BUT NOT SUPPLIED

- Deionized (distilled) water 0
- Test tubes for diluting samples 0
- Glassware (graduated cylinder and bottle) for Wash Solution (Dilution Buffer) 0
- Precision pipettes to deliver $(5)10-1000 \mu$ L with disposable tips 0
- Multichannel pipette to deliver 100 μ L with disposable tips 0
- Absorbent material (e.g. paper towels) for blotting the microtitrate plate after washing 0
- Vortex mixer 0
- Orbital microplate shaker capable of approximately 300 rpm 0
- Microplate washer (optional). [Manual washing is possible but not preferable.] 0





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

- \circ Microplate reader with 450 ± 10 nm filter preferably with reference wavelength 630 nm (alternatively another one from the interval 550 650nm)
- Software package facilitating data generation and analysis (optional)

9 PREPARATION OF REAGENTS

All reagents need to be brought to room temperature prior to use. Always prepare only the appropriate quantity of reagents for your test. Do not use components after the expiration date marked on their label.

9.1 Assay reagents supplied ready to use

1. Antibody Coated Microtiter Strips

Stability and storage:

Return the unused strips to the provided aluminium zip-sealed bag with desicant and seal carefully. Remaining Microtiter Strips are stable 3 month stored at 2-8°C and protected from the moisture.

- 2. Biotin Labelled Antibody
- 3. Streptavidin-HRP conjugate
- 4. Dilution Buffer
- 5. Substrate Solution
- 6. Stop Solution

<u>Stability and storage:</u> Opened reagents are stable 3 month when stored at 2-8°C.





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

9.2 Assay reagents supplied concentrated or lyophilized

1. Human Resistin Master Standard

Refer to the Cetrificate of Analysis for current volume of Dilution Buffer needed for reconstitution of standard!!!

Reconstitute the lyophilized Master Standard with Dilution Buffer just prior to the assay. Let it dissolve at least 15 minutes with occasionally gently shaking (not to foam). The resulting concentration of the resistin in the stock solution is **50 ng/mL**.

Volume of Standard	Dilution Buffer	Concentration
stock		50 ng/mL
500 µL of stock	750 μL	20 ng/mL
500 μL of std. 20 ng/mL	500 μL	10 ng/mL
500 μ L of std. 10 ng/mL	500 μL	5 ng/mL
500 μ L of std. 5 ng/mL	750 μL	2 ng/mL
500 μL of std. 2 ng/mL	500 μL	1 ng/mL

Prepare set of standards using Dilution Buffer as follows:

Dilute each concentration of standard 3x with Dilution Buffer prior to the assay, e.g. 50 μ L of standard + 100 μ L of Dilution Buffer for singlets, or preferably 100 μ L of standard + 200 μ L of Dilution Buffer for duplicates. Mix well (not to foam).

Stability and storage:

Set of Standards (50 - 1 ng/mL) should be aliquoted and frozen at -20°C for 3 months.

Avoid repeated freeze/thaw cycles.

Do not store the diluted Standard solutions.

2. Quality Controls High, Low

Reconstitute each Quality Control (High and Low) with **350** μ L of Dilution Buffer just prior to the assay. Let it dissolve at least 30 minutes with occasionally gently shaking (not to foam).

Dilute Quality Controls prior to the assay 3x with Dilution Buffer, e.g. 50 μ L of Control + 100 μ L of Dilution Buffer for singlets, or preferably 100 μ L of Control + 200 μ L of Dilution Buffer for duplicates

Stability and storage:

The reconstituted Quality Controls must be used immediately or aliquoted and frozen at -20°C for 3 month. Avoid repeated freeze/thaw cycles.

Do not store the diluted Quality Controls solutions.





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

3. Wash Solution Conc. (10x)

Dilute Wash Solution Conc. (10x) ten-fold in distilled water to prepare a 1x working solution. <u>Example:</u> 100 mL of Wash Solution Conc. (10x) + 900 mL of distilled water for use of all 96-wells. <u>Stability and storage:</u> The diluted Wash Solution is stable 1 month when stored at 2-8°C. Opened Wash Solution Conc. (10x) is stable 3 months when stored at 2-8°C.

10 PREPARATION OF SAMPLES

The kit measures human resistin (homodimeric) in serum or plasma.

Samples should be assayed immediately after collection or should be stored at -20°C. Mix thoroughly thawed samples just prior to the assay and avoid repeated freeze/thaw cycles, which may cause erroneous results.

Avoid using hemolyzed or lipemic samples.

Dilute samples 3x with Dilution Buffer just prior to the assay (e.g. 50 μ L of sample + 100 μ L of Dilution Buffer for singlets, or preferably 100 μ L of sample + 200 μ L of Dilution Buffer for duplicates). Mix well (not to foam). Vortex is recommended.

Stability and storage:

Samples should be stored at -20°, or preferably at -70°C for long-term storage.

Do not store the diluted samples.

See Chapter 13 for stability of serum or plasma samples if stored at 2-8°C and effect of freezing/thawing on the concentration of human resistin.

Note: It is recommended to use a precision pipette and a careful technique to perform the dilution in order to get precise results.

Ask for protocol at DRG if assaying tissue culture medium, synovial fluid and cerebrospinal fluid (CSF).

11 ASSAY PROCEDURE

- 1. Pipet **100 μL** of diluted Standards, Quality Controls, Dilution Buffer (=Blank) and samples, preferably in duplicates, into the appropriate wells.
- 2. Incubate the plate at room temperature (ca. 25°C) for 1 hour, shaking at ca. 300 rpm on an orbital microplate shaker.
- 3. Wash the wells **3-times** with Wash Solution (0.35 mL per well). After final wash, invert and tap the plate strongly against paper towel.
- 4. Add 100 μL of Biotin Labelled Antibody Solution into each well.
- 5. Incubate the plate at room temperature (ca. 25°C) for 1 hour, shaking at ca. 300 rpm on an orbital microplate shaker.





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

- 6. Wash the wells **3-times** with Wash Solution (0.35 mL per well). After final wash, invert and tap the plate strongly against paper towel.
- 7. Add 100 µL of Streptavidin-HRP Conjugate into each well.
- 8. Incubate the plate at room temperature (ca. 25°C) for 1 hour, shaking at ca. 300 rpm on an orbital microplate shaker.
- 9. Wash the wells **3-times** with Wash Solution (0.35 mL per well). After final wash, invert and tap the plate strongly against paper towel.
- 10. Add $100 \ \mu L$ of Substrate Solution into each well. Avoid exposing the microtiter plate to direct sunlight. Covering the plate with e.g. aluminium foil is recommended.
- 11. Incubate the plate for **10 minutes** at room temperature. The incubation time may be extended [up to 20 minutes] if the reaction temperature is below than 20°C. Do not shake the plate during the incubation.
- 12. Stop the colour development by adding $100 \ \mu L$ of Stop Solution.
- 13. Determine the absorbance of each well using a microplate reader set to 450 nm, preferably with the reference wavelength set to 630 nm (acceptable range: 550 650 nm). Subtract readings at 630 nm (550 650 nm) from the readings at 450 nm. The absorbance should be read within 5 minutes following step 12

Note: If some samples and standard/s have absorbances above the upper limit of your microplate reader, perform a second reading at 405 nm. A new standard curve, constructed using the values measured at 405 nm, is used to determine resistin concentration of off-scale standards and samples. The readings at 405 nm should not replace the readings for samples that were "in range" at 450 nm.

Note 2: Manual washing: Aspirate wells and pipet 0.35 ml Wash Solution into each well. Aspirate wells and repeat twice. After final wash, invert and tap the plate strongly against paper towel. Make certain that Wash Solution has been removed entirely.





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

	strip 1+2	strip 3 + 4	strip 5+ 6	strip 7+ 8	strip 9+10	strip 11+ 12
A	Standard 50	QC Low	Sample 8	Sample 16	Sample 24	Sample 32
В	Standard 20	Sample 1	Sample 9	Sample 17	Sample 25	Sample 33
С	Standard 10	Sample 2	Sample 10	Sample 18	Sample 26	Sample 34
D	Standard 5	Sample 3	Sample 11	Sample 19	Sample 27	Sample 35
Е	Standard 2	Sample 4	Sample 12	Sample 20	Sample 28	Sample 36
F	Standard 1	Sample 5	Sample 13	Sample 21	Sample 29	Sample 37
G	Blank	Sample 6	Sample 14	Sample 22	Sample 30	Sample 38
Н	QC High	Sample 7	Sample 15	Sample 23	Sample 31	Sample 39

Figure 1: Example of a work sheet.

12 CALCULATIONS

Most microtiter plate readers perform automatic calculations of analyte concentration. The standard curve is constructed by plotting the mean absorbance (Y) of Standards against the known concentration (X) of Standards in logarithmic scale, using the four-parameter algorithm.

Results are reported as concentration of resistin in ng/mL in samples.

Alternatively, the logit log function can be used to linearize the standard curve (i.e. logit of the mean absorbance (Y) is plotted versus log of the known concentration (X) of Standards).

Samples, Quality Controls and Standards are all diluted 3x prior to analysis, so there is no need to take this dilution factor into account.





9

DRG[®] Resistin (human) ELISA (EIA-4572)

REVISED 18 JUNE 2010 RM (VERS. 4.0)



Figure 2: Typical Standard Curve for Human Resistin ELISA.

DRG International, Inc. USA Fax: (908) 233 0758 e-mail: corp@drg-international.com

RUO IN THE USA AND NON-EU COUNTRIES





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

13 PERFORMANCE CHARACTERISTICS

Typical analytical data obtained with Human Resistin ELISA are presented in this chapter.

13.1 Sensitivity

Limit of Detection (LOD) (defined as concentration of analyte giving absorbance higher than mean absorbance of blank* plus three standard deviations of the absorbance of blank: $A_{blank} + 3x \text{ SD}_{blank}$) is calculated from the real human resistin values in wells and is 0.033 ng/mL.

*Dilution Buffer is pipetted into blank wells.

13.2 Limit of assay

Results exceeding resistin level of 50 ng/mL should be repeated with more diluted samples.

Dilution factor needs to be taken into consideration in calculating the resistin concentration.

13.3 Specificity

The antibodies used in this ELISA are specific for human resistin with no detectable cross reactivities to human leptin, leptin receptor, adiponectin, TNF-alfa, RELM-beta, A-FABP and E-FABP at 100 ng/mL and IL-6, AGRP and Asp (C3adesArg) at 2 ng/mL.

Sera of several mammalian species were measured in the assay. See results below.

For details please contact us at drg@drg-diagnostics.com.

Mammalian serum sample	Observed cross reactivity
Bovine	no
Cat	no
Dog	no
Goat	no
Hamster	no
Horse	yes
Monkey	yes
Mouse	no
Pig	yes
Rabbit	no
Rat	no
Sheep	no





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

13.4 Precision

Intra-assay (Within-Run, n=8)

Sample	Mean (ng/mL)	SD (ng/mL)	CV (%)
1	7.53	0.21	2.8
2	11.35	0.39	3.4

Inter-assay (Run-to-Run, n=3)

Sample	Mean (ng/mL)	SD (ng/mL)	CV(%)
1	6.46	0.33	5.1
2	13.35	0.93	6.9

13.5 Spiking Recovery

Serum samples were spiked with different amounts of human resistin and assayed.

Sample	Observed (ng/mL)	Expected (ng/mL)	Recovery O/E (%)
1	5.55	-	-
	8.99	10.55	85.2
	13.35	15.55	85.9
	25.34	25.55	99.2
2	7.47	-	-
	10.88	12.47	87.2
	16.79	17.47	96.1
	26.07	27.47	94.9





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

13.6 Linearity

Serum samples were diluted with Dilution Buffer and assayed.

Sample	Dilution	Observed (ng/mL)	Expected (ng/mL)	Recovery O/E (%)
1	-	12.56	-	-
	2x	6.51	6.28	103.7
	4x	2.98	3.14	94.9
	8x	1.74	1.57	110.8
2		28.46	-	-
	2x	14.02	14.23	98.5
	4x	7.27	7.12	102.2
	8x	3.88	3.56	109.1

13.7 Effect of sample matrix

Heparin, citrate and EDTA plasmas were compared to respective serum samples from the same 10 individuals. Results are shown below:

Volunteer	Serum (ng/mL)	Plasma (ng/mL)		
No.		Heparin	Citrate	EDTA
1	14.1	11.2	12.0	13.3
2	7.9	6.6	6.7	6.5
3	11.3	10.2	9.9	12.2
4	9.0	8.3	8.0	9.0
5	5.9	6.2	7.0	8.2
6	9.3	8.9	7.8	9.3
7	6.2	6.3	6.6	7.7
8	5.6	5.9	5.9	5.6
9	5.1	6.0	4.8	5.4
10	6.9	6.5	8.1	7.9
Mean (ng/mL)	8.13	7.61	7.68	8.51
Mean Plasma/Serum (%)	-	99.1	90.2	104.7
Correlation Coeff. R²	-	0,95	0.89	0.86

12





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES



Figure 3: Resistin levels measured using Human Resistin ELISA from10 individuals using serum, heparin, citrate and EDTA plasma, respectively.

13.8 Stability of samples stored at 2-8°C

Samples should be stored at -20° C. However, no decline in concentration of resistin was observed in serum and plasma samples after 10 days when stored at 2-8°C. To avoid microbial contamination, samples were treated with ε -aminocaproic acid and sodium azide, resulting in the final concentration of 0.03% and 0.1%, respectively.





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

Volunteer	Incubation:	Serum	Plasma (ng/mL)		
No.	Temperature, Period	(ng/mL)	Heparin	Citrate	EDTA
	- 20°C	14.1	11.2	12.0	13.3
1	2-8°C, 1 day	14.5	13.2	13.1	10.7
	2-8°C, 10 days	14.2	10.9	13.2	12.9
2	- 20°C	9.3	8.9	7.8	9.3
	2-8°C, 1 day	10.2	8.4	8.2	9.9
	2-8°C, 10 days	8.8	8.7	7.4	8.7
3	- 20°C	5.1	6.0	4.8	5.4
	2-8°C, 1 day	5.1	5.4	4.6	5.7
	2-8°C, 10 days	5.4	5.5	4.1	5.5

13.9 Effect of Freezing/Thawing

No decline was observed in concentration of human resistin in serum and plasma samples after repeated (3x) freeze/thaw cycles. However it is recommended to avoid unnecessary repeated freezing/thawing of the samples.

Volunteer	Number of f/t	Serum	Plasma (ng/mL)		L)
No.	cycles	(ng/mL)	Heparin	Citrate	EDTA
	1x	14.1	11.2	12	13.3
1	3x	12.8	12.2	11.8	13.1
	5x	11.6	10.5	10.4	11.5
2	1x	9.3	8.9	7.8	9.3
	3x	9.4	8.3	7.3	9.1
	5x	9.2	8.7	7.1	8.4
	1x	5.1	6	4.8	5.4
3	3x	5.3	4.9	4.7	4.7
	5x	3.8	4.5	3.9	4.1

13.10 Reference range

It is recommended that each laboratory include its own panel of control sample in the assay.

Each laboratory should establish its own normal and pathological references ranges for resistin levels with the assay.

14 DEFINITION OF THE STANDARD

A recombinant protein is used as the standard. The recombinant resistin is a 19,5 kDa dimeric protein consisting of two 92 amino acid polypeptide chains which are disulfide-linked.

DRG International, Inc. USA Fax: (908) 233 0758 e-mail: corp@drg-international.com 14





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

15 TROUBLESHOOTING AND FAQS

1. Weak signal in all wells

Possible explanations:

- Omission of a reagent or a step
- Improper preparation or storage of a reagent
- Assay performed before reagents were allowed to come to room temperature
- Improper wavelength when reading absorbance

2. High signal and background in all wells

Possible explanations:

- Improper or inadequate washing
- Overdeveloping; incubation time with Substrate Solution should be decreased before addition of Stop Solution
- Incubation temperature over 30°C

3. High coefficient of variation (CV)

Possible explanation:

- Improper or inadequate washing
- Improper mixing Standards, Quality Controls or samples

16 REFERENCES

- o Lazar MA. Resistin- and Obesity-associated metabolic diseases. Horm Metab Res; 39(10):710-6 (2007)
- Meier U, Gressner AM: Endocrine regulation of energy metabolism: review of pathobiochemical and clinical chemical aspects of leptin, ghrelin, adiponectin, and resistin. *Clin Chem*; **50**(9):1511-25 (2004)
- o Kershaw EE, Flier JS: Adipose tissue as an endocrine organ. J Clin Endocrinol Metab; 89(6):2548-56 (2004)
- Hartman HB, Lazar MA at al.: Mechanisms Regulating Adipocyte Expression of Resistin. *The Journal of Biological Chemistry*; 277 (22), 19754-19761, (2002)
- Steppan C.M. et al.: The Hormone Links Obesity to Diabetes. *Nature*; **409**, 307-312, (2001)
- Sheng CH, Di J, Jin Y, Zhang YC, Wu M, Sun Y, Zhang GZ. Resistin is expressed in human hepatocytes and induces insulin resistance. *Endocrine*; 33(2):135-43. (2008)
- Busetto L, Bassetto F, Zocchi M, Zuliani F, Nolli ML, Pigozzo S, Coin A, Mazza M, Sergi G, Mazzoleni F, Enzi G. The effects of the surgical removal of subcutaneous adipose tissue on energy expenditure and adipocytokine concentrations in obese women. *Nutr Metab Cardiovasc Dis;* 18(2):112-120 (2008)
- Mineo D, Ambrogi V, Frasca L, Elena Cufari M, Pompeo E, Claudio Mineo T. Effects of Lung Volume Reduction Surgery for Emphysema on Glyco-lipidic Hormones. *Chest*; [Epub ahead of print] (2008)

DRG International, Inc. USA Fax: (908) 233 0758 e-mail: corp@drg-international.com 15





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

- Dolezalova R, Lacinova Z, Dolinkova M, Kleiblova P, Haluzikova D, Housa D, Papezova H, Haluzik M. Changes of endocrine function of adipose tissue in anorexia nervosa: comparison of circulating levels versus subcutaneous mRNA expression. *Clin Endocrinol*; 67(5):674-8 (2007)
- Ellington AA, Malik AR, Klee GG, Turner ST, Rule AD, Mosley TH Jr, Kullo IJ. Association of plasma resistin with glomerular filtration rate and albuminuria in hypertensive adults. *Hypertension*; **50**(4):708-14 (2007)
- Kim SJ, Nian C, McIntosh CH. Resistin is a key mediator of glucose-dependent insulinotropic polypeptide (GIP) stimulation of lipoprotein lipase (LPL) activity in adipocytes. *J Biol Chem*; 282(47):34139-47 (2007)
- Chen D, Fang Q, Chai Y, Wang H, Huang H, Dong M. Serum resistin in gestational diabetes mellitus and early postpartum. *Clin Endocrinol;* **67**(2):208-11 (2007)
- Cortelazzi D, Corbetta S, Ronzoni S, Pelle F, Marconi A, Cozzi V, Cetin I, Cortelazzi R, Beck-Peccoz P, Spada A. Maternal and foetal resistin and adiponectin concentrations in normal and complicated pregnancies. *Clin Endocrinol*; 66(3):447-53 (2007)
- Hoefle G, Saely CH, Risch L, Koch L, Schmid F, Rein P, Aczél S, Berchtold S, Drexel H. Relationship between the adipose-tissue hormone resistin and coronary artery disease. *Clin Chim Acta*; 386(1-2):1-6 (2007)
- Senolt L, Housa D, Vernerova Z, Jirasek T, Svobodova R, Veigl D, Anderlova K, Muller-Ladner U, Pavelka K, Haluzik M. Resistin in rheumatoid arthritis synovial tissue, synovial fluid and serum. *Ann Rheum Dis;* 66(4):458-63 (2007)
- Pilz S, Weihrauch G, Seelhorst U, Wellnitz B, Winkelmann BR, Boehm BO, Marz W. Implications of resistin plasma levels in subjects undergoing coronary angiography. *Clin Endocrino*; 66(3):380-6 (2007)
- Derosa G, Fogari E, D'Angelo A, Cicero AF, Salvadeo SA, Ragonesi PD, Ferrari I, Gravina A, Fassi R, Fogari R. Metabolic effects of telmisartan and irbesartan in type 2 diabetic patients with metabolic syndrome treated with rosiglitazone. J Clin Pharm Ther; 32(3):261-8 (2007)
- Lubos E, Messow CM, Schnabel R, Rupprecht HJ, Espinola-Klein C, Bickel C, Peetz D, Post F, Lackner KJ, Tiret L, Munzel T, Blankenberg S. Resistin, acute coronary syndrome and prognosis results from the AtheroGene study. *Atherosclerosis*; **193**(1):121-128 (2007)
- Koebnick C, Wagner K, Garcia AL, Gruendel S, Lahmann PH, Weickert MO, Mohlig M, Harsch IA, Einig C, Speth M, Katz N, Trippo U, Zunft HJ. Increase in serum resistin during weight loss in overweight subjects is related to lipid metabolism. *Int J Obes;* 30(7):1097-103 (2006)
- Perseghin G, Lattuada G, De Cobelli F, Ntali G, Esposito A, Burska A, Belloni E, Canu T, Ragogna F, Scifo P, Del Maschio A, Luzi L. Serum resistin and intra-hepatic fat content in nondiabetic individuals. *J Clin Endocrinol Metab;* 91(12):5122-5 (2006)
- Pagano C, Soardo G, Pilon C, Milocco C, Basan L, Milan G, Donnini D, Faggian D, Mussap M, Plebani M, Avellini C, Federspil G, Sechi LA, Vettor R. Increased serum resistin in nonalcoholic fatty liver disease is related to liver disease severity and not to insulin resistance. *J Clin Endocrinol Metab*; 91(3):1081-6 (2006)





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

- Kremen J, Dolinkova M, Krajickova J, Blaha J, Anderlova K, Lacinova Z, Haluzikova D, Bosanska L, Vokurka M, Svacina S, Haluzik M. Increased subcutaneous and epicardial adipose tissue production of proinflammatory cytokines in cardiac surgery patients: possible role in postoperative insulin resistance. *J Clin Endocrinol Metab*; **91**(11):4620-7 (2006)
- Menzaghi C, Coco A, Salvemini L, Thompson R, De Cosmo S, Doria A, Trischitta V. Heritability of serum resistin and its genetic correlation with insulin resistance-related features in nondiabetic Caucasians. *J Clin Endocrinol Metab*; 91(7):2792-5 (2006)
- Nagaev I, Bokarewa M, Tarkowski A, Smith U. Human resistin is a systemic immunederived proinflammatory cytokine targeting both leukocytes and adipocytes. *PLoS ONE;* **20**(1):e31 (2006)
- Westphal S, Borucki K, Taneva E, Makarova R, Luley C. Adipokines and treatment with niacin. *Metabolism*; 55(10):1283-5 (2006)
- Hui-Bing H, Migita K, Miyashita T, Maeda Y, Nakamura M, Yatsuhashi H, Ishibashi H, Eguchi K, Kimura. Relationship between serum resistin concentrations and inflammatory markers in patients with type 2 diabetes mellitus. *Metabolism*; 55(12):1670-3 (2006)
- Rae C, Graham A. Human resistin promotes macrophage lipid accumulation. *Diabetologia*; 49(5):1112-4 (2006)





REVISED 18 JUNE 2010 RM (VERS. 4.0)

RUO IN THE USA AND NON-EU COUNTRIES

SYMBOLS USED WITH DRG ASSAYS

Symbol	English	Deutsch	Français	Español	Italiano
Ĩ	Consult instructions for use	Gebrauchsanweisung beachten	Consulter les instructions d'utilisation	Consulte las instrucciones de uso	Consultare le istruzioni per l'uso
CE	European Conformity	CE-Konfirmitäts- kennzeichnung	Conformité aux normes européennes	Conformidad europea	Conformità europea
IVD	In vitro diagnostic device	In-vitro-Diagnostikum	Usage Diagnostic in vitro	Para uso Diagnóstico in vitro	Per uso Diagnostica in vitro
RUO	For research use only	Nur für Forschungszwecke	Seulement dans le cadre de recherches	Sólo para uso en investigación	Solo a scopo di ricerca
REF	Catalogue number	Katalog-Nr.	Numéro de catalogue	Número de catálogo	Numero di Catalogo
LOT	Lot. No. / Batch code	Chargen-Nr.	Numéro de lot	Número de lote	Numero di lotto
Σ	Contains sufficient for <n> tests/</n>	Ausreichend für "n" Ansätze	Contenu suffisant pour "n" tests	Contenido suficiente para <n> ensayos</n>	Contenuto sufficiente per "n" saggi
X	Storage Temperature	Lagerungstemperatur	Température de conservation	Temperatura de conservación	Temperatura di conservazione
Σ	Expiration Date	Mindesthaltbarkeits- datum	Date limite d'utilisation	Fecha de caducidad	Data di scadenza
	Legal Manufacturer	Hersteller	Fabricant	Fabricante	Fabbricante
Distributed by	Distributor	Vertreiber	Distributeur	Distribuidor	Distributore
Content	Content	Inhalt	Conditionnement	Contenido	Contenuto
Volume/No.	Volume / No.	Volumen/Anzahl	Volume/Quantité	Volumen/Número	Volume/Quantità

DRG International, Inc. USA Fax: (908) 233 0758 e-mail: <u>corp@drg-international.com</u>