

#### DRG<sup>®</sup> CRP hs ELISA (EIA-4584)



**RUO** in the USA

Revised 27 Aug. 2010 rm (Vers. 2.1)

Please use only the valid version of the package insert provided with the kit.

#### **1 PRINCIPLE OF THE CRP ELISA**

Microtiter strips coated with anti-CRP antibody are incubated with diluted standard sera and donor samples. During this incubation step CRP is bound specifically to the wells. After removal of the unbound serum proteins by a washing procedure, the antigen-antibody complex in each well is detected with specific peroxidase-conjugated antibodies.

After removal of the unbound conjugate, the strips are incubated with a chromogen solution containing tetramethylbenzidine and hydrogen peroxide: a blue colour develops in proportion to the amount of immunocomplex bound to the wells of the strips. The enzymatic reaction is stopped by the addition of  $0.5M H_2SO_4$  and the absorbance values at 450 nm are determined.

A standard curve is obtained by plotting the absorbance values versus the corresponding standard values. The concentration of CRP in donor samples is determined by interpolation from the standard curve.

In the United States, this kit is intended for Research Use Only.

#### 2 REAGENTS

- Coated Microtiter strips

   x 8-well strips coated with monoclonal antibodies to human CRP.
- 2. Standard Sera 5 vials,

each containing 1/10 prediluted CRP standard solutions (0.2 ml): 0 - 0.4 - 1 - 5 - 10  $\mu$ g/ml. Calibrated against the NIBSC 1<sup>st</sup> International Standard, 85/506. Contain 0,09 % NaN<sub>3</sub> and antimicrobial agents as preservatives.

- **3.** Conjugate 1 vial, containing peroxidase conjugated monoclonal anti-human CRP antibodies (12 ml). Contains antimicrobial agents and an inert red dye.
- Specimen Dilution Buffer 1 vial, containing 40 ml dilution buffer 5x concentrated. Contains 0.09 % NaN<sub>3</sub> and antimicrobial agents and an inert green dye.

## 5. **Washing Solution -** 1 vial containing 50 ml 20 x concentrated phosphate buffered washing solution.

- 6. Chromogen Solution 1 vial, containing 15 ml of a solution containing  $H_2O_2$  and tetramethylbenzidin.
- 7. **Stopping Solution** -1 vial, containing 12 ml of 0.5M H<sub>2</sub>SO<sub>4</sub>



## DRG<sup>®</sup> CRP hs ELISA (EIA-4584)



# CE

#### Revised 27 Aug. 2010 rm (Vers. 2.1)

#### **3** MATERIALS REQUIRED BUT NOT SUPPLIED

- 1. Precision micropipettes and standard laboratory pipettes.
- 2. Clean standard laboratory volumetric glassware.
- 3. Clean glass tubes for the dilution of the samples.
- 4. A microtiterplate reader capable of measuring absorbencies at 450 nm

#### 4 WARNINGS AND PRECAUTIONS FOR USERS

- 1. For Research Use Only.
- 2. Human blood components used in the preparation of the standard sera have been tested and found to be nonreactive for hepatitis B surface antigen and HIV I. Since no known method can ever offer complete assurance that products derived from human blood will not transmit hepatitis or other viral infections, it is recommended to handle these standard sera in the same way as potentially infectious material. Dispose donor samples and all materials used to perform this test as if they contain infectious agents.
- 3. Do not mix reagents or coated microtiter strips from kits with different lot numbers.
- 4. Some kit components contain sodium azide as a preservative. In order to prevent the formation of potentially explosive metal azides in laboratory plumbing, flush drains thoroughly after disposal of these solutions.

#### **5 STORAGE CONDITIONS**

2 °C to 8 °C

- 1. Store the microtiter strips in their original package with the desiccant until all the strips have been used.
- 2. Never use any kit components beyond the expiration date.

#### 6 SPECIMEN COLLECTION AND PREPARATION

Human serum and plasma may be used in this assay.

Remove serum from clot as soon as possible to avoid haemolysis. Lipemic and/or haemolysed samples can cause false results.

Transfer the serum to a clean storage tube.

Specimens may be stored at  $2 \degree C - 8 \degree C$  for a few days, or they can be stored frozen for a longer period of time. Avoid repeated freezing and thawing.

#### 7 ASSAY PROCEDURE

#### 7.1 General Remarks

- 1. Use a separate disposable tip for each sample transfer to avoid cross-contamination.
- 2. All reagents must be allowed to come to room temperature before use. All reagents must be mixed without foaming.
- 3. Once the assay has been started, all steps should be completed without interruption.
- 4. Absorbance is a function of the incubation tome and temperature. Therefore the size of the assay run should be limited. It is suggested to run no more than 20 donor samples with one set of Reference Standards in duplicate.

#### DRG International, Inc., USA Fax: (908) 233 0758 e-mail: corp@drg-international.com









RUO in the USA

# \_ہزر ( ( Revised 27 Aug. 2010 rm (Vers. 2.1)

#### 7.2 **Reconstitution of the Reagents**

#### Washing Solution:

dilute 50 ml of concentrated Washing Solution (5) to 1000 ml with distilled water. Reconstituted solution can be stored at least 1 month, store at  $2 \degree C - 8 \degree C$ .

At higher temperatures, the concentrated Washing solution (5) may appear cloudy, without affecting its performance. Upon dilution, the solution will be clear.

#### Sample Diluent:

Dilute 40 ml of the concentrated Sample Diluent to 200 ml with distilled water. Reconstituted solution can be stored at least 3 months or as long as solution remains clear. Store at  $2 \degree C - 8 \degree C$ .

#### 7.3 **Assay Procedure**

- 1. The <u>10 x prediluted standard sera</u> (2) are diluted <u>1:100</u> as follows : pipette 10 µl of each calibrator into separate glass dilution tubes. Add 990 µl of diluted Specimen Dilution Buffer (4) and mix carefully.
- The donor samples are diluted 1:1000 in two consecutive steps: 2. pipette 10 µl of each donor sample into separate glass dilution tubes and add 990 µl of diluted Specimen Dilution Buffer (4). Mix thoroughly. Add 450 µl of diluted Specimen Dilution Buffer to 50 µl of these 100 x prediluted samples. Mix thoroughly. Warning: do not store the diluted samples for more than 8 hours.
- Pipette 100  $\mu$ l of the diluted calibrators and samples into each of a pair of adjacent wells (1). 3.
- Incubate the covered microtiter strips for  $30 \pm 2$  min at room temperature. 4.
- 5. Wash the microtiter strips three times with Washing Solution. This can either be performed with a suitable microtiter plate washer or by briskly shaking out the contents of the strips and immersing them in washing solution. During the third step, the washing solution is left in the strips for 2-3 min. Change washing solution for each cycle. Finally empty the microtiter strips and remove excess fluid by blotting the inverted strips on adsorbent paper.
- Add 100 µl of Conjugate Solution (3) and incubate the covered microtiter strips for  $30 \pm 2$  min at room 6. temperature.
- Repeat the washing procedure as described in step 5. 7.
- 8. Add 100 µl of Chromogen (6) Solution to each well.
- 9. Incubate for  $10 \pm 2$  min at room temperature. Avoid light exposure during this step.
- 10. Add 50 µl of Stopping Solution (7) to each well.
- 11. Determine the absorbance of each well at 450 nm within 30 min following the addition of acid.



CE

#### DRG<sup>®</sup> CRP hs ELISA (EIA-4584)



## **RUO** in the USA

#### Revised 27 Aug. 2010 rm (Vers. 2.1)

#### 8 RESULTS

The average absorbance value of each calibrator is plotted against the corresponding CRP-value and the best calibration curve (e.g. log/linear) is constructed.

Use the average absorbance of each donor sample obtained in the hsCRP ELISA to determine the corresponding value by simple interpolation from the curve.

Depending on the experience and/or availability of computer capability, other methods of data reduction may be used.

#### 9 MINIMAL DETECTABLE CONCENTRATION

The minimal detectable concentration is approximately 0.02 µg/ml.

#### **10 REFERENCES / LITERATURE**

- 1. MITRA B, PANJA M. -High sensitive C-reactive protein: a novel biochemical markers and its role in coronary artery disease, J Assoc Physicians India. 2005 Jan;53:25-32.
- 2. KANDA T. C-reactive protein (CRP) in the cardiovascular system, Rinsho Byori, 2001 Apr, 49(4), 395-401.
- URSELLA, M.MAZZONE, G.PORTALE, A.TESTA, G.PIGNATARO, M. COVINO, P. FENICI, G.B. GASBARRINI, N. GENTILONI SILVERI. How to use the C-reactive protein in cardiac disease ? Ital Heart J. 2001 Mar; 2(3): 164-171.
- 4. KRAUSE KJ. C-reactive protein: a screening test for coronary disease? J Insur Med. 2001, 33(1), 4-11.
- 5. RIDKER PM. High-sensitivity C-reactive protein and cardiovascular risk: rationale for screening and primary prevention, Am J Cardiol, 2003 Aug 21, 92(4B), 17K-22K.
- 6. SHAH SH., NEWBY LK. C-reactive protein: a novel marker of cardiovascular risk, Cardiol Rev. 2003 Jul-Aug, 11.