

As of 31 Aug. 2010 rm (Vers. 1.1)**For Veterinary Use Only**

1 INTRODUCTION

1.1 Intended Use

The **DRG PMSG ELISA** is an enzyme immunoassay for the semiquantitative *in vitro* measurement of Pregnant Mare Serum Gonadotropin (PMSG) in equine serum and plasma. Not for Use in Human Determinations.

1.2 Summary and Explanation

PMSG or equine chorionic gonadotropin is secreted by the endometrial cups of the pregnant mares uterus. The hormone is found in the blood of the pregnant mare between the 40th and 120th days of gestation, reaching a peak at approximately the 60th day.

Measurement of PMSG provides a specific test for pregnancy since the hormone is only found in the pregnant mare. After 150 days of gestation, hormone levels are no longer detectable.

The PMSG diagnostic test as a means of pregnancy testing is a suitable and convenient method to confirm earlier ultrasound scanning techniques and allows the practitioner to obtain this information quickly and economically in the laboratory.

2 PRINCIPLE OF THE TEST

The DRG PMSG ELISA Kit is a solid phase enzyme-linked immunosorbent assay (ELISA) based on the sandwich principle.

The microtiter wells are coated with a monoclonal antibody directed towards a unique antigenic site of the PMSG molecule. An aliquot of sample containing endogenous PMSG is incubated in the coated well. After washing a second incubation follows with enzyme conjugate, which is an anti-PMSG antibody conjugated with horseradish peroxidase. After incubation the unbound conjugate is washed off.

The amount of bound peroxidase is proportional to the concentration of PMSG in the sample.

Having added the substrate solution, the intensity of colour developed is proportional to the concentration of PMSG in the sample.

3 WARNINGS AND PRECAUTIONS

1. This kit is for in vitro diagnostic use only. For Veterinary use only!
2. All blood components and biological materials should be handled as potentially hazardous in use and for disposal. Follow universal precautions as established by the Centers for Disease Control and Prevention and by the Occupational Safety and Health Administration when handling and disposing of infectious agents.
3. Before starting the assay, read the instructions completely and carefully. Use the valid version of the package insert provided with the kit. Be sure that everything is understood.
4. The microplate contains snap-off strips. Unused wells must be stored at 2 °C to 8 °C in the sealed foil pouch and used in the frame provided.
5. Pipetting of samples and reagents must be done as quickly as possible and in the same sequence for each step.
6. Use reservoirs only for single reagents. This especially applies to the substrate reservoirs. Using a reservoir for dispensing a substrate solution that had previously been used for the conjugate solution may turn solution colored. Do not pour reagents back into vials as reagent contamination may occur.

As of 31 Aug. 2010 rm (Vers. 1.1)**For Veterinary Use Only**

7. Mix the contents of the microplate wells thoroughly to ensure good test results. Do not reuse microwells.
8. Do not let wells dry during assay; add reagents immediately after completing the rinsing steps.
9. Allow the reagents to reach room temperature (21-26°C) before starting the test. Temperature will affect the absorbance readings of the assay. However, values for the patient samples will not be affected.
10. Never pipet by mouth and avoid contact of reagents and specimens with skin and mucous membranes.
11. Do not smoke, eat, drink or apply cosmetics in areas where specimens or kit reagents are handled.
12. Wear disposable latex gloves when handling specimens and reagents. Microbial contamination of reagents or specimens may give false results.
13. Handling should be done in accordance with the procedures defined by an appropriate national biohazard safety guideline or regulation.
14. Do not use reagents beyond expiry date as shown on the kit labels.
15. All indicated volumes have to be performed according to the protocol. Optimal test results are only obtained when using calibrated pipettes and microtiterplate readers.
16. Do not mix or use components from kits with different lot numbers. It is advised not to exchange wells of different plates even of the same lot. The kits may have been shipped or stored under different conditions and the binding characteristics of the plates may result slightly different.
17. Some reagents contain Proclin 300, BND and/or MIT as preservatives. In case of contact with eyes or skin, flush immediately with water.
18. TMB substrate has an irritant effect on skin and mucosa. In case of possible contact, wash eyes with an abundant volume of water and skin with soap and abundant water. Wash contaminated objects before reusing them. If inhaled, take the person to open air.
19. Chemicals and prepared or used reagents have to be treated as hazardous waste according to the national biohazard safety guideline or regulation.
20. For information on hazardous substances included in the kit please refer to Material Safety Data Sheets. Material Safety Data Sheets for this product are available upon request directly from DRG.

As of 31 Aug. 2010 rm (Vers. 1.1)

For Veterinary Use Only

4 REAGENTS

4.1 Reagents provided

1. **Microtiterwells**, 12x8 (break apart) strips, 96 wells;
Wells coated with anti-PMSG antibody (monoclonal).
2. **Neg. Control**, 1 vial, 2 mL, ready to use, white cap;
0 mIU/mL PMSG.
Contains non-mercury preservative.
3. **Cut-off Control**, 1 vial, 2 mL, ready to use, yellow cap.
For exact value please refer to vial label or QC-Datasheet.
Contains non-mercury preservative.
4. **Pos. Control**, 1 vial, 2 mL, ready to use, red cap
800 mIU/mL PMSG,.
Contains non-mercury preservative.
5. **Enzyme Conjugate**, 1 dropping bottle, 11 mL, ready to use, green cap;
Anti-PMSG antibody conjugated to horseradish peroxidase;
Contains non-mercury preservative.
6. **Substrate Solution**, 1 vial, 11 mL, ready to use, blue cap,
Tetramethylbenzidine (TMB).
7. **Transfer pipettes for single use**, 100 pieces

4.2 Materials required but not provided

- Distilled or deionized water
- Wash bottle
- Absorbent paper.
- Timer

4.3 Storage Conditions

When stored at 2 °C to 8 °C unopened reagents will retain reactivity until expiration date. Do not use reagents beyond this date.

Opened reagents must be stored at 2 °C to 8 °C. Microtiter wells must be stored at 2 °C to 8 °C. Once the foil bag has been opened, care should be taken to close it tightly again.

Opened kits retain activity for six weeks if stored as described above.

4.4 Reagent Preparation

Bring all reagents and required number of strips to room temperature prior to use.

4.5 Disposal of the Kit

The disposal of the kit must be made according to the national regulations. Special information for this product is given in the Material Safety Data Sheet.

As of 31 Aug. 2010 rm (Vers. 1.1)**For Veterinary Use Only****4.6 Damaged Test Kits**

In case of any severe damage to the test kit or components, DRG has to be informed in writing, at the latest, one week after receiving the kit. Severely damaged single components should not be used for a test run. They have to be stored until a final solution has been found. After this, they should be disposed according to the official regulations.

5 SPECIMEN COLLECTION AND PREPARATION

Serum or plasma can be used in this assay.

Do not use haemolytic, icteric or lipaemic specimens.

Please note: Samples containing sodium azide should not be used in the assay.

5.1 Specimen Storage and Preparation

If samples cannot be assayed immediately, they should be frozen at -20°C (up to two years). Thawed samples should be inverted several times prior to testing.

6 ASSAY PROCEDURE**6.1 General Remarks**

- All reagents and specimens must be allowed to come to room temperature before use. All reagents must be mixed without foaming.
- Once the test has been started, all steps should be completed without interruption.
- Use a new transfer pipette for each, control or sample in order to avoid cross contamination.
- Absorbance is a function of the incubation time and temperature. Before starting the assay, it is recommended that all reagents are ready, caps removed, all needed wells secured in holder, etc. This will ensure equal elapsed time for each pipetting step without interruption.
- As a general rule the enzymatic reaction is linearly proportional to time and temperature.
- The size of the assay run should be limited. Pipetting of all samples and controls should be completed within 10 minutes

As of 31 Aug. 2010 rm (Vers. 1.1)

For Veterinary Use Only

6.2 Test Procedure

1. Secure the desired number of Microtiter wells in the frame holder.
2. Dispense **3 drops** of *Negative Control*, *Cut-off Control* and *Positive Control* into appropriate wells.
3. Dispense **3 drops** of each **sample** each with a fresh transfer pipette into appropriate wells.
4. Incubate for **60 minutes** at room temperature.
5. Briskly shake out the contents of the wells.
Rinse the wells **3 times** with distilled water (300 µL per well). Strike the wells sharply on absorbent paper to remove residual droplets.
6. Dispense **3 drops** of *Enzyme Conjugate* into each well.
7. Incubate for **60 minutes** at room temperature.
8. Briskly shake out the contents of the wells.
Rinse the wells **5 times** with distilled water (300 µL per well). Strike the wells sharply on absorbent paper to remove residual droplets.
9. Add **3 drops** of *Substrate Solution* to each well.
10. Incubate for **30 minutes** at room temperature.
11. The intensity of developed blue color has to be interpreted.

7 INTERPRETATION OF RESULTS

The developed blue color of the sample has to be compared with the color of the Cut-off Control.

NEGATIVE: color intensity of sample < color of Cut-off Control

POSITIVE: color intensity of sample > color of Cut-off Control

8 QUALITY CONTROL

Good laboratory practice requires that controls be run with each calibration curve. A statistically significant number of controls should be assayed to establish mean values and acceptable ranges to assure proper performance.

It is recommended to use control samples according to state and federal regulations. The use of control samples is advised to assure the day to day validity of results. Use controls at both normal and pathological levels.

The controls and the corresponding results of the QC-Laboratory are stated in the QC certificate added to the kit. The values and ranges stated on the QC sheet always refer to the current kit lot and should be used for direct comparison of the results.

Employ appropriate statistical methods for analysing control values and trends. If the results of the assay do not fit to the established acceptable ranges of control materials patient results should be considered invalid.

In this case, please check the following technical areas: Pipetting and timing devices; photometer, expiration dates of reagents, storage and incubation conditions, aspiration and washing methods.

After checking the above mentioned items without finding any error contact your distributor or DRG directly.

As of 31 Aug. 2010 rm (Vers. 1.1)**For Veterinary Use Only****9 LIMITATIONS OF USE**

Reliable and reproducible results will be obtained when the assay procedure is performed with a complete understanding of the package insert instruction and with adherence to good laboratory practice.

Any improper handling of samples or modification of this test might influence the results.

10 LEGAL ASPECTS**10.1 Reliability of Results**

The test must be performed exactly as per the manufacturer's instructions for use. Moreover the user must strictly adhere to the rules of GLP (Good Laboratory Practice) or other applicable national standards and/or laws. This is especially relevant for the use of control reagents. It is important to always include, within the test procedure, a sufficient number of controls for validating the accuracy and precision of the test.

The test results are valid only if all controls are within the specified ranges and if all other test parameters are also within the given assay specifications. In case of any doubt or concern please contact DRG.

10.2 Therapeutic Consequences

Therapeutic consequences should never be based on laboratory results alone even if all test results are in agreement with the items as stated under point 10.1. Any laboratory result is only a part of the total clinical picture of a specimen.

Only in cases where the laboratory results are in acceptable agreement with the overall clinical picture of the specimen should therapeutic consequences be derived.

The test result itself should never be the sole determinant for deriving any therapeutic consequences.

10.3 Liability

Any modification of the test kit and/or exchange or mixture of any components of different lots from one test kit to another could negatively affect the intended results and validity of the overall test. Such modification and/or exchanges invalidate any claim for replacement.

Claims submitted due to customer misinterpretation of laboratory results subject to point 10.2. are also invalid.

Regardless, in the event of any claim, the manufacturer's liability is not to exceed the value of the test kit. Any damage caused to the test kit during transportation is not subject to the liability of the manufacturer.