



DRG® LL-37 (human) (EIA-5140)



As of 25 Feb. 2010 rm (Vers. 1.0)

Please use only the valid version of the package insert provided with the kit.

ATTENTION

For use in laboratory research only. Not for clinical or diagnostic use.

Note that this user protocol is not lot-specific and is representative for the current specifications of this product. Please consult the vial label and the certificate of quality control for information on specific lots. Also note that shipping conditions may differ from storage conditions.

For research use only. Not for use in or on humans or animals or for diagnostics. It is the responsibility of the user to comply with all local/state and federal rules in the use of this product. DRG is not responsible for any patent infringements that might result from the use or derivation of this product.

1 INTENDED USE

The human LL-37 ELISA kit is to be used for the in vitro quantitative determination of human LL-37 in plasma and cell culture supernatant samples. This kit is intended for laboratory research use only and is not for use in diagnostic or therapeutic procedures. The analysis should be performed by trained laboratory professionals.

2 INTRODUCTION

Cathelicidins are a family of antimicrobial proteins predominantly found in the peroxidase-negative granules of neutrophils. The cathelicidins are synthesized as preproproteins. Within the neutrophils, they are stored in granules as inactive proforms after removal of the signal peptide. The active biologic domains of the cathelicidins generally reside in the C-terminus. The C-terminal antimicrobial peptides are activated when cleaved from the proforms of the cathelicidins by serine proteases from azurophil granules. Human cationic antimicrobial protein (hCAP)18 is the only human cathelicidin identified to date. The antibacterial C-terminus of hCAP-18, LL-37 (37 amino acids), has been shown to exert broad antimicrobial activity towards gram-negative as well as gram-positive bacteria and to have synergistic antibacterial effects with the defensins. For instance, deficiency in saliva LL-37 accords with occurrence of periodontal disease in patients with morbus Kostmann. Moreover, it functions as a chemotactic agent for neutrophils, monocytes and T cells. LL-37 is markedly resistant to proteolytic degradation and to a limited extent also cytotoxic towards mammalian cells. LL-37 was demonstrated to be present in plasma in levels from 1.2 – 1.8 µg/ml and to be enhanced in infectious diseases.

3 KIT FEATURES

Working time of 3½ hours.

Minimum concentration which can be measured is 0.1 ng/ml.

Measurable concentration range of 0.1 to 100 ng/ml..


Working volume of 100 µl/well.

Cross-reactivity

Cross-reactivity for other species or proteins/peptides has not been tested.

As of 25 Feb. 2010 rm (Vers. 1.0)

4 PROTOCOL OVERVIEW



- The human LL-37 ELISA is a ready-to-use solid-phase enzyme-linked immunosorbent assay based on the sandwich principle with a working time of 3½ hours.
- The efficient format of 2 plates with twelve disposable 8-well strips allows free choice of batch size for the assay.
- Samples and standards are incubated in micro titer wells coated with antibodies recognizing human LL-37.
- Biotinylated tracer antibody will bind to captured human LL-37.
- Streptavidin-peroxidase conjugate will bind to the biotinylated tracer antibody.

As of 25 Feb. 2010 rm (Vers. 1.0)

- Streptavidin-peroxidase conjugate will react with the substrate, tetramethylbenzidine (TMB).
- The enzyme reaction is stopped by the addition of oxalic acid.
- The absorbance at 450 nm is measured with a spectrophotometer. A standard curve is obtained by plotting the absorbance (linear) versus the corresponding concentrations of the human LL-37 standards (log).
- The human LL-37 concentration of samples, which are run concurrently with the standards, can be determined from the standard curve.

5 KIT COMPONENTS AND STORAGE INSTRUCTIONS

Item no.	Kit component	Quantity	Color code
Vial 1	Wash/dilution buffer A 20x	2 vial (20 ml)	Grey
Vial 2	Wash/dilution buffer B 40x	1 vial (20 ml)	Gold
Vial 3	Standard	2 vial, 0.5 ml lyophilized	Yellow
Vial 4	Tracer, biotinylated	2 vials, 1 ml lyophilized	Green
Vial 5	Streptavidin-peroxidase	1 vial, 1 ml lyophilized	Blue
Vial 6	TMB substrate	1 vial (20 ml)	Purple
Vial 7	Stop solution	1 vial (20 ml)	Red
Item 8	12 Microtiter strips, pre-coated	2 plates	
Item 9	Frame	1	
Item 10	Adhesive covers	4	
Item 11	Certificate of quality control	1	
Item 12	Manual	1	
Item 13	Data collection sheet	1	

Table 1

Upon receipt, store individual components at 2 - 8°C. Do not freeze.

Do not use components beyond the expiration date printed on the kit label.

The standard, tracer and streptavidin-peroxidase are stable in lyophilized form until the expiration date indicated on the kit label, if stored at 2 - 8°C.

The exact concentration of the standard is indicated on the label of the vial and the certificate of quality control.

Once reconstituted the standard is stable for 12 hours, if stored at 2 - 8°C. For longer stability we recommend to store aliquots at -20°C. Stored at -20°C the standard will be stable for 1 month.

Once reconstituted, tracer and streptavidin-peroxidase are stable for 1 month if stored at 2 - 8°C.

Upon receipt, foil pouch around the plate should be vacuum-sealed and unpunctured.

Any irregularities to aforementioned conditions may influence plate performance in the assay.

Return unused strips immediately to the foil pouch containing the desiccant pack and reseal along the entire edge of the zip-seal. Quality guaranteed until expiration date if stored at 2 - 8°C.

As of 25 Feb. 2010 rm (Vers. 1.0)

5.1 Materials required but not provided

Calibrated micropipettes and disposable tips.

Distilled or de-ionized water.

Plate washer: automatic or manual.

In case a plate washer is used the supplied wash buffer is not sufficient. Additional wash buffer can be ordered separately. Please contact your local distributor.

Polypropylene tubes.

Calibrated ELISA plate reader capable of measuring absorbance at 450 nm.

6 WARNINGS AND PRECAUTIONS

For research use only, not for diagnostic or therapeutic use.

This kit should only be used by qualified laboratory staff.

Do not under any circumstances add sodium azide as preservative to any of the components.

Do not use kit components beyond the expiration date.

Do not mix reagents from different kits and lots. The reagents have been standardized as a unit for a given lot. Use only the reagents supplied by manufacturer.

The assay has been optimized for the indicated standard range. Do not change the standard range.

Standard, tracer and streptavidin-peroxidase vials should be opened after reconstitution. Open vials carefully: vials are under vacuum.

Do not ingest any of the kit components.

Kit reagents contain 2-chloroacetamide as a preservative. 2-Chloroacetamide is harmful in contact with skin and toxic if swallowed. In case of accident or if you feel unwell, seek medical advice immediately.

The TMB substrate is light sensitive, keep away from bright light. The solution should be colorless until use.

The stop solution contains 2% oxalic acid and can cause irritation or burns to respiratory system, skin and eyes. Direct contact with skin and eyes should be strictly avoided. If contact occurs, rinse immediately with plenty of water and seek medical advice.

Incubation times, incubation temperature and pipetting volumes other than those specified may give erroneous results.

Do not reuse micro wells or pour reagents back into their bottles once dispensed.

Handle all biological samples as potentially hazardous and capable of transmitting diseases.

Hemolyzed, hyperlipemic, heat-treated or contaminated samples may give erroneous results.

Use polypropylene tubes for preparation of standard and samples. Do not use polystyrene tubes or sample plates.

As of 25 Feb. 2010 rm (Vers. 1.0)

7 SAMPLE PREPARATION

7.1 Collection and handling

Plasma

Please be aware that human LL-37 is released from neutrophils into serum in the process of blood coagulation. This will lead to false positive results. It is therefore advised to use 'careful plasma', which can be obtained as follows.

Keep freshly collected blood on ice. Within 20 minutes after blood sampling, separate plasma by centrifugation: 1500xg at 4°C for 15 min. Remove plasma and transfer to fresh polypropylene tube. Be careful to not disturb white cells in the buffy coat. Recentrifuge the transferred plasma in order to avoid every contamination with white blood cells: 1500xg at 4°C for 15 min.

Most reliable results are obtained if EDTA plasma is used.

7.2 Storage

Store samples below -20°C, preferably at -70°C in polypropylene tubes. Storage at -20°C can affect recovery of human LL-37. Use samples within 24 hours after thawing. Avoid multiple freeze-thaw cycles which may cause loss of human LL-37 activity and give erroneous results.

Do not use hemolyzed, hyperlipemic, heat-treated or contaminated samples.

Before performing the assay, samples should be brought to room temperature (18 – 25°C) and mixed gently. Prepare all samples (controls and test samples) prior to starting the assay procedure. Avoid foaming.

7.3 Dilution procedures

Plasma samples

Human LL-37 can be measured accurately if plasma samples are diluted at least 20x with supplied wash/dilution buffer in polypropylene tubes.

Note that most reliable results are obtained with EDTA plasma.

Remark regarding recommended sample dilution

The recommended dilution for samples should be used as a guideline. The recovery of human LL-37 from an undiluted sample is not 100% and may vary from sample to sample. When testing less diluted samples it is advisable to run recovery experiments to determine the influence of the matrix on the detection of human LL-37.

Do not use polystyrene tubes or sample plates for preparation or dilution of the samples.

As of 25 Feb. 2010 rm (Vers. 1.0)

8 REAGENT PREPARATION

Allow all the reagents to equilibrate to room temperature (20 – 25°C) prior to use. Return to proper storage conditions immediately after use.

Wash / dilution buffer

Dilute 20x concentrated wash/dilution buffer A by mixing 40 ml with 360 ml distilled or de-ionized water.

Dilute 40x concentrated wash/dilution buffer B by mixing 20 ml with 380 ml distilled or de-ionized water.

Finally combine both solutions equally and mix well. The wash/dilution buffer is sufficient for 2 x 96 tests.

In case less tests are required, prepare the required volume by dilution of 1 part 20x concentrated wash/dilution buffer A with 9 parts of distilled or de-ionized water and 1 part 40x concentrated wash/dilution buffer B with 19 parts of distilled or de-ionized water. Finally combine both solutions equally and mix well.

Concentrated buffer may contain crystals. In case the crystals do not disappear at room temperature within 1 hour, concentrated buffer can be warmed up to 37°C. Do not shake the solution.

Standard solution

The standard is reconstituted by injection of 1 ml of distilled or de-ionized water. Prepare each human LL-37 standard in polypropylene tubes by serial dilution of the reconstituted standard with wash/dilution buffer as shown in Table 2.

Tube	Volume wash/dilution buffer	Volume standard	Concentration (ng/ml)
1	See certificate of quality control	150 µl vial 3	100
2	250 µl	125 µl tube 1	33
3	250 µl	125 µl tube 2	11
4	250 µl	125 µl tube 3	3.7
5	250 µl	125 µl tube 4	1.2
6	250 µl	125 µl tube 5	0.4
7	250 µl	125 µl tube 6	0.1
8	250 µl	-	0

Table 2

Tracer solution

The tracer is reconstituted by injection of 1 ml distilled or de-ionized water. Dilute the reconstituted 1 ml tracer with 11 ml wash/dilution buffer, which is sufficient for 1 x 96 tests. In case less volume is required, prepare the desired volume of tracer by diluting 1 part of the reconstituted tracer with 11 parts of wash/dilution buffer.

As of 25 Feb. 2010 rm (Vers. 1.0)

Streptavidin-peroxidase solution

The streptavidin-peroxidase is reconstituted by injection of 1 ml distilled or de-ionized water. Dilute the reconstituted 1 ml streptavidin-peroxidase with 23 ml wash/dilution buffer, which is sufficient for 2 x 96 tests. In case less volume is desired, prepare the required volume of streptavidin-peroxidase solution by diluting 1 part of the reconstituted streptavidin-peroxidase with 23 parts of wash/dilution buffer.

9 ELISA PROTOCOL

Bring all reagents to room temperature (20 - 25°C) before use.

1. Determine the number of test wells required, put the necessary microwell strips into the supplied frame, and fill out the data collection sheet. Return the unused strips to the storage bag with desiccant, seal and store at 2 - 8°C.
2. Transfer 100 µl in duplicate of standard, samples, or controls into appropriate wells. Do not touch the side or bottom of the wells.
3. Apply an adhesive cover to the tray. Tap the tray to eliminate any air bubbles. Be careful not to splash liquid onto the cover.
4. Incubate the strips or plate for 1 hour at room temperature.
5. Wash the plates 4 times with wash/dilution buffer using a plate washer or as follows:
 - a. Carefully remove the plate sealer, avoid splashing.
 - b. Empty the plate by inverting plate and shaking contents out over the sink, keep inverted and tap dry on a thick layer of tissues.
 - c. Add 200 µl of wash/dilution buffer to each well, wait 20 seconds, empty the plate as described in 5b.
 - d. Repeat the washing procedure 5b/5c three times.
 - e. Empty the plate and gently tap on thick layer of tissues.
6. Add 100 µl of diluted tracer to each well using the same pipetting order as applied in step 2. Do not touch the side or bottom of the wells.
7. Cover the tray with an adhesive cover. Incubate the tray for 1 hour at room temperature.
8. Repeat the wash procedure described in step 5.
9. Add 100 µl of diluted streptavidin-peroxidase to each well, using the same pipetting order as applied in step 2. Do not touch the side or bottom of the wells.
10. Cover the tray with an adhesive cover, incubate the tray for 1 hour at room temperature.
11. Repeat the wash procedure described in step 5.
12. Add 100 µl of TMB substrate to each well, using the same pipetting order as applied in step 2. Do not touch the side or bottom of the wells.
13. Cover the tray with a new adhesive cover, incubate the tray for 20 – 30 minutes at room temperature. Avoid exposing the micro well strips to direct sunlight. Covering the plate with aluminum foil is recommended.
14. Stop the reaction by adding 100 µl of stop solution with the same sequence and timing as used in step 12. Mix solutions in the wells thoroughly by gently swirling the plate. Gently tap the tray to eliminate any air bubbles trapped in the wells.
15. Read the plate within 30 minutes after addition of stop solution at 450 nm using a plate reader, following the instructions provided by the instrument's manufacturer.

As of 25 Feb. 2010 rm (Vers. 1.0)

10 INTERPRETATION OF RESULTS

Calculate the mean absorbance for each set of duplicate standards, control and samples.

If individual absorbance values differ by more than 15% from the corresponding mean value, the result is considered suspect and the sample should be retested.

The mean absorbance of the zero standard should be less than 0.3.

Create a standard curve using computer software capable of generating a good curve fit. The mean absorbance for each standard concentration is plotted on the vertical (Y) axis versus the corresponding concentration on the horizontal (X) axis (logarithmic scale). For an example of the standard curve see certificate of quality control included with the kit. If the standard is out of range, the results of the test samples are not reliable. The test should be repeated.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

Samples that give a mean absorbance above the absorbance for the highest standard concentration are out of range of the assay. These samples should be retested at a higher dilution.

11 TECHNICAL HINTS

User should be trained and familiar with ELISA assays and test procedure.

If you are not familiar with the ELISA technique it is recommended to perform a pilot assay prior to evaluation of your samples. Perform the assay with a standard curve only following the instructions.

Improper or insufficient washing at any stage of the procedure will result in either false positive or false negative results.

Completely empty wells before dispensing wash buffer, fill with wash buffer as indicated for each cycle and do not allow wells to sit uncovered or dry for extended periods.

Since exact conditions may vary from assay to assay, a standard curve must be established for every run. If the standard is out of range, the results of the test samples are not reliable. The test should be repeated.

Do not mix reagents from different batches, or other reagents and strips. Remainders should not be mixed with contents of freshly opened vials.

Each time the kit is used, fresh dilutions of standard, sample, tracer, streptavidin peroxidase and buffers should be made.

Caps and vials are not interchangeable. Caps should be replaced on the corresponding vials.

To avoid cross-contaminations, change pipette tips between reagent additions of each standard, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.

To ensure accurate results, proper adhesion of supplied covers during incubation steps is necessary.

The waste disposal should be performed according to your laboratory regulations.

12 QUALITY CONTROL

The certificate of quality control included in this kit is lot specific and is to be used to verify results obtained by your laboratory. The absorption values provided on the certificate of quality control are to be used as a guideline only. The results obtained by your laboratory may differ.

This assay is designed to eliminate interference by soluble receptors, binding proteins, and other factors present in biological samples. Until all factors have been tested in the DRG immunoassay, the possibility of interference cannot be excluded.

For optimal performance of this kit, it is advised to work according to good laboratory practice.

As of 25 Feb. 2010 rm (Vers. 1.0)

13 TROUBLESHOOTING

Warranty claims and complaints in respect of deficiencies must be logged before expiry date of the product. A written complaint containing lot number of the product and experimental data shall be sent to drg@drg-diagnostics.de. Suggestions summarized below in Table 3 can be used as guideline in case of unexpected assay results.

Low absorbance	High absorbance	Poor duplicates	All wells positive	All wells negative	Possible cause
•	•		•	•	Kit materials or reagents are contaminated or expired
•					Incorrect reagents used
•		•	•		Lyophilized reagents are not properly reconstituted
•	•	•	•	•	Incorrect dilutions or pipetting errors
•		•			Improper plastics used for preparation of standard and/or samples
•	•				Improper incubation times or temperature
		•			Especially in case of 37°C incubation: plates are not incubated uniformly
•					Assay performed before reagents were brought to room temperature
•	•	•	•	•	Procedure not followed correctly
				•	Omission of a reagent or a step
		•			Poor mixing of samples
	•		•		Low purity of water
	•	•			Strips were kept dry for too long during/after washing
	•	•	•		Inefficient washing
	•	•			Cross-contamination from other samples or positive control
		•	•		TMB solution is not clear or colorless
•	•				Wrong filter in the micro titer reader
	•	•			Air bubbles
		•			Imprecise sealing of the plate after use
•					Wrong storage conditions

Table 3



DRG[®] LL-37 (human) (EIA-5140)













As of 25 Feb. 2010 rm (Vers. 1.0)











14 REFERENCES / LITERATURE

1. Martineau, A et al; IFN- γ - and TNF-independent vitamin D-inducible human suppression of mycobacteria: the role of cathelicidin LL-37, J Immunol 2007, 178: 7190

As of 25 Feb. 2010 rm (Vers. 1.0)

SYMBOLS USED WITH DRG ASSAYS

Symbol	English	Deutsch	Français	Español	Italiano
	Consult instructions for use	Gebrauchsanweisung beachten	Consulter les instructions d'utilisation	Consulte las instrucciones de uso	Consultare le istruzioni per l'uso
	European Conformity	CE-Konformitätskennzeichnung	Conformité aux normes européennes	Conformidad europea	Conformità europea
	In vitro diagnostic device	In-vitro-Diagnostikum	Usage Diagnostic in vitro	Para uso Diagnóstico in vitro	Per uso Diagnostica in vitro
	For research use only	Nur für Forschungszwecke	Seulement dans le cadre de recherches	Sólo para uso en investigación	Solo a scopo di ricerca
	Catalogue number	Katalog-Nr.	Numéro de catalogue	Número de catálogo	Numero di Catalogo
	Lot. No. / Batch code	Chargen-Nr.	Numéro de lot	Número de lote	Numero di lotto
	Contains sufficient for <n> tests/	Ausreichend für "n" Ansätze	Contenu suffisant pour "n" tests	Contenido suficiente para <n> ensayos	Contenuto sufficiente per "n" saggi
	Storage Temperature	Lagerungstemperatur	Température de conservation	Temperatura de conservación	Temperatura di conservazione
	Expiration Date	Mindesthaltbarkeits-datum	Date limite d'utilisation	Fecha de caducidad	Data di scadenza
	Legal Manufacturer	Hersteller	Fabricant	Fabricante	Fabbricante
Distributed by	Distributor	Vertreiber	Distributeur	Distribuidor	Distributore
Content	Content	Inhalt	Conditionnement	Contenido	Contenuto
Volume/No.	Volume / No.	Volumen/Anzahl	Volume/Quantité	Volumen/Número	Volume/Quantità

Symbol	Portugues	Dansk	Svenska	Ελληνικά
	Consulte as instruções de utilização	Se brugsanvisning	Se bruksanvisningen	Εγχειρίδιο χρήστη
	Conformidade com as normas europeias	Europaesk overensstemmelse	Europeisk överensstämmelse	Ευρωπαϊκή Συμμόρφωση
	Diagnóstico in vitro	In vitro diagnostik	Diagnostik in vitro	in vitro διαγνωστικό
				
	Catálogo n.º	Katalognummer	Katalog nummer	Αριθμός καταλόγου
	No do lote	Lot nummer	Batch-nummer	Αριθμός Παρτίδος
		Indeholder tilstrækkeligt til "n" test	Innehåller tillräckligt till "n" tester	Περιεχόμενο επαρκές για «n» εξετάσεις
	Temperatura de conservação	Opbevarings-temperatur	Förvaringstemperatur	Θερμοκρασία αποθήκευσης
	Prazo de validade	Udløbsdato	Bäst före datum	Ημερομηνία λήξης
	Fabricante	Producent	Tillverkare	Κατασκευαστής
Distributed by				
Content	Conteúdo	Indhold	Innehåll	Περιεχόμενο
Volume/No.	Volume/Número	Volumen/antal	Volym/antal	Όγκος/αριθ..