

As of 17 Mar. 2010 rm (Vers. 1.0)

Please use only the valid version of the package insert provided with the kit.

For research use only.

Not for diagnostic or therapeutic procedures.

1 INTENDED USE

The rat/mouse OPG ELISA Kit is a sandwich ELISA intended for the quantitative determination of Osteoprotegerin in serum, plasma, urine and cell culture supernatants. **It is for research use only.**

2 SUMMARY AND EXPLANATION

Osteoprotegerin (OPG) or **Osteoclastogenesis inhibitory factor (OCIF)** is a dimeric glycoprotein of the TNF receptor family with a molecular weight of 60 kD resp. 120 kD which shows an inhibitory effect on osteoclasts and osteoclast precursor cells.

Osteoprotegerin is a soluble "decoy"-receptor and is produced in different tissues, e.g. bone, skin, liver, stomach, intestine and lung. As a so-called "decoy receptor" OPG inhibits the binding of RANK to RANKL (OPG-L, osteoclast differentiation factor, ODF) and thus inhibits the recruitment, proliferation and activation of osteoclasts.

OPG shows an inhibitory effect on osteoclasts. Osteoclast formation activity may be determined principally by the relative concentration of OPGL/osteoclast differentiation factor (ODF) to OPG/OCIF in the bone marrow microenvironment. Alterations of this ratio may be the major cause of bone loss in many imbalances in bone metabolism such as osteoporosis, osteopetrosis, metastatic osteolytic lesions and rheumatic bone degradation.

Indication

- Postmenopausal and senile osteoporosis
- Glucocorticoid-induced osteoporosis
- Diseases with locally increased resorption activity
- Therapy monitoring after treatment with OPG
- Arthritis
- Oncology

3 PRINCIPLE OF THE TEST

This sandwich-type ELISA is an assay for the direct determination of OPG in serum, plasma and urine. In this assay two highly specific antibodies against OPG are used. The capture antibody is attached to the wells of the microtiter plate, the detection antibody is labeled with biotin.

In a first incubation step the samples and the biotinylated antibody against OPG react with the coated capture antibody on the microtiter plate. A sandwich-type complex is formed consisting of the binding antibody on the plate, OPG and the biotinylated detection antibody. To remove all unspecific bound substances a washing step is carried out.

In a second step streptavidin – peroxidase is added which reacts with the detection antibody. After another washing step, the solid phase is incubated with the substrate, TMB. An acidic stopping solution is subsequently added. The blue color changes to yellow. The intensity of the yellow color is directly proportional to the concentration of OPG in the sample.

As of 17 Mar. 2010 rm (Vers. 1.0)

A dose - response curve of the absorbance units at 450 nm versus concentration is generated. OPG in the samples is determined directly from this calibration curve.

4 MATERIAL SUPPLIED

Label	Kit Components	Quantity
PLATE	Microtiter plate, 12 x 8 strips	96
WASHBUF	ELISA washing buffer concentrate (10x)	100 ml
COATBUF	Coating buffer	30 ml
COATAB	Capture antibody (rat anti-mouse OPG), lyophilized	1 vial
2.AB	Detection antibody (goat anti-mouse OPG, biotinylated), lyophilized	1 vial
STD	Calibrator (4000 pg) lyophilized	1 vial
CONJ	Conjugate, (Streptavidin-HRP-labeled), ready-to-use	1 vial
DIL	Reagent Diluent	100 ml
SUB	TMB substrate (Tetramethylbenzidine), ready-to-use	15 ml
STOP	ELISA stop solution, ready-to-use	1 x 15 ml

5 MATERIAL AND EQUIPMENT REQUIRED BUT NOT PROVIDED

1,5 ml reaction vials (Eppendorf)
Precision pipettes calibrated to deliver 10 -1000 µl and disposable tips.
Centrifuge capable of 3000 x g
ELISA reader
Vortex-mixer
Distilled or deionized water

6 PREPARATION AND STORAGE OF REAGENTS

- To run assay more than once, ensure that reagents are stored at conditions stated on the label.
Prepare only the appropriate amount necessary for each assay.
The kit can be used up to 4 times within the expiry date stated on the label.
- Reagents with a volume less than **100 µl** should be centrifuged before use to avoid loss of volume.
- The **WASHBUF** (ELISA wash buffer concentrate) should be diluted with aqua bidist. **1:10** before use (100 ml WASHBUF + 900 ml aqua bidist.), mix well.
Crystals could occur due to high salt concentration in the stock solutions. The crystals must be redissolved at room temperature or at 37°C before dilution.
The **buffer concentrate** is stable at **2-8°C** until the expiry date stated on the label.
Diluted **buffer solution** can be stored in a closed flask at **2-8°C for one month**.

As of 17 Mar. 2010 rm (Vers. 1.0)

- **COATAB (Capture Antibody, 1 vial)**, 720 µg/ml of rat anti-mouse OPG.
Reconstitute with **70 µl bidistilled water**. After reconstitution, store at 2 - 8°C for up to 60 days or aliquot and store at -20°C to -70°C for up to 6 months.
Dilute immediately before use to a **working concentration of 4.0 µg/mL in COATBUF** (coating buffer).
- **2.AB (Detection Antibody, 1 vial)**, 36 µg/ml of biotinylated goat antimouse OPG.
Reconstitute with **70 µl bidistilled water**. After reconstitution, store at 2 – 8°C for up to 60 days or aliquot and store at - 20°C to -70°C for up to 6 months.
Dilute to a **working concentration of 200 ng/ml in DIL** (Reagent Diluent).
- **STD (Standard, 1 vial)**, 4000 pg/ml of recombinant mouse OPG.
Reconstitute with **0.6 mL bidistilled water**. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions.
Store reconstituted standard at 2 - 8° C or aliquot and store at -70° C for up to 2 months.
A seven point standard curve using **2-fold serial dilutions in DIL** (Reagent Diluent), and a high standard of 4000 pg/mL is recommended.
- **CONJ (Streptavidin-HRP, 1 vial)**, 60 µl of streptavidin conjugated to horseradish-peroxidase.
Store at 2 - 8° C. **Do not freeze**.
Dilute to the **working concentration 1:200 in DIL** (Reagent Diluent).

7 PRECAUTIONS

For research use only.

Human materials used in kit components were tested and found to be negative for HIV, Hepatitis B and Hepatitis C.

However, for safety reasons, all kit components should be treated as potentially infectious.

Stop solution is composed of sulfuric acid, which is a strong acid. Even diluted, it still must be handled with care. It can cause acid burns and should be handled with gloves, eye protection, and appropriate protective clothing. Any spills should be wiped out immediately with copious quantities of water.

Reagents should not be used beyond the expiration date stated on kit label.

8 SPECIMEN COLLECTION AND PREPARATION

Serum, plasma and urine samples

Serum, plasma and urine samples can be used without any dilution. Serum must be centrifuged and aliquoted within 90 minutes after collection and stored at -20°C until use.

9 ASSAY PROCEDURE

9.1 Procedural notes

Do not mix different lot numbers of any kit component.

Guidelines for medical laboratories should be followed.

Incubation time, incubation temperature and pipetting volumes of the components are defined by the producer. Any variation of the test procedure, which is not coordinated with the producer, may influence the results of the test. DRG can therefore not be held responsible for any damage resulting from wrong use.

As of 17 Mar. 2010 rm (Vers. 1.0)

The assay should always be performed according the enclosed manual.

9.2 Plate Preparation

Coat a 96-well microplate with **100 µL per well of the diluted COATAB** (capture antibody). Seal the plate and incubate overnight at room temperature.

Aspirate and wash the wells **5 x with 250 µl** ELISA wash buffer. Complete removal of liquid at each step is essential for good performance. After the last wash, remove any remaining wash buffer by aspirating or by inverting the plate and blotting it against clean paper towels.

Block plates by adding **250 µl of DIL** (reagent diluent) to each well. Incubate at room temperature for 1 hour.

Aspirate and wash the wells **5 x with 250 µl** ELISA wash buffer. After the last wash, remove any remaining wash buffer by aspirating or by inverting the plate and blotting it against clean paper towels. The plates are now ready for sample addition.

9.3 Assay Procedure

1. Add **100 µl of SAMPLE** (sample) or **STD** (standards) in Reagent Diluent, per well.
Cover with an adhesive strip and incubate 2 hours at room temperature.
2. Aspirate and wash the wells **5 x with 250 µl** ELISA wash buffer. After the last wash, remove any remaining wash buffer by aspirating or by inverting the plate and blotting it against clean paper towels.
3. Add **100 µl of the 2. AB** (detection antibody), diluted in Reagent Diluent, to each well.
Cover with a new adhesive strip and incubate 2 hours at room temperature.
4. Aspirate and wash the wells **5 x with 250 µl** ELISA wash buffer. After the last wash, remove any remaining wash buffer by aspirating or by inverting the plate and blotting it against clean paper towels.
5. Add **100 µl of the working dilution of CONJ** (streptavidin-HRP) to each well.
Cover the plate and incubate for 20 minutes at room temperature. Avoid placing the plate in direct light.
6. Aspirate and wash the wells **5 x with 250 µl** ELISA wash buffer. After the last wash, remove any remaining wash buffer by aspirating or by inverting the plate and blotting it against clean paper towels.
7. Add **100 µl of SUB** (substrate solution) to each well.
Incubate for 20 minutes at room temperature. Avoid placing the plate in direct light.
8. Add **50 µl of STOP** (stop solution) to each well. Gently tap the plate to ensure thorough mixing.
9. Determine absorption immediately with an ELISA reader at **450 nm** against 620 nm (or 690 nm) as a reference. If no reference wavelength is available, read only at 450 nm. If the extinction of the highest standard exceeds the range of the photometer, absorption must be measured immediately at 405 nm against 620 nm as a reference.

10 RESULTS

The following algorithms can be used alternatively to calculate the results. We recommend to use the "4-Parameter-algorithm".

1. 4-parameter-algorithm

It is recommended to use a linear ordinate for optical density and a logarithmic abscissa for concentration. When using a logarithmic abscissa, the zero calibrator must be specified with a value less than 1 (e. g. 0.01).

As of 17 Mar. 2010 rm (Vers. 1.0)

2. Point-to-point-calculation

We recommend a linear ordinate for optical density and a linear abscissa for concentration.

3. Spline-algorithm

We recommend a linear ordinate for optical density and a logarithmic abscissa for concentration. When using a logarithmic abscissa, the zero calibrator must be specified with a value less than 1 (e. g. 0.01). The plausibility of the pairs of values should be examined before the automatic evaluation of the results. If this option is not available with the used program, a control of the paired values should be done manually.

11 LIMITATIONS

Samples with OPG levels greater then the highest standard value should be further diluted with wash buffer and re-assayed.

12 QUALITY CONTROL

It is recommended to use control samples for internal quality control.

Control samples or serum pools should be analyzed with each run of calibrators and patient samples. Results, generated from the analysis of control samples, should be evaluated for acceptability using appropriate statistical methods. In assays in which one or more of the quality control sample values are located outside the acceptable limits, the results for the patient sample may not be valid.

13 REFERENCES

1. Hofbauer LC, Osteoprotegerin ligand and osteoprotegerin: novel implications for osteoclast biology and bone metabolism. *European Journal of Endocrinology* (1999), 141: 195-210
2. Aubin JE, and E Bonnellye, Osteoprotegerin and its Ligand: A New Paradigm for Regulation of Osteoclastogenesis and Bone Resorption. *Medscape Women Health* (2000), 5(2)
3. Simonet WS, Lacey DL, Dunstan CR, Kelley M, Chang M-S, Lüthi R, et al., Osteoprotegerin, a novel secreted protein involved in the regulation of bone density. (1997) *Cell* 89: 309-319
4. Bucay N, Sarosi I, Dunstan CR, Morony S, Tarpley J, Capparelli C et al., Osteoprotegerin-deficient mice develop early onset osteoporosis and arterial calcification. (1998) *Genes and Development* 12: 1260- 1268
5. Bekker PJ, Holloway D, Nakanishi A, Arrighi HM, and CR Dunstan, Osteoprotegerin (OPG) has Potent and Sustained Anti-Resorptive Activity in Postmenopausal Women. *21st Annual Meeting of the ASBMR* (1999), Abstract No. 1190
6. Makhluif HA, Mueller SM, Mizuno S, and J Glowacki, Age-Related Decline in Osteoprotegerin Expression by Human Bone Marrow Cells Cultured in Three-Dimensional Collagen Sponges. *Biochemical and Biophysical Research Communications* (2000), 268: 669-672
7. Honore P et al., Osteoprotegerin blocks bone cancer-induced skeletal destruction, skeletal pain and pain-related neurochemical reorganization of the spinal cord. *Nature Medicine* (2000), 6(5): 521-528



DRG® Osteoprotegerin (OPG) ELISA (EIA-5155)



As of 17 Mar. 2010 rm (Vers. 1.0)

14 GENERAL NOTES ON THE TEST AND TEST PROCEDURE

The test components which are made of human serum are tested for Australia antigen and HIV and found to be negative.

However, since no test method can offer complete assurance that infectious agents are absent; these reagents should be handled as recommended for any potentially infectious human serum or blood specimen. The normal precautions for laboratory working should be observed.

Kit reagents contain sodium azide or thimerosal as bactericides. Sodium azide and thimerosal are toxic. Substrates for the enzymatic color reactions are toxic and carcinogenic. Avoid contact with skin or mucous membranes.

For research use only.

Reagents should not be used beyond the expiration date shown on the kit label.

Do not mix different lot numbers of any kit component.











Quality control guidelines should be followed.











Incubation time, incubation temperature and pipetting volumes of the components are defined by the producer. Any variation of the test procedure, which is not coordinated with the producer, may influence the results of the test. DRG can therefore not be held responsible for any damage resulting from wrong use.

Warranty claims and complaints in respect of deficiencies must be logged within 14 days after receipt of the product. The product should be send to DRG along with a written complaint.

As of 17 Mar. 2010 rm (Vers. 1.0)

SYMBOLS USED WITH DRG ASSAYS

Symbol	English	Deutsch	Français	Español	Italiano
	Consult instructions for use	Gebrauchsanweisung beachten	Consulter les instructions d'utilisation	Consulte las instrucciones de uso	Consultare le istruzioni per l'uso
	European Conformity	CE-Konformitätskennzeichnung	Conformité aux normes européennes	Conformidad europea	Conformità europea
	In vitro diagnostic device	In-vitro-Diagnostikum	Usage Diagnostic in vitro	Para uso Diagnóstico in vitro	Per uso Diagnostica in vitro
	For research use only	Nur für Forschungszwecke	Seulement dans le cadre de recherches	Sólo para uso en investigación	Solo a scopo di ricerca
	Catalogue number	Katalog-Nr.	Numéro de catalogue	Número de catálogo	Numero di Catalogo
	Lot. No. / Batch code	Chargen-Nr.	Numéro de lot	Número de lote	Numero di lotto
	Contains sufficient for <n> tests/	Ausreichend für "n" Ansätze	Contenu suffisant pour "n" tests	Contenido suficiente para <n> ensayos	Contenuto sufficiente per "n" saggi
	Storage Temperature	Lagerungstemperatur	Température de conservation	Temperatura de conservación	Temperatura di conservazione
	Expiration Date	Mindesthaltbarkeits-datum	Date limite d'utilisation	Fecha de caducidad	Data di scadenza
	Legal Manufacturer	Hersteller	Fabricant	Fabricante	Fabbricante
Distributed by	Distributor	Vertreiber	Distributeur	Distribuidor	Distributore
Content	Content	Inhalt	Conditionnement	Contenido	Contenuto
Volume/No.	Volume / No.	Volumen/Anzahl	Volume/Quantité	Volumen/Número	Volume/Quantità

Symbol	Portugues	Dansk	Svenska	Ελληνικά
	Consulte as instruções de utilização	Se brugsanvisning	Se bruksanvisningen	Εγχειρίδιο χρήστη
	Conformidade com as normas europeias	Europaesk overensstemmelse	Europeisk överensstämmelse	Ευρωπαϊκή Συμμόρφωση
	Diagnóstico in vitro	In vitro diagnostik	Diagnostik in vitro	in vitro διαγνωστικό
				
	Catálogo n.º	Katalognummer	Katalog nummer	Αριθμός καταλόγου
	No do lote	Lot nummer	Batch-nummer	Αριθμός Παρτίδος
		Indeholder tilstrækkeligt til "n" test	Innehåller tillräckligt till "n" tester	Περιεχόμενο επαρκές για «n» εξετάσεις
	Temperatura de conservação	Opbevarings-temperatur	Förvaringstemperatur	Θερμοκρασία αποθήκευσης
	Prazo de validade	Udløbsdato	Bäst före datum	Ημερομηνία λήξης
	Fabricante	Producent	Tillverkare	Κατασκευαστής
Distributed by				
Content	Conteúdo	Indhold	Innehåll	Περιεχόμενο
Volume/No.	Volume/Número	Volumen/antal	Volym/antal	Όγκος/αριθ..