

Barbiturates Screen Test (RAP-2867)

INTENDED USE

The LiveSure™ Barbiturates Screen Test Card and Test Strip devices are rapid, in vitro diagnostic (IVD), qualitative, lateral flow, immunochromatographic urinary assays for the detection of barbiturate drugs (e.g., secobarbital, pentobarbital, phenobarbital, etc.), their analogs and their metabolites (collectively termed: BAR) in human urine at the Substance Abuse and Mental Health Services Administration (SAMHSA) cut-off level of 300 ng/ml. The tests are designed to obtain a visual, qualitative result and are intended for professional use only. They are not intended for quantitative analysis, or for over-the-counter sale. The LiveSure™ Barbiturates Screen Test provides only preliminary analytical data. A more specific, alternative method is required to obtain a confirmed analytical result. SAMHSA has established gas chromatography / mass spectrometry (GC/MS) as the preferred confirmatory method. Clinical considerations and professional judgment should be applied to any drug of abuse test result, particularly when preliminary positive results are indicated.

SUMMARY AND EXPLANATION OF TEST

Barbiturates (secobarbital, pentobarbital, phenobarbital, etc.), their analogs and their metabolites (BAR) are central nervous system depressants. Barbiturates are usually taken orally, rectally, or by intravenous or intramuscular injection. High dose abuse of barbiturates may impair mental and motor coordination, causing confusion, injury and mental disorders, and may lead to respiratory collapse, coma and even death. Short-acting barbiturates generally are excreted as metabolites of pentobarbital and secobarbital (e.g., barbital) within 12 to 24 hours, while the long acting barbiturates primarily appear unchanged (e.g., phenobarbital) in urine within 12 to 48 hours. Barbiturates, their analogs, and their metabolites (BAR) are the primary urine markers for detecting barbiturate use or abuse. The LiveSure™ Barbiturates Screen Tests are rapid, visual, lateral flow, competitive IVD immunochromatographic assays for the qualitative detection of BAR in human urine. These IVD, qualitative screening tests are based on immunoassay principles and designed specifically for the assay and identification of BAR at the cut-off level of 300 ng/ml or higher, as set by SAMHSA.

PRINCIPLE OF THE PROCEDURE

The LiveSure™ Barbiturates Screen Test is a competitive, IVD, immunochromatographic urinary assay in which chemically modified and bound barbiturate conjugate competes with urinary BAR for limited specific barbiturate antibody binding sites. Each test device contains a sample reaction unit, a pink-colored, antibody-colloidal gold conjugate unit pre-labeled with specific antibody, and a chromatographic membrane pre-coated with the drug conjugate in the device test region. When free drug is absent from urine or at concentrations lower than the urinary detection cut-off level, the pink-colored, antibody-colloidal gold conjugate is free to bind to the immobilized drug conjugate pre-coated on the chromatographic membrane test region. Consequently, a pink-colored band will form in the test region indicating a negative result. When drug is present in urine at concentrations at or exceeding the detection cut-off level, it will bind to limited epitopes on the pink-colored, antibody-colloidal gold conjugate, completely competing with drug conjugate in the test region. In such cases, no band forms in the test region, indicating a positive result. The LiveSure™ Barbiturates Screen Test also provides a built-in process control. A pink-colored band should always appear in the control region, regardless of the presence of any urinary barbiturates, their analogs and/or their metabolites. This pink-colored control band verifies that: 1) sufficient urine volume was added, and 2) proper flow was obtained. If the control band is missing, the test was not performed correctly or failed to function correctly. In summary, negative urine will produce two pink-colored bands, one in the control region and one in the test region. Positive urine will produce one pink-colored band in the control region only.

REAGENTS AND MATERIALS SUPPLIED

1. Test device with the following components: a sample reaction unit, a pink-colored, colloidal gold conjugate unit, and a chromatographic membrane unit. A disposable sample pipette comes with the test card.
2. Test instructions

Barbiturates Screen Test (RAP-2867)

MATERIALS REQUIRED BUT NOT PROVIDED

1. Urine sample collection containers
2. Timer or clock

WARNINGS AND PRECAUTIONS

1. For in vitro diagnostic use.
2. Avoid cross contamination of urine samples by using a new sample collection container and pipette for each urine sample.
3. Urine specimens may be potentially infectious. Avoid contact with skin by wearing gloves and proper laboratory attire. Properly handle and dispose of all used test devices in an approved biohazard container.
4. Do not use a test device beyond the expiration date imprinted on the outside of the foil pouch.

STORAGE AND STABILITY

The test devices supplied can be stored at room temperature or refrigerated (2-28°C) and will remain stable until the expiration date. **Do Not Freeze.**

SAMPLE COLLECTION AND PREPARATION

The fresh urine samples should be collected in a clean, dry container such that same-day testing may be performed. Urine specimens may be refrigerated at 2-8°C for 48 hours, or frozen at -20°C for assaying at a later date. Specimens that have been refrigerated or frozen must be equilibrated to room temperature before testing. Urine samples exhibiting visible precipitates should be filtered, centrifuged or allowed to settle so that clear aliquots can be obtained for testing.

ASSAY PROCEDURE

1. Do not open foil pouch until ready to begin testing. Refrigerated test device should be allowed to equilibrate to room temperature (15-28°C) before pouch is opened.
2. Remove the test device from the sealed foil pouch by tearing along the notch.
3. **FOR TEST STRIP:** Immerse the strips in urine with the arrow end pointing towards the urine. Do not introduce urine above the maximum level, as indicated by the arrows.
FOR TEST CARD: Draw the urine sample into the pipette and dispense four drops (approximately 0.2ml) into the sample well of test device.
4. Read the test result at five minutes.

IMPORTANT

In order to prevent an incorrect reading, do not read the test results after more than 10 minutes. After 10 minutes, the intensity of the colored lines may change or a new line may appear. To avoid confusion, discard the test device after interpreting the result.

INTERPRETATION OF RESULTS

Negative: Two pink-colored bands appear, one in the control region and one in the test region. A negative result indicates free drug is absent from urine or at concentrations lower than the detection cut-off level of the test.


Positive: One pink-colored band appears in the control region with no apparent band appearing in the test region. A positive result indicates free drug is present in urine at concentrations at or exceeding the detection cut-off level of the test.

Invalid: No band appears in the control region, or a pink-colored band appears in the test region only. An invalid test result may be due to improper assay procedures or damage to the device. With an invalid result, the assay is inconclusive and the specimen should be re-tested using a new test device.


Note: The test band intensity may be weaker or stronger than that of the control band.

Barbiturates Screen Test (RAP-2867)

TEST STRIPS



TEST CARDS



PROCESS AND QUALITY CONTROL

Good laboratory practice requires use of control materials that ensure proper test device performance and reliability. Quality control standards are available for the validation of device functionality from commercial sources such as BioRad, Alltech and Biopool. To test the quality of control standards, use the assay procedure for testing urine samples. The SAMHSA recommended guidelines for drugs of abuse screening test devices indicate that controls should contain the drug of abuse analyte at levels at least 25% above cut-off values.

AFTER TESTING

Urine specimens may be potentially infectious. Avoid contact with skin by wearing gloves and proper laboratory attire. Properly handle and dispose of all used test devices in an approved biohazard container. Residual urine should be disposed of in a medically approved manner after the completion of all testing, including the confirmatory assay.

PERFORMANCE CHARACTERISTICS

SENSITIVITY

The LiveSure™ Barbiturates Screen Test has been designed for the detection of urinary BAR at the detection sensitivity of 300 ng/ml as suggested for immunoassay methods by SAMHSA. In sensitivity studies performed, samples with concentrations of BAR equal to or lower than 150 ng/ml were identified as negative for all samples. Concentrations of BAR equal to or higher than 450 ng/ml were identified as positive results for all samples. Thus the cut-off level of the LiveSure™ Barbiturates Screen Test was determined to be 300 ng/ml for both the Test Card and Test Strip devices.

PRECISION

In order to determine the precision of both the LiveSure™ Barbiturates Screen Test Card and Test Strip devices, both within-lot and inter-lot reproducibility studies were performed. Results of the within-lot reproducibility studies clearly showed excellent repeatability for all positive and negative urine samples using one lot of LiveSure™ Barbiturates Screen Test Card devices and one of Test Strip devices. The results of inter-lot reproducibility studies clearly demonstrated that there was no appreciable inter-lot variation when testing both positive and negative samples across three different lots of LiveSure™ Barbiturates Screen Test Card and Test Strip devices.

SPECIFICITY STUDY

Device specificity was tested against compounds related to, or not associated with, barbiturates, prepared in drug-free, normal human urine. The following compounds produced positive results with the LiveSure™ Barbiturates Screen Test at levels at or greater than the concentrations listed below.

Barbiturates Screen Test (RAP-2867)

REACTIVE BAR ANALOGS, METABOLITES OR OTHER SUBSTANCES

<i>Alphenal</i>	300 ng/ml
<i>Aprobarbital</i>	300 ng/ml
<i>Barbital</i>	300 ng/ml
<i>Bromocriptine</i>	300 ng/ml
<i>Butabarbital</i>	300 ng/ml
<i>Butethal</i>	300 ng/ml
<i>Pentobarbital</i>	300 ng/ml
<i>Phenobarbital</i>	300 ng/ml
<i>Secobarbital</i>	300 ng/ml
<i>Zoloft</i>	300 ng/ml

LESS REACTIVE BAR ANALOGS, METABOLITES OR OTHER SUBSTANCES

<i>Allobarbital</i>	1,000 ng/ml
<i>Amobarbital</i>	1,000 ng/ml
<i>Sodium Dodecyl Sulfate (SDS)</i>	1,000 ng/ml
<i>Butalbital</i>	2,000 ng/ml
<i>Ketamine</i>	10,000 ng/ml
<i>D-Lysergic Acid Diethylamide(LSD)</i>	100,000 ng/ml

The following compounds were found not to have any impact on negative results when tested at levels of up to 100 µg drug/ml (100,000 ng drug/ml).

NON-REACTIVE BAR ANALOGS, METABOLITES OR OTHER SUBSTANCES

Acetaminophen (N-Acetyl-p-aminophenol)	Guaiacol glyceryl ether
Acetone	Hemoglobin
Albumin	Hydromorphone
Amitriptyline	Imipramine
Amphetamine	(+/-)-Isoproterenol
Ampicillin	(1R,2S)-(-)-N-Methyl-Ephedrine
Aspartame (Asp-Phe Methyl Ester)	Lidocaine
Aspirin (Acetylsalicylic Acid)	Meperidine
Atropine	Methadone
Benzocaine (Ethyl p-Aminobenzoate)	Methamphetamine
Benzoyllecgonine	Morphine
Bilirubin	(+)-Naproxen
Caffeine	(+/-)-Norephedrine
Chloroquine	Oxalic Acid
(+)-Chlorpheniramine	Oxazepam
(+/-)-Chlorpheniramine	Penicillin-G (Benzylpenicillin)
Cocaine	Pheniramine
Codeine	Phenothiazine (Thiodiphenylamine)
Creatine	L-Phenylephrine
Dexbrompheniramine ([+]-Brompheniramine)	-Phenylethylamine
Dextromethorphan	([+/-]-PhenylproDRGolamine)
4-Dimethylaminoantipyrine	Procaine

Barbiturates Screen Test (RAP-2867)

Dopamine (3-Hydroxytyramine)	Promethazine
Doxylamine	Quinidine
(+/-)-Ephedrine	Ranitidine
(-)-Ephedrine	Sulindac
(+)-Epinephrine	Thioridazine
Erythromycin	Trifluoperazine
Ethanol	Trimethobenzamide
Furosemide	Tyramine
Glucose	Vitamin C (L-Ascorbic Acid)

ASSAY COMPARISONS & EQUIVALENCY

Accuracy and equivalency comparisons of LiveSure™ Barbiturates Screen Test Cards and Test Strips were evaluated against 144 individual in-house laboratory urine samples, as well as against 292 individual external SAMHSA-certified clinical laboratory urine samples. Results have been tabulated below.

Table 1. LiveSure™ Barbiturates Card vs. Emit® II

LiveSure™	EMIT II (+)	EMIT II (-)	Row Totals
(+)	144	1	145
(-)	0	291	291
Col. Totals	144	292	436

When compared to Emit® II, the percent agreement with LiveSure™ Barbiturates Screen Test Card positive samples WAS 144/144 or 100%. Negative samples recovered at 291/292 or 99.7%, while the overall relative accuracy obtained was 435/436 or 99.8%.

Table 2. LiveSure™ Barbiturates Strip vs. Emit® II

LiveSure™	EMIT II (+)	EMIT II (-)	Row Totals
(+)	144	2	146
(-)	0	290	290
Col. Totals	144	292	436

When compared to Emit® II, the percent agreement with LiveSure™ Barbiturates Screen Test Strip positive samples was 144/144 or 100%. Negative samples recovered at 290/292 or 99.3%, while the overall relative accuracy obtained was 434/436 or 99.5%.

Table 3. LiveSure™ Barbiturates Card vs. GC/MS

LiveSure™	GC/MS (+)	GC/MS (-)	Row Totals
(+)	104	0	104
(-)	0	231	231

When compared to the GC/MS data, the relative sensitivity or percent agreement of LiveSure™ Barbiturates Screen Test Card positive samples with the external clinical study was 104/104 or 100%. Negative samples recovered a relative specificity of agreement of 231/231 or 100%. The overall relative accuracy obtained was 335/335 or 100%.

Table 4. LiveSure™ Barbiturates Strip vs. GC/MS

LiveSure™	GC/MS (+)	GC/MS (-)	Row Totals
(+)	104	1	105
(-)	0	230	230
Col. Totals	104	231	335

Barbiturates Screen Test (RAP-2867)

When compared to the GC/MS data, the relative sensitivity or percent agreement of LiveSure™ Barbiturates Screen Test Strip positive samples with the external clinical study was 104/104 or 100%. Negative samples recovered a relative specificity of agreement of 230/231 or 99.6%. Finally, the overall relative accuracy obtained was 334/335 or 99.7%.

LIMITATIONS OF PROCEDURE

1. This diagnostic test is designed for the medical or forensic in vitro detection of barbiturates, their analogs and their metabolites (BAR) in human urine.
2. This diagnostic test only provides a qualitative screening for BAR in urine. It is not to be used for the quantitative determination of BAR in urine. A positive result with this test indicates only the presence of BAR and should be confirmed by GC/MS. It does not indicate intoxication.
3. There is a possibility of false positive results due to user error or the presence of interfering factors in urine.
4. See "SPECIFICITY STUDY" for substances that produce positive results at high concentrations and those substances that do not interfere with test performance.
5. Adulterants, such as bleach or other strong oxidizing agents, may produce erroneous test results when added to urine specimens, regardless of the analysis method used. If an adulteration is suspected, a fresh urine specimen should be used.

BIBLIOGRAPHY

1. Baselt, R. C. *Disposition of Toxic Drugs and Chemicals in Man*, Biomedical Publications, Davis, CA, 1982.
2. *Urine Testing for Drugs of Abuse*. National Institute on Drug Abuse (NIDA), Research Monograph 73, 1986.
3. Fed. Register, Department of Health and Human Services, *Mandatory Guidelines for Federal Workplace Drug Testing Programs*, 53, 69, 11970-11979, 1988.
4. McBay, A. J. Clin. Chem. 33, 33B-40B, 1987.
5. Gilman, A. G., and Goodman, L. S. *The Pharmacological Basis of Therapeutics*, eds. MacMillan Publishing, New York, NY, 1980.