

DRG[®] Human sIL-6R ELISA (EIA-4423)

REVISED 17 AUG. 2007

RUO

1 INTENDED USE

The DRG Human Soluble Interleukin-6 Receptor (hsIL-6R) (80 kDa) ELISA is to be used for the in vitro quantitative determination of hsIL-6R in human serum, plasma, buffered solution, or cell culture medium. The assay will recognize both natural and recombinant hsIL-6R.

**This kit has been configured for research use only and is not to be used in diagnostic procedures.
Read entire protocol before use.**

2 PRINCIPLE OF THE METHOD

The DRG hsIL-6R kit is a solid phase sandwich Enzyme Linked-Immuno-Sorbent Assay (ELISA). An antibody specific for hsIL-6R has been coated onto the wells of the microtiter strips provided. Samples, including standards of known hsIL-6R content, control specimens, and unknowns, are pipetted into these wells, followed by the addition of a biotinylated second antibody. During the first incubation, the hsIL-6R antigen binds simultaneously to the immobilized (capture) antibody on one site, and to the solution phase biotinylated antibody on a second site. After removal of excess second antibody, Streptavidin-Peroxidase (enzyme) is added. This binds to the biotinylated antibody to complete the four-member sandwich. After a second incubation and washing to remove all the unbound enzyme, a substrate solution is added, which is acted upon by the bound enzyme to produce color. The intensity of this colored product is directly proportional to the concentration of hsIL-6R present in the original specimen.

3 REAGENTS PROVIDED

Note: Store all reagents at 2 - 8°C.

Reagent	96 Test Kit
hsIL-6R Standard , recombinant hsIL-6R. Refer to vial label for quantity and reconstitution volume.	2 vials
Standard Diluent Buffer . Contains 15 mM sodium azide; 125 mL per bottle.	1 bottle
hsIL-6R Antibody-Coated Wells , 96 wells per plate.	1 plate
sIL-6R Biotin Conjugate , (Biotin-labeled anti-sIL-6R). Contains 15 mM sodium azide; 6 mL per bottle.	1 bottle
Streptavidin-Peroxidase (HRP), (100x) concentrate . Contains 3.3 mM thymol; 0.125 mL per vial.	1 vial
Streptavidin-Peroxidase (HRP) Diluent . Contains 3.3 mM thymol; 25 mL per bottle.	1 bottle
Wash Buffer Concentrate (25x) ; 100 mL per bottle.	1 bottle
Stabilized Chromogen , Tetramethylbenzidine (TMB); 25 mL per bottle.	1 bottle
Stop Solution ; 25 mL per bottle.	1 bottle
Plate Covers , adhesive strips.	3

Disposal Note: This kit contains materials with small quantities of sodium azide. Sodium azide reacts with lead and copper plumbing to form explosive metal azides. Upon disposal, flush drains with a large volume of water to prevent

DRG[®] Human sIL-6R ELISA (EIA-4423)

REVISED 17 AUG. 2007

RUO

azide accumulation. Avoid ingestion and contact with eyes, skin and mucous membranes. In case of contact, rinse affected area with plenty of water. Observe all federal, state and local regulations for disposal.

4 SUPPLIES - NOT PROVIDED

1. Microtiter plate reader capable of measurement at or near 450 nm.
2. Calibrated adjustable precision pipettes, preferably with disposable plastic tips. (A manifold multi-channel pipette is desirable for large assays.)
3. Deionized or distilled H₂O.
4. Plate washer: automated or manual (squirt bottle, manifold dispenser, etc.).
5. Graph paper: linear (Cartesian), log-log, or semilog, as desired.
6. Glass or plastic tubes for diluting and aliquoting standard.
7. Absorbent paper towels.
8. Calibrated beakers and graduated cylinders in various sizes.

5 PROCEDURAL NOTES/LAB QUALITY CONTROL

1. When not in use, kit components should be refrigerated. All reagents should be warmed to room temperature before use.
2. **Microtiter plates should be allowed to come to room temperature before opening the foil bags.** Once the desired number of strips has been removed, immediately reseal the bag and store at 2 - 8°C to maintain plate integrity.
3. Samples should be collected in pyrogen/endotoxin-free tubes.
4. Samples should be frozen if not analyzed shortly after collection. Avoid multiple freeze-thaw cycles of frozen samples. Thaw completely and mix well prior to analysis.
5. When possible, avoid use of badly hemolyzed or lipemic sera. If large amounts of particulate matter are present, centrifuge or filter prior to analysis.
6. It is recommended that all standards, controls and samples be run in duplicate.
7. Samples that are >4000 pg/mL should be diluted with Standard Diluent Buffer.
8. When pipetting reagents, maintain a consistent order of addition from well-to-well. This ensures equal incubation times for all wells.
9. Cover or cap all reagents when not in use.
10. Do not mix or interchange different reagent lots from various kit lots.
11. Do not use reagents after the kit expiration date.
12. Read absorbances within 2 hours of assay completion.
13. In-house controls should be run with every assay. If control values fall outside pre-established ranges, the accuracy of the assay is suspect.
14. All residual wash liquid must be drained from the wells by efficient aspiration or by decantation followed by tapping the plate forcefully on absorbent paper. Never insert absorbent paper directly into the wells.
15. Because Stabilized Chromogen is light sensitive, avoid prolonged exposure to light. Also avoid contact between Stabilized Chromogen and metal, or color may develop.

SAFETY

All blood components and biological materials should be handled as potentially hazardous. Follow universal precautions as established by the Centers for Disease Control and Prevention and by the Occupational Safety and Health Administration when handling and disposing infectious agents.

DRG® Human sIL-6R ELISA (EIA-4423)

REVISED 17 AUG. 2007

RUO

6 DIRECTIONS FOR WASHING

Incomplete washing will adversely affect the test outcome. All washing must be performed with Wash Buffer provided.

Washing can be performed manually as follows:

completely aspirate the liquid from all wells by gently lowering an aspiration tip (aspiration device) into the bottom of each well. Take care not to scratch the inside of the well. After aspiration, fill the wells with at least 0.4 mL of diluted wash solution. Let soak for 15 to 30 seconds, then aspirate the liquid. Repeat as directed under ASSAY METHOD. After the washing procedure, the plate is inverted and tapped dry on absorbent tissue.

Alternatively, the wash solution may be put into a squirt bottle. If a squirt bottle is used, flood the plate with wash buffer, completely filling all wells. After the washing procedure, the plate is inverted and tapped dry on absorbent tissue.

If using an automated washer, the operating instructions for washing equipment should be carefully followed.

7 REAGENT PREPARATION AND STORAGE

7.1 Reconstitution and Dilution of hsIL-6R Standard

Note: Either glass or plastic tubes may be used for standard dilutions.

1. Reconstitute standard to 10,000 pg/mL with Standard Diluent Buffer. Refer to standard vial label for instructions. Swirl or mix gently and allow to sit for 10 minutes to ensure complete reconstitution. Use standard within 1 hour of reconstitution.
2. Add 0.150 mL of the reconstituted standard to a tube containing 0.225 mL Standard Diluent Buffer. Label as 4000 pg/mL hsIL-6R. Mix.
3. Add 0.125 mL of Standard Diluent Buffer to each of 6 tubes labeled 2000, 1000, 500, 250, 125 and 62.5 pg/mL hsIL-6R.
4. Make serial dilutions of the standard as described in the following dilution table. Mix thoroughly between steps.

7.2 Dilution of hsIL-6R Standard

Standard:	Add:	Into:
4000 pg/mL	Prepare as described in Step 2.	
2000 pg/mL	0.125 mL of the 4000 pg/mL std.	0.125 mL of the Diluent Buffer
1000 pg/mL	0.125 mL of the 2000 pg/mL std.	0.125 mL of the Diluent Buffer
500 pg/mL	0.125 mL of the 1000 pg/mL std.	0.125 mL of the Diluent Buffer
250 pg/mL	0.125 mL of the 500 pg/mL std.	0.125 mL of the Diluent Buffer
125 pg/mL	0.125 mL of the 250 pg/mL std.	0.125 mL of the Diluent Buffer
62.5 pg/mL	0.125 mL of the 125 pg/mL std.	0.125 mL of the Diluent Buffer
0 pg/mL	0.125 mL of the Diluent Buffer	An empty tube

Discard all remaining reconstituted and diluted standards after completing assay.

Return the Standard Diluent Buffer to the refrigerator.

DRG[®] Human sIL-6R ELISA (EIA-4423)

REVISED 17 AUG. 2007

RUO

7.3 Storage and Final Dilution of Streptavidin-HRP

Please Note: The Streptavidin-HRP 100x concentrate is in 50% glycerol. This solution is viscous. To ensure accurate dilution, allow Streptavidin-HRP concentrate to reach room temperature. Gently mix. Pipette Streptavidin-HRP concentrate slowly. Remove excess concentrate solution from pipette tip by gently wiping with clean absorbent paper.

1. Dilute 10 μ L of this 100x concentrated solution with 1 mL of Streptavidin-HRP Diluent for each 8-well strip used in the assay. Label as Streptavidin-HRP Working Solution. For Example:

# of 8-Well Strips	Volume of Streptavidin-HRP Concentrate	Volume of Diluent
2	20 μ L solution	2 mL
4	40 μ L solution	4 mL
6	60 μ L solution	6 mL
8	80 μ L solution	8 mL
10	100 μ L solution	10 mL
12	120 μ L solution	12 mL

2. Return the unused Streptavidin-HRP concentrate to the refrigerator.

7.4 Dilution of Wash Buffer

Allow the 25x concentrate to reach room temperature and mix to ensure that any precipitated salts have redissolved. Dilute 1 volume of the 25x wash buffer concentrate with 24 volumes of deionized water (e.g., 50 mL may be diluted up to 1.25 liters, 100 mL may be diluted up to 2.5 liters).

Label as Working Wash Buffer.

Store both the concentrate and the Working Wash Buffer in the refrigerator.

The diluted buffer should be used within 14 days.

8 ASSAY METHOD: PROCEDURE AND CALCULATIONS

Be sure to read the Procedural Notes/Lab Quality Control section before carrying out the assay.

Allow all reagents to reach room temperature before use. Gently mix all liquid reagents prior to use.

Note: A standard curve must be run with each assay.

1. Determine the number of 8-well strips needed for the assay. Insert these in the frame(s) for current use. (Re-bag extra strips and frame. Store these in the refrigerator for future use.)
2. **Dilute** serum or plasma samples **1:100** with Standard Diluent Buffer (e.g., add 20 μ L of sample to 1980 μ L Standard Diluent Buffer). Please Note: Individual samples may require a greater or lesser dilution to fall within the range of the assay.
3. Add 50 μ L of the Standard Diluent Buffer to zero wells. Well(s) reserved for chromogen blank should be left empty.
4. Add 50 μ L of standards, samples or controls to the appropriate microtiter wells. (See REAGENT PREPARATION AND STORAGE, Section 7.2.)
5. Pipette 50 μ L of biotinylated anti-sIL-6R (Biotin Conjugate) solution into each well except the chromogen blank(s). Tap gently on the side of the plate to mix.

DRG[®] Human sIL-6R ELISA (EIA-4423)

REVISED 17 AUG. 2007

RUO

6. Cover plate with plate cover and **incubate for 2 hours at room temperature.**
7. Thoroughly aspirate or decant solution from wells and discard the liquid. Wash wells 4 times.
See DIRECTIONS FOR WASHING.
8. Add 100 μ L Streptavidin-HRP Working Solution to each well except the chromogen blank(s). (Prepare the working dilution as described in REAGENT PREPARATION AND STORAGE, Section 7.3.)
9. Cover plate with the plate cover and **incubate for 30 minutes at room temperature.**
10. Thoroughly aspirate or decant solution from wells and discard the liquid. Wash wells 4 times.
See DIRECTIONS FOR WASHING.
11. Add 100 μ L of Stabilized Chromogen to each well. The liquid in the wells will begin to turn blue.
12. **Incubate for 30 minutes at room temperature and in the dark. Please Note: Do not cover the plate with aluminum foil or metalized mylar.** The incubation time for chromogen substrate is often determined by the microtiter plate reader used. Many plate readers have the capacity to record a maximum optical density (O.D.) of 2.0. The O.D. values should be monitored and the substrate reaction stopped before the O.D. of the positive wells exceed the limits of the instrument. The O.D. values at 450 nm can only be read after the Stop Solution has been added to each well. If using a reader that records only to 2.0 O.D., stopping the assay after 20 to 25 minutes is suggested.
13. Add 100 μ L of Stop Solution to each well. Tap side of plate gently to mix. The solution in the wells should change from blue to yellow.
14. Read the absorbance of each well at 450 nm having blanked the plate reader against a chromogen blank composed of 100 μ L each of Stabilized Chromogen and Stop Solution.
Read the plate within 2 hours after adding the Stop Solution.
15. Plot on graph paper the absorbance of the standards against the standard concentration. (Optimally, the background absorbance may be subtracted from all data points, including standards, unknowns and controls, prior to plotting.) Draw the best smooth curve through these points to construct the standard curve. If using curve fitting software, the four parameter algorithm provides the best curve fit.
16. Read the hsIL-6R concentrations for unknown samples and controls from the standard curve plotted in step 15.
Multiply value(s) obtained for serum or plasma sample(s) by 100 to correct for the 1:100 dilution in step 2. (Samples producing signals greater than that of the highest standard (4000 pg/mL) should be diluted in Standard Diluent Buffer and re-analyzed, multiplying the concentration found by the appropriate dilution factor.)

DRG[®] Human sIL-6R ELISA (EIA-4423)

REVISED 17 AUG. 2007

RUO

TYPICAL DATA

The following data were obtained for the various standards over the range of 0 to 4000 pg/mL hsIL-6R.

Standard hsIL-6R (pg/mL)	Optical Density (450 nm)
0	0.058 0.061
62.5	0.109 0.117
125	0.151 0.160
250	0.240 0.253
500	0.372 0.398
1000	0.728 0.651
2000	1.336 1.153
4000	2.639 2.512

9 LIMITATIONS OF THE PROCEDURE

Do not extrapolate the standard curve beyond the 4000 pg/mL standard point; the dose-response is non-linear in this region and accuracy is difficult to obtain. Dilute samples >4000 pg/mL with Standard Diluent Buffer; re-analyze these and multiply results by the appropriate dilution factor.

The influence of various drugs, aberrant sera (hemolyzed, hyperlipidemic, jaundiced, etc.) and the use of biological fluids in place of serum samples have not been thoroughly investigated. The rate of degradation of native hsIL-6R in various matrices has not been investigated. The immunoassay literature contains frequent references to aberrant signals seen with some sera, attributed to heterophilic antibodies. Though such samples have not been seen to date, the possibility of this occurrence cannot be excluded.

This kit is for research use only.

Not for human therapeutic or diagnostic use.

10 PERFORMANCE CHARACTERISTICS

10.1 SENSITIVITY

The minimum detectable dose of hsIL-6R is <8 pg/mL. This was determined by adding two standard deviations to the mean O.D. obtained when the zero standard was assayed 30 times.

DRG[®] Human sIL-6R ELISA (EIA-4423)

REVISED 17 AUG. 2007

RUO

10.2 PRECISION

10.2.1 Intra-Assay Precision

Samples of known hsIL-6R concentration were assayed in replicates of 16 to determine precision within an assay.

	Sample 1	Sample 2	Sample 3
Mean (pg/mL)	314.2	885.0	1878.0
SD	15.7	51.5	99.8
%CV	5.0	5.8	5.3

SD = Standard Deviation, CV = Coefficient of Variation

10.2.2 Inter-Assay Precision

Samples were assayed 48 times in multiple assays to determine precision between assays.

	Sample 1	Sample 2	Sample 3
Mean (pg/mL)	306.0	978.6	1976.1
SD	17.2	92.0	133.4
%CV	5.6	9.4	6.8

SD = Standard Deviation, CV = Coefficient of Variation

10.3 LINEARITY OF DILUTION

A human serum pool containing 4000 pg/mL of measured hsIL-6R was serially diluted in Standard Diluent Buffer over the range of the assay.

Linear regression analysis of samples versus the expected concentration yielded a correlation coefficient of 0.99.

10.4 RECOVERY

The recovery of hsIL-6R added to diluted human serum and plasma averaged 102% (range: 90% to 108%). The recovery of hsIL-6R added to tissue culture medium containing 1% fetal calf serum averaged 108%, while the recovery of hsIL-6R added to tissue culture medium containing 10% fetal calf serum averaged 98%.

10.5 SPECIFICITY

Buffered solutions of a panel of substances at 10,000 pg/mL were assayed with the DRG hsIL-6R kit.

The following substances were tested and found to have no cross-reactivity:

- human IL-1 β , IL-2, IL-3, IL-4, IL-6, IL-7, IL-8, IL-10, IL-13, TNF- α , SCF, IFN- γ ;
- mouse IL-1b, IL-2, IL-4, IL-6, IL-10, TNF- α , IFN- γ ;
- rat TNF- α , MIP-2.