

Sample Preparation	Blood collection																		
	Tissue extraction																		
	CSF extraction																		
Catalog #	EK-033-20																		
Standard Size	96 wells																		
Sensitivity	Minimum Detectable Concentration = 0.78ng/ml																		
Range	0.78-50 ng/ml																		
Cross Reactivity	<table><tr><th>Peptide</th><th>% Cross-reactivity</th></tr><tr><td>IGFBP-3 (Human)</td><td>100</td></tr><tr><td>IGFBP-1 (Human)</td><td>0</td></tr><tr><td>IGFBP-2 (Human)</td><td>0</td></tr><tr><td>IGF-I (Human)</td><td>0</td></tr><tr><td>IGF-II (Human)</td><td>0</td></tr><tr><td>IGF-I (Mouse)</td><td>0</td></tr><tr><td>IGF-II (Mouse)</td><td>0</td></tr><tr><td>Insulin (Human)</td><td>0</td></tr></table>	Peptide	% Cross-reactivity	IGFBP-3 (Human)	100	IGFBP-1 (Human)	0	IGFBP-2 (Human)	0	IGF-I (Human)	0	IGF-II (Human)	0	IGF-I (Mouse)	0	IGF-II (Mouse)	0	Insulin (Human)	0
Peptide	% Cross-reactivity																		
IGFBP-3 (Human)	100																		
IGFBP-1 (Human)	0																		
IGFBP-2 (Human)	0																		
IGF-I (Human)	0																		
IGF-II (Human)	0																		
IGF-I (Mouse)	0																		
IGF-II (Mouse)	0																		
Insulin (Human)	0																		

Standard Curve Image

