



LDN Labor Diagnostika Nord GmbH & Co. KG

Am Eichenhain 1, 48531 Nordhorn

Telefon: +49-5921-8197 0

Telefax: +49-5921-8197 222

e-mail: info@ldn.de

Internet: <http://www.ldn.de>

LDN[®]

Instructions for use

PAP ELISA

REF

TM E-4200



IVD



PAP ELISA

NAME AND INTENDED USE

The PAP ELISA is a solid phase enzyme-linked immuno-sorbent assay. This test provides quantitative measurement of human prostatic acid phosphatase (PAP) in serum.
(For Professional Use Only)

SUMMARY AND EXPLANATION OF TEST

The phosphatases include two main types. The alkaline phosphatase has a pH optimum of about 9 while acid phosphatase (AcP) has its optimal activity at a pH of about 5^{1, 2}. The AcP, first demonstrated in the urine in 1925, was found to be much more prevalent in male than in female urine². Prostatic acid phosphatase (PAP) enzyme activity was found to be localized in organs of the male genital tract^{2, 3}. The presence of PAP in sera of patients with prostatic cancer was first demonstrated by Gutman⁴. It was shown that PAP concentration was elevated in many men with primary prostatic carcinoma and metastatic lesions of the prostate^{4, 5}.

The release of PAP into the blood stream by primary or metastatic carcinoma of the prostate may provide the clinician with a means of following remission or relapse of a prostatic malignancy^{5, 6}. Therefore, sensitive and accurate measurement of serum PAP is essential in monitoring the effectiveness of various therapeutic treatments^{5, 9}.

Because moderate elevation of serum PAP is observed to accompany many nonprostatic diseases, a specific assay for determination of PAP is established⁷. The current methodologies for determination of PAP activity using the colorimetric methods^{8, 9} and immunochemical methods¹⁰ have many limitations. PAP ELISA provides an enzyme immunoassay system and establishes an ELISA method for quantitative measurement of PAP in human serum¹¹.

PRINCIPLE OF THE ASSAY

PAP ELISA is a solid phase enzyme linked immunosorbent assay (ELISA). The wells are coated with specific anti-PAP antibodies. The Samples, Standards and Controls are incubated in the wells and bound with enzyme conjugate which is a mixture of anti-PAP antibodies with different affinity toward epitopes of PAP molecule and chemically conjugated with horseradish peroxidase. Unbound enzyme conjugate is washed off. The amount of bound peroxidase is proportional to the concentration of the PAP present in the Samples, Standards and Controls. Upon addition of the chromogen substrate, the intensity of color developed is proportional to the concentration of PAP.

MATERIAL PROVIDED

WELLS 96

TM E-4231 Micro-well Strips:

Rabbit anti-PAP antibodies coated wells, 96 wells.

CONJUGATE

TM E-4240 Enzyme Conjugate (11 ml):

Mouse IgG(anti-PAP) conjugated to horseradish peroxidase

DILUENT

TM E-4260 Sample Diluent or Zero Standard (11 mL).

Standards

	Cat. no.	Standard	Concentration	Volume/Vial
STANDARD A	TM E-4201	Standard 1	0 ng/ml	0.75 ml
STANDARD B	TM E-4202	Standard 2	1 ng/ml	0.75 ml
STANDARD C	TM E-4203	Standard 3	3 ng/ml	0.75 ml
STANDARD D	TM E-4204	Standard 4	5 ng/ml	0.75 ml
STANDARD E	TM E-4205	Standard 5	15 ng/ml	0.75 ml
STANDARD F	TM E-4205	Standard 6	30 ng/ml	0.75 ml

In BSA-containing diluent.

SUBSTRATE

TM E-0055 TMB Solution (11 mL):

Buffer solution containing hydrogen peroxide and TMB.

WASH-CONC

TM E-0030 Wash Buffer concentrate 100X (10 mL):

Prepare working washing solution by adding 10 mL wash buffer concentrate into 990 mL distilled water.

STOP-SOLN

TM E-0080 Stop Solution:

1 vial, 6 mL, ready to use, contains 2 M HCl

Avoid contact with the stop solution. It may cause skin irritations and burns.

Well holder: for securing wells.

Package Insert.

MATERIALS REQUIRED BUT NOT PROVIDED

1. Micro-well reader.
2. Disposal tips and pipettor of 25 µL and 100 µL.
3. Quality controls such as Bio-Rad Lyphocheck were recommended

WARNING AND PRECAUTION

1. PAP ELISA is designed for in Vitro Diagnostic Use only.
2. The Components in the kits are intended for usage as an integral unit. The components from different lots should not be mixed, and not be used beyond expiration date.
3. The material should be used in a designated work area, the bench surface should be cleaned with detergent and the contaminated materials should be disposed properly.
4. Some components have been tested using FDA-approved methods and has been found negative for antibody to human immuno-deficiency virus (HIV-I, HIV-II), antibody to Hepatitis C and Hepatitis B surface antigen (HBsAg). No known test method can offer total assurance that HIV-I, HIV-II, Hepatitis B & C virus or other infectious agents are absent. Handle these reagents as if they were potentially infectious. Information on handling human serum is provided in the CDC/NIH manual ABiosafety in Microbiological and Biomedical Laboratories@ (U.S.A. HHS publication No. (NIH 88-8395.)
5. Avoid microbial contamination of reagents when removing aliquots from the vials.

STORAGE AND STABILITY

1. Store the kits at 2-8 °C in a refrigerator.
2. Keep micro-wells in a dry bag with desiccants.
3. The reagents are stable until expiration of the kit. TMB Solution should be colorless; if the solution turns blue, it must be replaced. Do not expose these reagents to strong light during storage usage. If reagent is turbid or contains precipitates, contact the technical support.

SPECIMEN AND COLLECTION

Collect blood by venipuncture, allow to clot, and separate the serum by centrifugation at room temperature. Serum is required for the PAP ELISA, and do not add sodium azide as preservative. PAP is unstable at 4 °C and 20 °C⁸. Samples should not be stored at room temperature or 4 °C for more than 24 hours. Serum samples are recommended to be frozen for longer storage. Avoid repeated freezing and thawing of serum samples.

PREPARATION FOR ASSAY

1. Bring all reagents and samples to room temperature (24 °C ± 3 °C) and shake gently before beginning the test. Have all reagents and samples ready before the start of the assay. Once the test is begun it must be performed without any interruption to get the most reliable and consistent results.
2. Use new disposable tips for each specimen.

ASSAY PROCEDURE

1.	Secure the desired number of coated wells in the holder.
2.	Dispense 25 µL of standards, controls or serum samples into the appropriate wells.
3.	Dispense 100 µL of enzyme conjugate into wells. Incubate for 30 minutes at room temperature.
4.	Remove incubation mixture and rinse the wells 5 times with wash buffer (300 µL/well /each rinse).
5.	Dispense 100 µL of TMB Solution into each well and incubate for 15 minutes at room temperature.
6.	Stop reaction by adding 50 µL of Stop Solution into each well.
7.	Read O.D. at 450 nm with a microwell reader in 5 minutes .

PROCEDURE NOTE

1. Wash the microwells and remove washing buffer thoroughly.
2. Pipet all reagents and samples into bottom of the well. Vortex-mixing or shaking is not required.
3. Absorbance is a function of the time and temperature of incubations. It is recommended to have reagents, samples and needed wells ready for ensure the equal elapsed time for each pipetting without interruption.
4. For the same reason run no more than 20 patient samples with a set of reference standards in duplicate for each assay.
5. If a serum specimen contains greater than 30 ng/ml of PAP, the sample must be diluted with sample diluent and re-assayed as described in the assay procedure.

QUALITY CONTROL

Good laboratory practices include the use of control specimens to ensure that all reagents and protocols are performing properly. PAP ELISA does not include serum controls. Bio-Rad Lyphochek series control serum is recommended.

CALCULATION OF RESULTS

1. Plot the concentration (X) of each reference standards against its absorbance (Y) on a full logarithmic graph paper.
2. Obtain the PAP value of patient by reference to the standard curve as follows: (These data are for demonstration purpose only and must not be used in place of data generated for each assay).

Well No.	Description (ng/mL)	Absorbance 450 nm	PAP (ng/mL)
A1 B1	0 (Blank)	0.000 0.000	
A2 B2	1	0.102 0.089	
A3 B3	3	0.215 0.241	
A4 B4	5	0.371 0.383	
A5 B5	15	1.019 1.039	
A6 B6	30	2.054 2.060	
A7 B7	PATIENT A	1.653 1.455	23.19

EXPECTED VALUE

1. Normal serum concentrations of prostatic acid phosphatase in men have been reported to range from zero to 5 ng/mL¹². Elevated serum concentrations have been observed frequently in patients with prostatic carcinoma and metastatic lesions of the prostate^{4, 5, 12}.
2. A clinical study of the PAP ELISA was conducted and results are summarized as following:
 - a. Serum samples from 100 apparently healthy men were assayed, and 99% of the observations were less than 3 ng/mL.
 - b. Serum samples from 82 patients with benign Hypertrophy (BPH) were assayed and 79% of observations were less than 3 ng/mL.
 - c. Serum samples from 51 patients with prostatic carcinoma (PC) were assayed and 80% of observations were higher than 3 ng/mL.
3. Elevated PAP levels can be an indication of the presence of prostate cancer, they can also be the result of some other prostate diseases such as BPH; therefore, the test must be used in conjunction with a digital rectal examination (DRE) and that if either test is positive. Confirmatory testing with transrectal ultrasound and biopsy is needed to diagnose prostate cancer. Conversely, low PAP levels do not necessarily indicate an absence of prostate cancer.

LIMITATION

1. For diagnostic purpose, the PAP values should be used as an adjunct to other data available to the physician.
2. The ELISA kit is designed to avoid "hook effect" up to 1000 ng/mL.
3. Samples with PAP level above 30 ng/mL should be diluted to obtain accurate value.

PERFORMANCE CHARACTERISTICS

ACCURACY

Recovery studies were performed by mixing an aliquot of pooled serum and PAP Standards. The PAP values were measured and percentage of recovery determined.

Initial Values ng/mL	Concentration Spiked ng/mL	Expected Value ng/mL	Observed Values ng/mL	Recovery (%)
A: 6.5	5	5.7	5.8	102
B: 6.5	30	18.2	18.1	99
C: 21	5	13	12.9	99
D: 21	30	25.5	25.8	101

PRECISION

Intra-assay and inter-assay coefficient of variation were evaluated at three different pooled serum samples.

Intra-assay	Pool A	Pool B	Pool C
N	10	10	10
Mean (ng/mL)	1.01	17.2	29.1
S.D. (ng/mL)	0.09	0.99	1.8
C.V. (%)	8.9	5.7	6.2

Inter-assay	Pool A	Pool B	Pool C
N	10	10	10
Mean (ng/mL)	1.11	18.7	30.1
S.D. (ng/mL)	0.06	0.98	1.91
C.V. (%)	5.4	5.2	6.3

SPECIFICITY













The test is specific for human PAP.

Cross-reactivity with other serum proteins, human hormones and the tumor makers were not found.

MINIMAL DETECTABLE CONCENTRATION

The minimal detectable concentration of PAP is defined as that concentration of prostatic acid phosphatase which corresponds to the absorbance that is two standard deviations greater than the mean absorbance value of 20 replicate determinations of the zero diluent.

Symbols:

	Storage temperature		Manufacturer		Contains sufficient for <n> tests
	Expiry date		Batch code		For in-vitro diagnostic use only!
	Consult instructions for use		Content		CE labelled
	Caution		Catalogue number		For research use only!