

# High Sensitivity Human Amyloid β42 ELISA

96-Well Plate

Cat. # EZHS42

# High Sensitivity Human Amyloid β42 ELISA Kit 96-Well Plate (Cat. # EZHS42)

Ι.	Intended Use	2
II.	Principles Of Procedure	2
III.	Reagents Supplied	3
IV.	Storage and Stability	4
۷.	Technical Guidelines	4
VI.	Reagent Precautions	5
VII.	Materials Required But Not Provided	5
VIII.	Sample Collection And Storage	5
IX.	Preparation of Reagents	6
Х.	Assay Procedure	8
XI.	Microtiter Plate Arrangement	12
XII.	Interpretation	13
XIII.	Assay Characteristics	13
XIV.	Quality Controls	14
XV.	Troubleshooting Guide	14
XVI.	Replacement Reagents	15
XVII.	Ordering Information	16
XVIII.	References	17

## I. INTENDED USE

Amyloid beta peptides have been implicated in the etiology of Alzheimer's disease. Amyloid beta 40 is the most prominent peptide and Amyloid beta 42 is the neurotoxic form. The Amyloid beta 42/40-ratio (AB ratio) has been reported as a better indicator of the Alzheimer pathology. Millipore's High Sensitivity Human Amyloid  $\beta$ 42 ELISA kit is used for the measurement of Amyloid beta 42 in cerebrospinal fluid, cell culture supernatants, primary neurons and plasma in a 96-well format.

## This kit is for research purpose only.

## II. PRINCIPLES OF PROCEDURE

With this Amyloid  $\beta$  test-kit, the principle of a solid phase enzyme immunoassay, an <u>Enzyme-Linked-Immunos</u>orbent-<u>A</u>ssay (ELISA), for the quantitative analysis of hA $\beta$ 42 is applied. The antigen hA $\beta$ 42 to be tested, is detected by selective monoclonal anti-A $\beta$ -antibodies at two different binding sites (Epitopes), forming a "Sandwich-Complex". The polystyrene surface of the microtiter plate is layered with an antibody (capture antibody) which selectively recognizes the C-terminal end of the antigen. During the test procedure, a selective anti-A $\beta$ -antibody conjugate (detection antibody) is incubated together with the standard/sample and forms an antibody-Amyloid-antibody-complex. This complex is indirectly linked with a Biotin-Streptavidin bridge to an enzyme in the next step. In a follow-up reaction the enzyme catalyzes the conversion of a substrate (Chromogen) into a colored product and the color intensity is measured by a spectrophotometer. The Amyloid  $\beta$ 42 concentration in unknown samples are calculated from the standard curve.

Characteristics of the A $\beta$ 42 ELISA (HS):

- $\circ$  Highly sensitive, selective quantitative analysis of human Amyloid  $\beta$ 42
- Test range from 16 to 500 pg/mL
- High reproducibility and accurate linearity of the standard curve
- Precoated strips (12x8) for flexible usage of samples according to individual customer requirements
- Low sample volumes (50 μL or less)

## III. REAGENTS SUPPLIED

Each kit is sufficient to run one 96-well plate and contains the following reagents:

## A. Human Amyloid β 1-42 ELISA Plate

Antibody-coated microtiter plate Quantity: 1 strip plate Preparation: Ready to Use Note: Unused strips should be resealed in the foil pouch with the desiccant provided and stored at 2-8℃.

## B. Adhesive Plate Sealer

Quantity: 2 sheets Preparation: Ready to Use

## C. Synthetic Amyloid β1-42 Standard

Concentration: 1000 ng/mL, lyophilized Quantity: 1 bottle containing 250  $\mu$ L / vial upon hydration.

Preparation: Reconstitute vial with 0.25 mL distilled or deionized water.

## D. Amyloid $\beta$ 1-42 Quality Controls 1 and 2

One vial each, lyophilized, containing amyloid  $\beta$ 1-42 at two different levels. Quantity: 0.25 ml/vial upon hydration. Preparation: Reconstitute each vial with 0.25 ml de-ionized water immediately before use.

## E. Standard & Sample Diluent

Quantity: 1 bottle containing 25mL Preparation: Ready to use, for dilution of standards or samples

## F. Antibody Conjugate (100X)

Quantity: 1 bottle containing 100  $\mu$ L at 100 fold concentrate Preparation: Dilute with Antibody Conjugate Diluent 1:100 before use

## G. Antibody Conjugate Diluent

Quantity: 1 bottle containing 8 mL Preparation: Ready to Use, for use in diluting the Antibody Conjugate.

**H. Enzyme Conjugate** (100X) (Steptavidin-Peroxidase-Conjugate) Quantity: 1 bottle containing 150 μL at 100 fold concentrate Preparation: Dilute with Enzyme Conjugate Diluent 1:100 before use

## I. Enzyme Conjugate Diluent

Pre-titered Streptavidin-Horseradish Peroxidase Conjugate in Buffer Quantity: 1 bottle containing 13 mL Preparation: Ready to Use, for dilution of Enzyme Conjugate

## J. Washing Solution (20X)

Quantity: 2 bottles containing 25 mL/bottle at 20 fold concentrate Preparation: Dilute 1:20 with deionized water before use

## K. Substrate Solution

Quantity: 1 bottle containing 13 mL Preparation: Ready to Use

## L. Stop Solution

0.3 M HCI Quantity: 1 bottle containing 12 mL Preparation: Ready to Use (Caution: Corrosive Solution)

## IV. STORAGE & STABILITY

Prior to use, all components in the kit can be stored up to 2 weeks at 2-8°C. For longer storage (> 2 weeks), freeze Wash Buffer and reconstituted Standards and Quality Controls at  $\leq -20^{\circ}$ C. Minimize repeated freeze and thaw of the Standards and Quality Controls. Refer to expiration dates on all reagents prior to use. Do not mix reagents from different kits unless they have the same lot numbers

## V. TECHNICAL GUIDELINES

- The hAmyloid β42-ELISA (HS) is for *in vitro* use only.
- Carefully read and follow the test-instructions in this user guide included in every testkit. Test performance and data calculation should always be done by qualified staff.
- Do not mix reagents from different test-kits.
- Some of the test components are concentrated solutions. After dilution the working solution should be used within 14 days (2 to 8 °C). Standard dilutions have to be diluted always just before the test starts.
- To calibrate the test-system (standard), the dilutions should be made according to the description in the test procedure. The resulting internal standard curve is a fixed component of each measurement. A transfer of the absorbance data from one test plate to another is not suitable.
- To avoid a cross contamination and carryover of reagents, the use of clean pipet tips for each sample pipetting is necessary.
- The pipetting of reagents and samples starts / stops kinetic reactions. To obtain a high precision for the test, be sure to treat each well of the microtiter plate in an identical manner.
- The washing solution has to be tapped out of the wells after the last washing step to assure the removal of buffer residues from the wells completely.

## VI. REAGENT PRECAUTIONS

## A. Hydrochloric Acid

Hydrochloric Acid is corrosive and can cause eye and skin burns. It is harmful if swallowed and can cause respiratory and digestive tract burns. Avoid contact with skin and eyes. Do not swallow or ingest

## VII. MATERIALS REQUIRED BUT NOT PROVIDED

In addition to the reagents provided in the test-kit, the following materials are essential for the performance of the ELISA-Test:

- Deionized water for dilution of Washing Solution
- Variable precision pipets (suitable for volumes from 10 to 1000  $\mu$ L) \*
- Vortex-mixer
- Timer
- Microtiter plate shaker
- Microtiter plate washer \*
- Microtiter Plate Reader capable of reading absorbancy at 450 nm
- Ice and ice container for sample preparation

\*Note: We recommend the use of multi channel pipets and automated plate-washers to achieve parallel working steps and simultaneous incubation times for best reproducibility.

## VIII. SAMPLE COLLECTION AND STORAGE

- We recommend diluting the working solutions only for the intended use. The test plate is subdivided in strips of 8 wells for flexible sample handling
- The stability of Amyloid β is critical, because the peptides tend to aggregate in samples. For this reason, the preanalytic sample preparation is a major influencing parameter within the analysis of Amyloid peptides. Samples should be collected according to clinical approved standard procedures and immediately stored at -20 °C. During the handling of thawed samples, it is important to keep these chilled (for example working on ice).
- For the preparation of the samples polypropylene vials are recommended to avoid interaction with sample materials during storage.
- Avoid repeated thawing and freezing of samples and standards.
- For research use only the test can be applied directly or in dilution to different sample materials (cell supernatants). More complex sample materials sometimes require a change of the test procedure.
- Plasma samples do not require dilution for the assay.
- CSF samples should be diluted 1:2 1:5 using Standard & Sample Diluent (#SD).
- Cell culture supernatants can be used neat or diluted 1:2 using Standard & Sample Diluent (#SD).

## IX. REAGENT PREPARATION

## A. <u>Preparation of Human Amyloid β 1-42 Standard (DAY 1):</u>

Use care in opening the lyophilized Standard vial. Using a pipette, reconstitute the Human Amyloid 1-42 Standard with 0.25 ml distilled or deionized water to give a concentration of 1000 ng/ml. Invert and mix gently, let sit for 5 minutes then vortex gently.

 Label eight tubes as Stock (S), 500 pg/ml, 250 pg/ml, 125 pg/ml, 62.5 pg/ml, 31.25 pg/ml, 16 pg/ml, and Blank. Using the chart below, add appropriate volumes of Standard and Sample Diluent to each of the eight tubes. Prepare dilutions according to the chart below and mix well.

Note: Change tip for every dilution. Wet tip with Standard before dispensing. Unused portions of standard should be stored at  $\leq$  -20°C. Avoid multiple freeze/thaw cycles.

Volume of Deionized	Volume of Standard	Standard Concentration
Water to Add	to Add	(ng/ml)
0.25 ml	0	1000 ng/ml

Standard	Concentration of Amyloid β 1-42	Volume of Standard and Sample Diluent to Add	Volume of Standard to Add
Stock (S)	25,000 pg/ml	780 µl	20 µl of reconstituted std
Std 1	500 pg/ml	1470 µl	30 µl of Stock (S)
Std 2	250 pg/ml	150 µl	150 µl Std 1
Std 3	125 pg/ml	150 µl	150 µl Std 2
Std 4	62.5 pg/ml	150 µl	150 µl Std 3
Std 5	31.25 pg/ml	150 µl	150 µl Std 4
Std 6	16 pg/ml	150 µl	150 µl Std 5
Blank	0 pg/ml	150 µl	0

## B. Quality Control 1 and 2 Preparation

Use care in opening the lyophilized Quality Control vials. Reconstitute each Amyloid  $\beta$  1-42 Quality Control 1 and Quality Control 2 with 0.25 ml distilled or de-ionized water and gently invert to ensure complete hydration. Unused portions of the reconstituted Quality Controls should be stored in small aliquots at -20°C. Avoid further freeze/thaw cycles.

## C. Preparation of Antibody Conjugate Solution (DAY 1):

Dilute Antibody Conjugate (100x) 1:100 with Antibody Conjugate Diluent. Example: 60  $\mu$ L Antibody Conjugate (100x) + 5940  $\mu$ L Antibody Conjugate Diluent = 6000  $\mu$ L

## D. Preparation of Washing Solution (DAY 2) :

Dilute Washing Solution (20x) 1:20 with deionized water Example: 50 mL Washing Solution (20x) + 950 mL deionized water = 1000 mL

## E. Preparation of Enzyme Conjugate Solution (DAY 2):

Dilute Enzyme Conjugate (100x) 1:100 with Enzyme Conjugate Diluent Example: 110  $\mu$ L Enzyme Conjugate (100x) + 10890  $\mu$ L Enzyme Conjugate = 11000  $\mu$ L

## X. ASSAY PROCEDURE

## <u>Day 1</u>

• The following kit components are required for day 1:

8 well test strips Standard & Sample Diluent Synthetic Aß1-42 Standard Aß 1-42 Quality Controls 1,2 Antibody Conjugate Diluent Antibody Conjugate (100x)

- We recommend diluting the test reagents just before each application. The samples shall be chilled (at <4 ℃, working on ice) during the complete test procedure to achieve high stability and optimal data results.
- All standards, quality controls or samples should be mixed gently just before pipetting. Accurate mixing and pipetting of the standard solutions are essential to the precision of the assay.
- 1. Add 50 µL Antibody Conjugate Solution into all wells.
- 2. Add 50 µL of Standard and Sample Diluent to the background (0 pg/ml) wells.
- 3. Add in duplicate 50  $\mu$ L standard and quality controls in order of ascending concentration to the appropriate wells. Add sequentially 50  $\mu$ L of samples in duplicate to the remaining wells. For best results all additions should be completed within 30 minutes.
- Cover the plate with a plate sealer and thoroughly mix the contents of the wells for a period of 5 minutes on an orbital plate shaker (500-600 rpm/min).
  Incubate without shaking overnight (16-20 hours) at 2 to 8 ℃.

Note: The test reagents needed on the following day can be taken out of the refrigerator to allow them to reach room temperature overnight

## <u>Day 2</u>

• The following kit components are required for day 2:

Washing Solution (20x) Enzyme Conjugate (100x) Enzyme Conjugate Diluent Substrate Solution Stop Solution

- Caution- All reagents must be at room temperature before use
- 1. Remove plate sealer and decant solutions from the plate. Tap as before to remove residual solutions in the wells.
- 2. Wash test plate 5 times with 300 μL Washing Solution per well, remove the remaining fluid by tapping the plate on an absorbing paper.

- 3. Add 100 μL Enzyme Conjugate Solution to each well. Cover the plate with a plate sealer and incubate for 30 minutes at room temperature (20 to 28 °C) on an orbital shaker (800 rpm/min).
- 4. Remove plate sealer and decant solutions from the plate. Tap as before to remove residual solutions in the wells.
- 5. Wash test plate 5 times with 300 µL Washing Solution per well, remove the remaining fluid by tapping the plate on an absorbing paper.
- 6. Add 100  $\mu$ L Substrate Solution to each well. Cover plate with sealer and shake on the plate shaker for 5 30 minutes (A longer development time may be needed if using a plate washer). Blue color should be formed in wells of standards with intensity proportional to increasing concentrations of Amyloid  $\beta$  1-42.

**NOTE:** Please be aware that the color may develop more quickly or more slowly than the recommended incubation time depending on the localized room temperature. Please visually monitor the color development to optimize the incubation time.

7. Remove sealer and add 100  $\mu$ L Stop Solution (Caution: Corrosive solution) and shake plate by hand to ensure complete mixing of solution in all wells. The blue color should turn to yellow after acidification. Read absorbance at 450 nm and 590 nm in a plate reader within 5 minutes and ensure that there are no air bubbles in any well. Record the difference in absorbance units. The absorbance of the highest Amyloid  $\beta$  1-42 standard should be approximately 2.0 – 3.2, or not to exceed the capability of the plate reader used.

# Assay Procedure

	Day 1							Day 2			
	Step 1	Step 2	Step 3	Step 4	Step 1-2	Step      Step 3      Step 4-5      Step 6		Step 7	Step 7		
Well #	Ab. Conj Solution	Standard & Sample Diluent	Standards/QCs & Samples		e	Enzyme Conj Solution	sidual Buffer	Substrate		Stop Solution	
A1, B1	50 μL	50 μL		ې نې	nove tl	100 μL	ove re Nash I paper	100 μL		100 μL	
C1, D1	50 μL		50 μl of 16 pg/mL Standard	2 to 8	ell, ren paper.		Remo 00 µL \ orbing		rature		
E1, F1	50 μL		50 μl of 31.25 pg/mL Standard	jht at 2	per we		nutes vith 30 n abso		tempe		Ë.
G1, H1	50 μL		50 μl of 62.5 pg/mL Standard	overniç	olution tn absc		r 30 mi sh 5X v te on a		t room		1d 590
A2, B2	50 μL		50 μl of 125 pg/mL Standard	ubate (	hing So te on a		ture fo s. Was the pla		min at		) nm ar
C2, D2	50 μL		50 μl of 250 pg/mL Standard	ien inc	L Wasl the pla		mpera t towels pping t		5 to 30		at 450
E2, F2	50 μL		50 μl of 500 pg/mL Standard	utes, th	300 µl pping 1		oom te sorbeni d by ta		ubate		bance
G2, H2	50 μL		50 μl of Quality Control 1	r 5 minu	ies with iid by ta		ate at r / on abs ing fluid		and inc		d Absor
A3, B3	50 μL		50 μl of Quality Control 2	itate fo	te 5 tin ning flu		d incub smartly remair		agitate		Rea
C3, D3	50 μL		50 µl of Sample	eal, ag	est plai remai		ate, and apping ve the		Seal, ¿		
E3, F3	50 μL		50 µl of Sample	Ō	Vash te		I., agita er by ta Remo				
G3, H3 ↓	50 μL		50 µl of Sample		>		Seal buffe	•			

# XI. MICROTITER PLATE ARRANGEMENT

	1	2	3	4	5	6	7	8	9	10	11	12
А	(Blank) 0 pg/mL	125 pg/mL	QC 2	Etc.								
В	(Blank) 0 pg/mL	125 pg/mL	QC 2									
С	16 pg/mL	250 pg/mL	Sample 1									
D	16 pg/mL	250 pg/mL	Sample 1									
E	31.25 pg/mL	500 pg/mL	Sample 2									
F	31.25 pg/mL	500 pg/mL	Sample 2									
G	62.5 pg/mL	QC 1	Sample 3									
Н	62.5 pg/mL	QC 1	Sample 3									

## XII. INTERPRETATION

Analysis of the measured absorbance data (mean, standard deviation) for the standards and for the samples is performed with the help of a microtiter plate reader software.

The **blank** (zero standards) is **<u>not</u>** integrated into the calculation of the standard curve. The blank is taken only as a control for a non-specific binding of the antibody; the mean absorbance of the blank shall be below 0.2.

Construct a standard curve by plotting the mean absorbance of standard 1-6 on the vertical axis versus the corresponding A $\beta$ 1-42 concentration on the horizontal axis. The data can be calculated by linear fit (linear regression) or by a point to point fit (cubic spline). The test results are not valid if the standard 1 (500 pg/mL) shows an absorbance below 0.6 in magnitude. Please control your test handling (XIV. TROUBLESHOOTING GUIDE).

The assay will be considered accepted when all Quality Control values fall within the calculated QC range. If any QCs fall outside of the control range, review results with a supervisor.

The resulting A $\beta$ 42-concentrations of the samples can be calculated with this standard curve. Only samples that are in the measured range of the standard curve can be calculated.

If the A $\beta$ -concentration value of a sample exceeds 500 pg/mL, the test sample must be measured again by using a higher sample dilution (with appropriate amount of Standard & Sample Diluent).

## XIII. ASSAY CHARACTERISTICS

## A. Sensitivity

The lowest level of Amyloid ß 1-42 standard used in this assay is 8 pg/mL (50  $\mu$ L sample size).

## **B.** Specificity

The Amyloid  $\beta$ 42 ELISA (HS) uses monoclonal anti-A $\beta$  antibodies with high selectivity for human A $\beta$ . The capture antibody recognizes the C-terminal end of Amyloid  $\beta$ 1-42, which causes a high selectivity for A $\beta$ 42. The cross-reactivity of the used antibodies to other Amyloid peptides was tested by ELISA and BIACORE and shows no significant cross-reactivity to A $\beta$ 1-38, A $\beta$ 1-39, A $\beta$ 1-40, A $\beta$ 1-43 and A $\beta$ 1-44.

## C. Precision

Analyte	Intra-Assay (%CV)	Inter-Assay (% CV)
Amyloid β 1-42	<10%	<10%

## XIV. QUALITY CONTROLS

The ranges for Quality Control 1 and 2 are provided on the card insert or can be located at the Millipore website www.millipore.com\bmia.

Problem	Cause	Recommend
No signal	wrong test reagents used	Ensure that only the reagents for the specific test-lot are used.
	test reagents damaged	Don't use the test-kit after expiration date.
Weak signal	test reagents used in a wrong dilution	Control used test dilutions carefully (usually a dilution factor of 100 is used).
	wrong filter (wavelength)	Check your wavelength in your microtiter plate photometer.
	incubation time too short temperature too low	Check the information of incubation times of the lot in the product data sheet. (The incubation time of the enzyme substrate is applied for temperatures from 20 to 28 °C); extend the substrate incubation time, if absorption is below 1.0
	reagents not at right temperature	Make sure that the reagents used for day 2 have reached room temperature (20 to 28 °C) before using within the test-kit.
	sodium azide, mercaptoethanol or DTT can interfere with peroxidase activity at high concentrations	Only use samples which contain no or low contents (< 0.1 %) of sodium azide, mercaptoethanol or DTT.
High signal	test reagents used in a wrong dilution	Check used test dilutions carefully (usually a dilution factor of 100 is used).
	incubation time too long temperature too high	Check the information of incubation times of the lot in the product data sheet. (The incubation time of the enzyme substrate is applied for temperatures from 20 to 28 °C); shorten the substrate incubation time, if absorption is above 3.0

## **XV. TROUBLESHOOTING GUIDE**

Problem	Cause	Recommend
High background (blank)	insufficient washing steps	Wash plate carefully and remove the liquid after each washing carefully.
	contamination of the washing solution	Confirm that the water is not contaminated. Use always double distilled water for the reconstitution and dilution of the washing solution.
	contamination of reagents or vials/tubes from previous experiments	Avoid pipetting directly out of the reagent vials, if test reagents should be used in further measurements. (Oxidative active contaminants can influence the enzyme substrate by non- specific color development).
	test reagents (antibody- and enzyme conjugate) used in wrong dilutions	Check used test dilutions for antibody- and enzyme conjugate carefully (usually a dilution factor of 100 is used).
<i>Low precision (= random error)</i>	non-homogeneous samples e.g. cloudy solution, particles in the sample	Check that the samples are taken, prepared and stored according to a recommended sample procedure (polypropylene tubes, storage of clear samples at -20 °C).
	insufficient mixing of samples and standards	Mix samples and standards before pipetting carefully.
	variation in pipetting	Check your pipettes and calibrate if necessary.
	carry over between samples and/or standards	Change pipet tips after each pipetting.
	insufficient mixing of reagents during incubation	Mix reagents on the test plate after pipetting by moving the test plate carefully; use an orbital microtiter plate shaker on the recommended test steps for optimal mixing of reagents.
	insufficient washing	Check that the automatic microtiter plate washer is working correctly; residues of liquids must be removed completely after each washing step.
	evaporation of liquids	Check the contact of the cover seal with the plate during the incubation steps.

Problem	Cause	Recommend
Calculated data are too high or too low (=systematic error, deviation of data from "typical data")	calculation of the dilution factor is not correct	Check the dilution factor used for the sample dilution within the data calculation.
	modification of the test procedure	Follow the instructions in the product data sheet carefully (incubation time, dilution etc.).
	incorrect sample treatment	Check that the samples are taken, prepared and stored according to a recommended sample procedure (polypropylene tubes, storage of clear samples at -20 °C).

# XVI. REPLACEMENT REAGENTS

Reagents	Cat. #
Human Amyloid B 1-42 ELISA Plate	4TS
Synthetic Aβ 1-42 Standard	2STM
Amyloid β 1-42 Quality Control 1,2	2QC
Standard & Sample Diluent	SD
Antibody Conjugate	HSAC
Antibody Conjugate Diluent	HSAD
Enzyme Conjugate	2ED
Enzyme Conjugate Diluent	2EC
Washing Solution	WS
Substrate Solution	ES
Stop Solution	ET-TMB

## XVII. ORDERING INFORMATION

#### A. To place an order:

#### For USA Customers:

Please provide the following information to our customer service department to expedite your telephone, fax or mail order:

- 1. Your name, telephone and/or fax number
- 2. Customer account number
- 3. Shipping and billing address
- 4. Purchase order number
- 5. Catalog number and description of product
- 6. Quantity and product size

TELEPHONE ORDERS: Toll Free US (800) MILLIPORE FAX ORDERS: (636) 441-8050 MAIL ORDERS: Millipore 6 Research Park Drive St. Charles, Missouri 63304 U.S.A.

## For International Customers:

To best serve our international customers, it is Millipore's policy to sell our products through a network of distributors. To place an order or to obtain additional information about Millipore products, please contact your local distributor.

#### B. Conditions of Sale

All products are for research or manufacturing use only. They are not intended for use in clinical diagnosis or for administration to human or animals. All products are intended for *in vitro* use only.

#### C. Material Safety Data Sheets (MSDS)

Material safety data sheets for Millipore products may be ordered by fax or phone. See Section A above for details on ordering.

## XVIII. REFERENCES

Ida N., Hartmann T., Pantel J., Schröder J., Zerfass R., Förstl H., Sandbrink R., Masters C.L., Beyreuther K., Analysis of Heterogeneous BA4 Peptides in Human Cerebrospinal Fluid and Blood by a Newly Developed Sensitive Western Blot Assay, *J. Biol. Chem.* 271 (37): 22908–22914 (1996).

Jensen M., Schröder J., Blomberg M., Engvall B., Pantel J., Ida N., Basun H., Wahlund L., Werle E., Jauss M., Beyreuther K., Lannfelt L., Hartmann T., Cerebrospinal Fluid AB42 is Increased Early in Sporadic Alzheimer's Disease and Declines with Disease Progression, *Ann. Neurol.* 45: 504-511 (1999).

Jensen M., Hartmann T., Engvall B., Wang R., Uljon S.N., Sennvik K., Näslund J., Muehlhauser F., Nordstedt C., Beyreuther K., Lannfelt L., Quantification of Alzheimer Amyloid  $\beta$  Peptides Ending at Residues 40 and 42 by Novel ELISA Systems, *Mol Medicine* 6: 291-302 (2000)

Shoji M, Cerebrospinal Fluid AB40 and AB42: Natural Course and Clinical Usefulness, *Frontiers in Bioscience* 7: 997-1006 (2002).

Lewczuk P, Esselmann H, Otto M, Maler JM, Henkel AW, Henkel MK, Eikenberg O, Antz C, Krause WR, Reulbach U, Kornhuber J, Wiltfang J., Neurochemical diagnosis of Alzheimer's dementia by CSF Abeta42, Abeta42/Abeta40 ratio and total tau, *Neurobiol Aging* 25(3):273-81 (2004)