



**MOUSE LEPTIN ELISA KIT**  
**96-Well Plate (Cat. # EZML-82K)**

I. Intended Use	2
II. Principles Of Procedure	2
III. Reagents Supplied	2
IV. Storage and Stability	4
V. Reagent Precautions	4
VI. Materials Required But Not Provided	4
VII. Sample Collection And Storage	5
VIII. Assay Procedure	6
IX. Microtiter Plate Arrangement	8
X. Calculations	9
XI. Interpretation	9
XII. Graph of Typical Reference Curve	10
XIII. Correlation Graph	11
XIV. Assay Characteristics	12
XV. Quality Controls	14
XVI. Troubleshooting Guide	14
XVII. Replacement Reagents	15
XVIII. Ordering Information	15

**MOUSE LEPTIN ELISA KIT**  
**96-Well Plate (Cat.# EZML-82K)**

**I. INTENDED USE**

This kit is used for the non-radioactive quantification of leptin in mouse sera. Plasma samples may also be used but application to samples of other biological fluids may need validation by the user. One kit is sufficient to measure 37 unknown samples in duplicate. ***This kit is for research purpose only.***

**II. PRINCIPLES OF PROCEDURE**

This assay is a Sandwich ELISA based, sequentially, on: 1) binding of leptin in the sample by a pre-titered antiserum and immobilization of the resulting complexes in the wells of a microtiter plate, 2) after washing purified biotinylated detection antibody is allowed to bind to the immobilized leptin, 3) binding of horseradish peroxidase to the immobilized biotinylated antibodies after free detection antibodies are washed off, 4) wash away of free enzyme conjugates, and 5) quantification of immobilized antibody-enzyme conjugates by monitoring horseradish peroxidase activities in the presence of the substrate 3,3',5,5'-tetramethylbenzidine. The enzyme activity is measured spectrophotometrically by the increased absorbency at 450 nm, corrected from the absorbency at 590nm, after acidification of formed products. Since the increase in absorbency is directly proportional to the amount of captured leptin in the unknown sample, the latter can be derived by interpolation from a reference curve generated in the same assay with reference standards of known concentrations of mouse leptin.

**III. REAGENTS SUPPLIED**

Each kit is sufficient to run one 96-well plate and contains the following reagents:

**A. Rat/Mouse Leptin Microtiter Plate**

Coated with pre-titered capture antibodies

Quantity: 1 plate

Preparation: Ready to use.

Note: Unused strips should be resealed in the foil pouch with the desiccant provided and stored at 2-8 °C.

**B. Adhesive Plate Sealer**

Quantity: 2 sheets

Preparation: Ready to use.

**C. Rat/Mouse Leptin Antiserum**

Pre-titered anti-rodent leptin serum

Quantity: 6 mL

Preparation: Ready to use.

### III. REAGENTS SUPPLIED (continued)

#### D. 10X HRP Wash Buffer Concentrate

10X concentrate of 50 mM Tris Buffered Saline containing Tween-20.  
Quantity: Two bottles containing 50 mL each  
Preparation: Dilute 1:10 with distilled or de-ionized water.

#### E. Mouse Leptin ELISA Standards

Mouse leptin in buffer: 0.2, 0.5, 1, 2, 5, 10, 20 and 30 ng/mL.  
Quantity: 0.5 mL/vial  
Preparation: Ready to use.

#### F. Quality Controls 1 and 2

Various peptides including leptin in QC buffer.  
Quantity: 0.5 mL/vial  
Preparation: Ready to use.

#### G. Rat/Mouse Leptin Matrix Solution

Matrix containing 0.08% Sodium Azide  
Quantity: 0.5 mL/vial  
Preparation: Ready to use.

#### H. Assay Buffer

0.05 M phosphosaline, pH 7.4, containing 0.025 M EDTA, 0.08% sodium azide, 0.05% Triton X-100 and 1% BSA.  
Quantity: 20 mL/vial  
Preparation: Ready to use.

#### I. Rat/Mouse Leptin Detection Antibody

Pre-titered biotinylated anti-mouse leptin antibody.  
Quantity: 12 mL/vial  
Preparation: Ready to use.

#### J. Enzyme Solution

Pre-titered streptavidin-horseradish peroxidase conjugate in buffer.  
Quantity: 12 mL/vial  
Preparation: Ready to use.

#### K. Substrate (Light Sensitive: avoid unnecessary exposure to light)

3, 3',5,5'-tetramethylbenzidine in buffer.  
Quantity: 12 mL/vial  
Preparation: Ready to use.

#### L. Stop Solution (Caution: Corrosive Solution)

0.3 M HCl  
Quantity: 12 mL/vial  
Preparation: Ready to use.

#### **IV. STORAGE AND STABILITY**

All components of the kit can be stored up to two weeks at 2-8°C. For longer storage (>2 weeks), freeze antiserum, standards, quality controls, and matrix solution at  $\leq -20^{\circ}\text{C}$  and avoid repeated freeze and thaw. Unused strips should be resealed in the foil pouch with the desiccant provided and stored at 2-8°C. Refer to expiration dates on all reagents prior to use. Do not mix reagents from different kits unless they have the same lot numbers.

#### **V. REAGENT PRECAUTIONS**

##### **A. Sodium Azide**

Sodium Azide has been added to certain reagents as a preservative at a concentration of 0.08%. Although it is at a minimum concentration, Sodium Azide may react with lead and copper plumbing to form explosive metal azides. On disposal, flush with large volume of water to prevent azide build up.

##### **B. Hydrochloric Acid**

Hydrochloric Acid is corrosive and can cause eye and skin burns. It is harmful if swallowed and can cause respiratory and digestive tract burns. Avoid contact with skin and eye. Do not swallow or ingest.

#### **VI. MATERIALS REQUIRED BUT NOT PROVIDED**

1. Pipettes and pipette tips: 10  $\mu\text{L}$  ~ 20  $\mu\text{L}$  and 20  $\mu\text{L}$  ~ 100  $\mu\text{L}$
2. Multi-channel Pipettes and pipette tips: 0 ~ 50  $\mu\text{L}$  and 50 ~ 300  $\mu\text{L}$
3. Buffer and Reagent Reservoirs
4. Vortex Mixer
5. De-ionized Water
6. Microtiter Plate Reader capable of reading absorbency at 450 nm and 590nm
7. Orbital Microtiter Plate Shaker
8. Absorbent Paper or Cloth

## VII. SAMPLE COLLECTION AND STORAGE

1. To prepare serum samples, whole blood is directly drawn into a centrifuge tube that contains no anti-coagulant. Let blood clot at room temperature for 30 min.
2. Promptly centrifuge the clotted blood at 2,000 to 3,000 x g for 15 minutes at  $4 \pm 2^{\circ}\text{C}$ .
3. Transfer and store serum samples in separate tubes. Date and identify each sample.
4. Use freshly prepared serum or aliquot and store samples at  $\leq -20^{\circ}\text{C}$  for later use. Avoid multiple ( $> 3$ ) freeze/thaw cycles.
5. To prepare plasma sample, whole blood should be collected into centrifuge tubes containing enough  $\text{K}_3\text{EDTA}$  to achieve a final concentration of 1.735 mg/mL and centrifuge immediately after collection. Observe the same precautions in the preparation of serum samples.
6. If heparin is to be used as anti-coagulant, the effect on the assay outcome at the dose of heparin used should be pre-determined.
7. Avoid using samples with gross hemolysis or lipemia.

## VIII. ASSAY PROCEDURE

**Pre-warm all reagents to room temperature immediately before setting up the assay.**

1. Dilute the 10X concentrated HRP Wash Buffer 10 fold by mixing the entire contents of both buffer bottles with 900 mL de-ionized or distilled water.
2. Remove the required number of strips from the Microtiter Assay Plate. Unused strips should be resealed in the foil pouch and stored at 2-8°C. Assemble strips in an empty plate holder and wash each well 3 times with 300  $\mu$ L of diluted Wash Buffer per wash. Decant Wash Buffer and remove the residual amount from all wells by inverting the plate and tapping it smartly onto absorbent towels several times. **Do not let wells dry before proceeding to the next step. If automated machine is used for assay, follow the manufacturer's instructions for all washing steps described in this protocol.**
3. Add 30  $\mu$ L Assay Buffer to Background wells, Standard wells, and QC1 and QC2 wells. Add 40  $\mu$ L Assay Buffer to sample wells.
4. If samples to be assayed are serum or plasma, add 10  $\mu$ L Matrix Solution to the Background wells, Standard wells, and QC1 and QC2 wells. If samples are free of significant serum matrix components, add 10  $\mu$ L Assay Buffer instead.
5. Add 10  $\mu$ L Assay Buffer to the Background wells and add in duplicates 10  $\mu$ L Mouse Leptin Standards in the order of ascending concentration to the appropriate wells.
6. Add 10  $\mu$ L QC1 and 10  $\mu$ L QC2 to the appropriate wells.
7. Add sequentially 10  $\mu$ L of the unknown samples in duplicate to the remaining wells.
8. Transfer Antiserum Solution to a reagent reservoir and add 50  $\mu$ L of this solution to each well with a multi-channel pipette. Cover the plate with plate sealer and incubate at room temperature for 2 hours on an orbital microtiter plate shaker set to rotate at moderate speed, about 400 to 500 rpm.
9. Remove plate sealer and decant solutions from the plate. Tap as before to remove residual solutions in well.
10. Wash wells 3 times with diluted Wash Buffer, 300  $\mu$ L per well per wash. Decant and tap after each wash to remove residual buffer.
11. Add 100  $\mu$ L Detection Antibody to each well. Cover plate with sealer and incubate with moderate shaking at room temperature for 1 hour on an orbital microtiter plate shaker set to rotate at moderate speed, approximately 400-500 rpm.
12. Remove plate sealer and decant solutions from the plate. Tap as before to remove residual solutions in well.

## VIII. ASSAY PROCEDURE (continued)

13. Wash wells 3 times with diluted Wash Buffer, 300  $\mu$ L per well per wash. Decant and tap after each wash to remove residual buffer.
14. Add 100  $\mu$ L Enzyme Solution to each well. Cover plate with sealer and incubate with moderate shaking at room temperature for 30 minutes on the micro-titer plate shaker.
15. Remove plate sealer and decant solutions from the plate. Tap as before to remove residual solutions in well.
16. Wash wells 6 times with diluted Wash Buffer, 300  $\mu$ L per well per wash. Decant and tap after each wash to remove residual buffer.
17. Add 100  $\mu$ L of Substrate solution to each well, cover plate with sealer and shake in the plate shaker for approximately 10 to 15 minutes. Blue color should be formed in wells of leptin standards with intensity proportional to increasing concentrations of leptin.

**NOTE:** Please be aware that the color may develop more quickly or more slowly than the recommended incubation time depending on the localized room temperature. Please visually monitor the color development to optimize the incubation time. One can monitor color development using 370 nm filter, if available on the spectrophotometer. When the absorbance is between 1.2 and 1.8 at 370 nm, the stop solution can be added to terminate the color development.

18. Remove sealer and add 100  $\mu$ L Stop Solution [**CAUTION: CORROSIVE SOLUTION**] and shake plate by hand to ensure complete mixing of solution in all wells. The blue color should turn into yellow after acidification. Read absorbance at 450 nm and 590 nm in a plate reader within 5 minutes and ensure that there is no air bubbles in any well. Record the difference of absorbance units.

## Assay Procedure for Intact Mouse Leptin ELISA kit (Cat. # EZML-82K)

	Step 1	Step 2	Step 3	Step 4	Step 5-7	Step 8	Step 9-10	Step 11	Step 12-13	Step 14	Step 15-16	Step 17	Step 18					
Well #	<b>Dilute both bottles of 10X Wash Buffer with 900mL Deionized Water.</b>	<b>Wash plate 3X with 300 µL Wash Buffer. Remove residual buffer by tapping smartly on absorbent towels.</b>	Assay Buffer	Matrix*	Standards/ Controls/ Samples	Antiserum	<b>Seal, Agitate, Incubate 2 hours at Room Temperature. Wash 3X with 300 µL Wash Buffer</b>	Detection Antibody	<b>Seal, Agitate, Incubate 1 hour at Room Temperature. Remove residual buffer by tapping smartly on absorbent towels. Wash 3X with 300 µL Wash Buffer</b>	Enzyme Solution	<b>Seal, Agitate, Incubate 30 minutes at Room Temperature. Wash 6X with 300 µL Wash Buffer</b>	Substrate	Stop Solution					
A1, B1			40 µL	10 µL	-----	50 µL		100 µL		100 µL		100 µL						
C1, D1			30 µL	10 µL	10 µL of 0.2 ng/mL Standard	50 µL		↓		↓		↓						
E1, F1			30 µL	10 µL	10 µL of 0.5 ng/mL Standard	50 µL												
G1, H1			30 µL	10 µL	10 µL of 1 ng/mL Standard	50 µL												
A2, B2			30 µL	10 µL	10 µL of 2 ng/mL Standard	50 µL												
C2, D2			30 µL	10 µL	10 µL of 5 ng/mL Standard	50 µL												
E2, F2			30 µL	10 µL	10 µL of 10 ng/mL Standard	50 µL												
G2, H2			30 µL	10 µL	10 µL of 20 ng/mL Standard	50 µL												
A3, B3			30 µL	10 µL	10 µL of 30 ng/mL Standard	50 µL												
C3, D3			30 µL	10 µL	10 µL of QC 1	50 µL												
E3, F3			30 µL	10 µL	10 µL of QC 2	50 µL												
G3, H3			40 µL	-----	10 µL of Sample	50 µL												
A4, B4...			40 µL	-----	10 µL of Sample	50 µL							↓	↓	↓			
<b>Seal, Agitate, Incubate 10-15 minutes at Room Temperature.</b>																		
<b>Shake by hand. Read Absorbance at 450 nm and 590 nm within 5 minutes.</b>																		

\* See Section VIII. Assay Procedure Step 4: If samples are free of significant matrix components, add 10 µL assay buffer instead.



## IX. MICROTITER PLATE ARRANGEMENT

### Mouse Leptin ELISA

	1	2	3	4	5	6	7	8	9	10	11	12
A	Blank	2.0 ng/mL	30 ng/mL	Sample 2								
B	Blank	2.0 ng/mL	30 ng/mL	Sample 2								
C	0.2 ng/mL	5.0 ng/mL	QC 1	Etc.								
D	0.2 ng/mL	5.0 ng/mL	QC 1									
E	0.5 ng/mL	10 ng/mL	QC 2									
F	0.5 ng/mL	10 ng/mL	QC 2									
G	1.0 ng/mL	20 ng/mL	Sample 1									
H	1.0 ng/mL	20 ng/mL	Sample 1									

## X. CALCULATIONS

Graph a reference curve by plotting the absorbance unit of 450nm, less unit at 590nm, on the Y-axis against the concentrations of mouse leptin standard on the X-axis. The dose-response curve of this assay fits best to a sigmoidal 4- or 5-parameter logistic equation. The results of unknown samples can be calculated with any computer program having a 4- or 5-parameter logistic function.


**Note:** When sample volumes assayed differ from 10  $\mu\text{L}$ , an appropriate mathematical adjustment must be made to accommodate for the dilution factor (e.g., if 5  $\mu\text{L}$  of sample is used, then calculated data must be multiplied by 2). When sample volume assayed is less than 10  $\mu\text{L}$ , compensate the volume deficit with either matrix solution or assay buffer, whichever is appropriate.

## XI. INTERPRETATION

1. The assay will be considered accepted when all Quality Control values fall within the calculated Quality Control Range; if any QC's fall outside the control range, review results with the supervisor.
2. If the difference between duplicate results of a sample is  $>15\%$  CV, repeat the sample.
3. The limit of sensitivity of this assay is 0.05 ng/mL ( $\sim 3.13$  pM) leptin (10  $\mu\text{L}$  sample size).
4. The appropriate range of this assay is 0.2 ng/mL to 30 ng/mL leptin (10  $\mu\text{L}$  sample size). Any result greater than 30 ng/mL in a 10  $\mu\text{L}$  sample assayed should be repeated on dilution using either matrix solution or assay buffer, whichever is appropriate, as diluent until it falls within range.

## XII. GRAPH OF TYPICAL REFERENCE CURVE

(n = 7 assays)




For Demonstration Only - Do not use for calculations

### III. CORRELATION GRAPH

#### Mouse leptin Assays:

#### Correlation of Results by RIA and ELISA Methods



Serum samples from 22 mice were assayed for Leptin using Millipore Mouse RIA Kit (Cat# ML-82K) and Mouse Leptin ELISA Kit (Cat# EZML-82K). Correlation of the two kits are derived by linear regression analysis of paired results from each sample.

## XIV. ASSAY CHARACTERISTICS

### A. Sensitivity

The lowest level of mouse leptin that can be detected by this assay is 0.05 ng/mL using a 10  $\mu$ L sample size.

### B. Specificity

The specificity (also known as selectivity) of the analytical test is its ability to selectively measure the analytes in the presence of other like components in the sample matrix.

Mouse Leptin	100%
Rat Leptin	70%
Human Leptin	9%
Porcine Leptin	< 0.05%
Ovine Leptin	< 0.05%
Chicken Leptin	< 0.05%
Rat Insulin	0%
Rat C-peptide	0%
Human Proinsulin	0%
Bovine Proinsulin	0%
Porcine Proinsulin	0%
Glucagon	0%
Human Ghrelin	0%

### C. Precision

Sample Number	Mean Leptin Level (ng/mL)	Assay Variation (CV)	
		Intra-assay	Inter-assay
1	1.66	1.06%	4.59%
2	5.78	1.64%	3.96%
3	17.60	1.76%	3.01%

The assay variations of Millipore Mouse Leptin ELISA kit were studied on three mouse serum samples with varying concentrations of spiked analyte. The intra-assay variations are calculated from eight duplicate determinations in an assay. The inter-assay variations are calculated from results of 6 separate assays with duplicate samples in each assay.

#### XIV. ASSAY CHARACTERISTICS (continued)

##### D. Recovery

Spike and Recovery of Mouse Leptin in Mouse Serum

Serum Sample #	Mouse Leptin		Recovery (%) of Spiked Insulin
	Added (ng/mL)	Observed (ng/mL)	
Mouse Serum # 1	0	1.33	--
	0.5	1.83	100
	2.0	3.14	91
	10.0	10.04	87
Mouse Serum # 2	0	1.76	--
	0.5	2.24	96
	2.0	3.49	87
	10.0	10.48	87

Mouse leptin at indicated levels was added to two pooled mouse serum samples and the resulting leptin content of each sample was assayed by ELISA. The % of recovery = [(observed leptin level after spike - observed leptin level before spike) / spiked level of leptin] x 100%. Mean recovery rate at spiked leptin level of 0.5, 2, and 10 ng/mL is 98%, 89%, and 87%, respectively.

##### E. Linearity

Effect of Serum Dilution

Serum Sample #	Dilution Factor	Leptin Level		
		Observed (ng/mL)	Expected (ng/mL)	% Of Expected
Mouse Serum # 1	--	16.43	16.43	100
	2x	16.42		100
	5x	16.60		101
	10x	17.50		107
	20x	17.20		105
Mouse Serum # 2	--	16.44	16.44	100
	2x	16.56		101
	5x	17.60		107
	10x	17.70		108
	20x	18.20		111

Two leptin-spiked pooled mouse serum samples are diluted each with matrix solution to various degrees as indicated and assayed for leptin levels along with neat samples of each serum. Measured leptin levels are corrected for dilution factors and reported as observed leptin level.

## **XV. QUALITY CONTROLS**

The ranges for Quality Control 1 and 2 are provided on the card insert or can be located at the Millipore website [www.millipore.com/bmia](http://www.millipore.com/bmia).

## **XVI. TROUBLESHOOTING GUIDE**

1. To obtain reliable and reproducible results the operator should carefully read this manual and fully understand all aspects of each assay step before attempting to run the assay.
2. Throughout the assay the operator should adhere strictly to the procedures with good laboratory practice.
3. Have all necessary reagents and equipment ready on hand before starting. Once the assay has been started all steps should be completed with precise timing and without interruption.
4. Avoid cross contamination of any reagents or samples to be used in the assay.
5. Make sure that all reagents and samples are added to the bottom of each well.
6. Careful and complete mixing of solutions in the well is critical. Poor assay precision will result from incomplete mixing or cross well contamination due to inappropriate mixing.
7. Remove any air bubble formed in the well after acidification of substrate solution because bubbles interfere with spectrophotometric readings.
8. Do not let the absorbance reading of the highest standard fall beyond the limit of your microtiterplate reader's capacity. Adjust the length of substrate incubation time accordingly.
9. High absorbance in background or blank wells could be due to 1) cross well contamination by standard solution or sample and 2) inadequate washing of wells with wash buffer.

## **XVII. REPLACEMENT REAGENTS**

<b>Reagents</b>	<b>Cat#</b>
Rat/Mouse Leptin Microtiter Plate	EP83
Rat/Mouse Leptin Antiserum	EAS83
10X HRP Wash Buffer Concentrate	EWB-HRP
Mouse Leptin ELISA Standards	E8082-K
Quality Controls 1 and 2	E6082-K
Rat/Mouse Leptin Matrix Solution	EPS0016
Assay Buffer	AB-PTRHK
Rat/Mouse Leptin Detection Antibody	E1083
Enzyme Solution	EHRP
Substrate	ESS-TMB2
Stop Solution	ET-TMB

## **XVIII. ORDERING INFORMATION**

### **A. To place an order:**

#### **For USA Customers:**

Please provide the following information to our customer service department to expedite your telephone, fax or mail order:

1. Your name, telephone and/or fax number
2. Customer account number
3. Shipping and billing address
4. Purchase order number
5. Catalog number and description of product
6. Quantity and product size

#### **TELEPHONE ORDERS:**

Toll Free US (866) 441-8400  
(636) 441-8400

**FAX ORDERS:** (636) 441-8050

**MAIL ORDERS:** Millipore

6 Research Park Drive  
St. Charles, Missouri 63304 U.S.A.

#### **For International Customers:**

To best serve our international customers, it is Millipore's policy to sell our products through a network of distributors. To place an order or to obtain additional information about Millipore products, please contact your local distributor.



## **XVIII. ORDERING INFORMATION (continued)**

### **B. Conditions of Sale**

All products are for research or manufacturing use only. They are not intended for use in clinical diagnosis or for administration to human or animals. All products are intended for *in vitro* use only.

### **C. Material Safety Data Sheets (MSDS)**

Material safety data sheets for Millipore products may be ordered by fax or phone. See Section A above for details on ordering.