


BioVendor

Research
and Diagnostic Products



Human TGFbeta 1 ELISA

Product Data Sheet

Cat. No.: RBMS249/3R

For Research Use Only

CONTENTS

1.	INTENDED USE	3
2.	SUMMARY	3
3.	PRINCIPLES OF THE TEST	6
4.	REAGENTS PROVIDED	7
5.	STORAGE INSTRUCTIONS – ELISA KIT	8
6.	SPECIMEN COLLECTION AND STORAGE INSTRUCTIONS	8
7.	MATERIALS REQUIRED BUT NOT PROVIDED	8
8.	PRECAUTIONS FOR USE	9
9.	PREPARATION OF REAGENTS	10
10.	TEST PROTOCOL	13
11.	CALCULATION OF RESULTS	16
12.	LIMITATIONS	19
13.	PERFORMANCE CHARACTERISTICS	19
14.	REFERENCES	23
15.	REAGENT PREPARATION SUMMARY	25
16.	TEST PROTOCOL SUMMARY	26

**»» This kit is manufactured by:
BioVendor – Laboratorní medicína a.s.**

»» Use only the current version of Product Data Sheet enclosed with the kit!

1. INTENDED USE

The human TGF- β 1 ELISA is an enzyme-linked immunosorbent assay for the quantitative detection of human TGF- β 1. **The human TGF- β 1 ELISA is for research use only. Not for diagnostic or therapeutic procedures.**

2. SUMMARY

Transforming growth factor- β (TGF- β) is a pleiotropic cytokine that exhibits a broad spectrum of biological and regulatory effects on the cellular and organism level. It plays a critical role in cellular growth, development, differentiation, proliferation, extracellular matrix (ECM) synthesis and degradation, control of mesenchymal-epithelial interactions during embryogenesis, immune modulation, apoptosis, cell cycle progression, angiogenesis, adhesion and migration and leukocyte chemotaxis. It has both tumor suppressive and tumor promoting activities and is highly regulated at all levels (e.g.: mRNA turnover, latent protein activation or post-translational modifications).

TGF- β is the first recognized protein of at least 40 of the TGF- β superfamily of structurally related cytokines.

Three isoforms (TGF- β 1-3) have been described in mammals. (Each isoform is encoded by a unique gene on different chromosomes. All bind to the same receptors.) They are synthesized by most cell types and tissues. Cells of the immune system mainly express TGF- β 1.

The initially sequestered, inactive LTGF- β (latent TGF- β) requires activation (cleavage and dissociation of its LAP (latency associated peptide) region) before it can exert biological activity. LTGF- β can also be bound to LTB (latent TGF- β binding protein) to form a large latent complex (LLC). TGF- β forms homodimers, and its subunits of 12.5 kDa each are bound via disulphide bridges.

TGF- β signal transduction is mediated via the TGF- β receptors Type II and I, phosphorylation and conformational changes, followed by different pathways:

SMAD (- pathway: TGF- β recruitment finally leads to phosphorylation of receptor-regulated SMADs (R-SMADs = SMAD 2, 3) and binding of common SMAD (coSMAD = SMAD 4).

The R-SMAD/ coSMAD complex enters the nucleus and interacts with a number of transcription factors, coactivators and corepressors.

TGF- β induces MAPK- and MAP/ERK kinase dependent signal transduction (JNK/MAPK-, JNK/SPAK-, p38-, ERK1/2 - pathway) and the NF- κ B – pathway. TGF- β mediates cell cycle growth arrest via the phosphoinositide 3-kinase/Akt pathway.

TGF- β signaling is highly regulated e.g. via interaction with inhibitory SMADs (I-SMADs = SMAD 6, 7) or binding of the E3-ubiquitin ligases Smurf1 and Smurf2 or/and coreceptors.

TGF- β 1 is the key mediator in the pathophysiology of tissue repair and human fibrogenesis: balance between production and degradation of type I collagen, and fibrosis and scarring in organ and tissue.

TGF- β 1 exhibits important immunoregulatory features of partially adverse character: TGF- β 1 inhibits B and T cell proliferation, differentiation and antibody production as well as maturation and activation of macrophages. It inhibits activity of NK cells and lymphokine activated killer cells and blocks production of cytokines. TGF- β 1 promotes Treg cell differentiation resulting in IL-10/TGF- β 1 production and Th1 cell and Th2 cell suppression.

TGF- β 1 was recently shown to promote Th17 development in the presence of IL-6 or IL-21 in mice and probably plays a role in human Th17 development together with IL-1 β , IL-21 and IL-23. In this context TGF- β 1 is involved in induction and mediation of proinflammatory and allergic responses.

Cancer: TGF- β 1 is overexpressed in a high percentage of human tumors (e.g.: breast, prostate, renal cell, pancreatic, ovarian, cervical and gastric cancer and melanoma, non-hodgkin's lymphoma, multiple myeloma) and has been correlated with poor prognosis.

TGF- β 1 acts as a tumor suppressor (particularly in the early stage of carcinogenesis) and as a tumor promoter (namely, tumor progression, invasion and metastasis). Malignant cells secrete TGF- β 1, suppressing antitumor immune responses and creating immune tolerance. Mutations in the TGF- β 1 signaling pathway (e.g.: loss of cell surface receptors, decreased SMAD expression) render tumors refractory to growth inhibitory and apoptotic effects of TGF- β 1.

Autoimmune diseases: TGF- β 1 is functionally connected to major immune system abnormalities such as Systemic Lupus Erythematosus (SLE). In Multiple Sclerosis (MS) up-regulation of TGF- β 1 seems to correlate with a benign course and minor disability of MS. In autoimmune hepatitis (AIH) up-regulated serum TGF- β 1 has been observed. Pathological remodeling of connective tissue in systemic sclerosis (SSc) is attributed to the activation of the TGF- β 1/ SMAD pathway. In mice, TGF- β 1 gene transfer to the colon leads to intestinal fibrosis and serves as a mouse model for Crohn's disease (CD).

Liver: Increased TGF- β 1 expression was shown in hepatic fibrosis of chronic liver diseases (chronic hepatitis, alcoholic cirrhosis). Hepatitis C virus upregulates TGF- β 1 transcription in the liver and elevates circulating TGF- β 1 levels. Disrupting TGF- β 1 synthesis and/or signaling pathways prevents scar formation in experimental liver fibrosis. Removal of excess collagen after cessation of liver disease is regulated by TGF- β 1.

Kidney diseases: Glomerulonephritis and diabetic nephropathy due to excessive accumulation of ECM within the mesangium of the glomeruli is attributed to high TGF- β 1 levels. Urinary TGF- β 1 levels of patients with these diseases are elevated.

Diabetes: Expression levels and kinase activities of components of the TGF- β signaling pathway are altered in diseases of the pancreas. Low TGF- β levels in patients with Type I diabetes may contribute to a lack of immunosuppression and to disease propagation and maintenance.

Cardiovascular Diseases: TGF- β 1 is anti-atherogenic and atheroprotective, but loses its protective role and exhibits pathogenic effects in chronic disease. Increased TGF- β 1 levels were found in atherosclerotic clinical specimens. Dilated, ischaemic and hypertrophic cardiomyopathies are associated with high TGF- β 1 levels.

TGF- β 1 induces endothelial cells to undergo EndMT (endothelial-mesenchymal transition), which contributes to the progression of cardiac fibrosis associated with chronic heart disease. Decreased serum levels of TGF- β 1 in sepsis and acute stroke patients may reflect the changing immunological-inflammatory status of these patients. Remodeling after myocardial infarction is attributed to TGF- β 1. An upregulation of TGF- β 1 in the central nervous system after ischemia-induced brain damage has been described. A neuroprotective role of TGF- β 1 against ischemia-induced neuronal cell death was found.


Asthma, Chronic Obstructive Pulmonary Disorder (COPD), Cystic fibrosis (CF): TGF- β 1 plays an important role in chronic airway diseases, particularly in airway remodeling. Increased TGF- β 1 levels have been described in patients with severe asthma and airway eosinophilic inflammation, and have been correlated to the degree of sub-epithelial fibrosis. In mouse models decreased TGF- β 1 lead to reduced peribronchial fibrosis, airway smooth muscle proliferation and mucus production. TGF- β 1 was shown to induce apoptosis in airway epithelial cells. Integrin mediated local activation of TGF- β is critical for the development of pulmonary edema in acute lung injury.

Others: In Alzheimer's Disease (AD) increased TGF- β 1 immunoreactivity and TGF- β 1 mRNA levels correlate with beta-amyloid deposition in damaged cerebral blood vessels. A number of proinflammatory chemokines including TGF- β 1 are consistently elevated in brains of autistic patients. Decreased TGF- β 1 serum levels were described for patients with acute malaria. In periodontitis increased TGF- β 1 levels were measured in gingival crevicular fluid.


Patients with duchenne muscular dystrophy showed elevated TGF- β 1 expression levels and fibrosis. Increased TGF- β signaling events in scleroderma fibroblasts were shown. TGF- β 1 is a potent stimulator of chondrocyte matrix production versus tissue fibrosis and thus might play an important role in bone metabolism (Osteoarthritis).

3. PRINCIPLES OF THE TEST


An anti-human TGF- β 1 coating antibody is adsorbed onto microwells.




Human TGF- β 1 present in the sample or standard binds to antibodies adsorbed to the microwells.



A biotin-conjugated anti-human TGF- β 1 antibody is added and binds to human TGF- β 1 captured by the first antibody.




Following incubation unbound biotin-conjugated anti-human TGF- β 1 antibody is removed during a wash step. Streptavidin-HRP is added and binds to the biotin-conjugated anti-human TGF- β 1 antibody.



Following incubation unbound Streptavidin-HRP is removed during a wash step, and substrate solution reactive with HRP is added to the wells.


Figure 5

Fourth Incubation



A coloured product is formed in proportion to the amount of human TGF- β 1 present in the sample or standard. The reaction is terminated by addition of acid and absorbance is measured at 450 nm. A standard curve is prepared from 7 human TGF- β 1 standard dilutions and human TGF- β 1 sample concentration determined.

Figure 6



4. REAGENTS PROVIDED

- 1 aluminium pouch with a **Antibody Coated Microtiter Strips** with monoclonal antibody to human TGF- β 1
- 1 vial (120 μ l) **Biotin-Conjugate** anti-human TGF- β 1 polyclonal antibody
- 1 vial (150 μ l) **Streptavidin-HRP**
- 2 vials human TGF- β 1 **Standard** lyophilized, 4 ng/ml upon reconstitution
- 1 vial (5 ml) **Assay Buffer Concentrate** 20x (PBS with 1% Tween 20 and 10% BSA)
- 1 bottle (50 ml) **Wash Buffer Concentrate** 20x (PBS with 1% Tween 20)
- 1 vial (3 ml) 1 N HCl
- 1 vial (3 ml) 1 N NaOH
- 1 vial (15 ml) **Substrate Solution** (tetramethyl-benzidine)
- 1 vial (15 ml) **Stop Solution** (1M Phosphoric acid)
- 1 vial (0.4 ml) **Blue-Dye**
- 1 vial (0.4 ml) **Green-Dye**
- 1 vial (0.4 ml) **Red-Dye**
- 6 **Adhesive Films**

5. STORAGE INSTRUCTIONS – ELISA KIT

Store kit reagents between 2° and 8°C. Immediately after use remaining reagents should be returned to cold storage (2° to 8°C).

Expiry of the kit and reagents is stated on labels. Expiry of the kit components can only be guaranteed if the components are stored properly, and if, in case of repeated use of one component, this reagent is not contaminated by the first handling.

6. SPECIMEN COLLECTION AND STORAGE INSTRUCTIONS

Cell culture supernatant *, serum and plasma (EDTA, citrate, heparin) were tested with this assay. Other biological samples might be suitable for use in the assay. Remove serum or plasma from the clot or cells as soon as possible after clotting and separation.

Samples containing a visible precipitate must be clarified prior to use in the assay. Do not use grossly hemolyzed or lipemic specimens.

Samples should be aliquoted and must be stored frozen at -20°C to avoid loss of bioactive human TGF-β1. If samples are to be run within 24 hours, they may be stored at 2° to 8°C (for sample stability refer to 0).

Avoid repeated freeze-thaw cycles. Prior to assay, the frozen sample should be brought to room temperature slowly and mixed gently.

* Pay attention to a possibly elevated blank signal in cell culture supernatant samples containing serum components (e.g. FCS), due to latent TGF-β levels in animal serum.

7. MATERIALS REQUIRED BUT NOT PROVIDED

- 5 ml and 10 ml graduated pipettes
- 5 µl to 1000 µl adjustable single channel micropipettes with disposable tips
- 50 µl to 300 µl adjustable multichannel micropipette with disposable tips
- Multichannel micropipette reservoir
- Beakers, flasks, cylinders necessary for preparation of reagents
- Device for delivery of wash solution (multichannel wash bottle or automatic wash system)
- Microplate shaker
- Microwell strip reader capable of reading at 450 nm (620 nm as optional reference wave length)
- Glass-distilled or deionized water
- Statistical calculator with program to perform regression analysis

8. PRECAUTIONS FOR USE

- All chemicals should be considered as potentially hazardous. We therefore recommend that this product is handled only by those persons who have been trained in laboratory techniques and that it is used in accordance with the principles of good laboratory practice. Wear suitable protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken to avoid contact with skin or eyes. In the case of contact with skin or eyes wash immediately with water. See material safety data sheet(s) and/or safety statement(s) for specific advice.
- Reagents are intended for research use only and are not for use in diagnostic or therapeutic procedures.
- Do not mix or substitute reagents with those from other lots or other sources.
- Do not use kit reagents beyond expiration date on label.
- Do not expose kit reagents to strong light during storage or incubation.
- Do not pipette by mouth.
- Do not eat or smoke in areas where kit reagents or samples are handled.
- Avoid contact of skin or mucous membranes with kit reagents or specimens.
- Rubber or disposable latex gloves should be worn while handling kit reagents or specimens.
- Avoid contact of substrate solution with oxidizing agents and metal.
- Avoid splashing or generation of aerosols.
- In order to avoid microbial contamination or cross-contamination of reagents or specimens which may invalidate the test use disposable pipette tips and/or pipettes.
- Use clean, dedicated reagent trays for dispensing the conjugate and substrate reagent.
- Exposure to acid inactivates the conjugate.
- Glass-distilled water or deionized water must be used for reagent preparation.
- Substrate solution must be at room temperature prior to use.
- Decontaminate and dispose specimens and all potentially contaminated materials as they could contain infectious agents. The preferred method of decontamination is autoclaving for a minimum of 1 hour at 121.5°C.
- Liquid wastes not containing acid and neutralized waste may be mixed with sodium hypochlorite in volumes such that the final mixture contains 1.0% sodium hypochlorite. Allow 30 minutes for effective decontamination. Liquid waste containing acid must be neutralized prior to the addition of sodium hypochlorite.

9. PREPARATION OF REAGENTS

Buffer Concentrates should be brought to room temperature and should be diluted before starting the test procedure.

If crystals have formed in the **Buffer Concentrates**, warm them gently until they have completely dissolved.

9.1 Wash Buffer (1x)

Pour entire contents (50 ml) of the **Wash Buffer Concentrate** (20x) into a clean 1000 ml graduated cylinder. Bring to final volume of 1000 ml with glass-distilled or deionized water. Mix gently to avoid foaming. The pH of the final solution should adjust to 7.4.

Transfer to a clean wash bottle and store at 2° to 25°C. Please note that Wash Buffer (1x) is stable for 30 days.

Wash Buffer (1x) may also be prepared as needed according to the following table:

Number of Strips	Wash Buffer Concentrate (20x) (ml)	Distilled Water (ml)
1 - 6	25	475
1 - 12	50	950

9.2 Assay Buffer (1x)

Pour the entire contents (5 ml) of the **Assay Buffer Concentrate** (20x) into a clean 100 ml graduated cylinder. Bring to final volume of 100 ml with distilled water. Mix gently to avoid foaming.

Store at 2° to 8°C. Please note that the Assay Buffer (1x) is stable for 30 days.

Assay Buffer (1x) may also be prepared as needed according to the following table:

Number of Strips	Assay Buffer Concentrate (20x) (ml)	Distilled Water (ml)
1 - 6	2.5	47.5
1 - 12	5.0	95.0

9.3 Biotin-Conjugate

Please note that the Biotin-Conjugate should be used within 30 minutes after dilution.

Make a 1:100 dilution of the concentrated **Biotin-Conjugate** solution with Assay Buffer (1x) in a clean plastic tube as needed according to the following table:

Number of Strips	Biotin-Conjugate (ml)	Assay Buffer (1x) (ml)
1 - 6	0.06	5.94
1 - 12	0.12	11.88

9.4 Streptavidin-HRP

Please note that the Streptavidin-HRP should be used within 30 minutes after dilution.

Make a 1:100 dilution of the concentrated **Streptavidin-HRP** solution with Assay Buffer (1x) in a clean plastic tube as needed according to the following table:

Number of Strips	Streptavidin-HRP (ml)	Assay Buffer (1x) (ml)
1 - 6	0.06	5.94
1 - 12	0.12	11.88

9.5 Human TGF- β 1 Standard

Reconstitute **human TGF- β 1 standard** by addition of distilled water.

Reconstitution volume is stated in the Quality Control Sheet. Swirl or mix gently to insure complete and homogeneous solubilization (concentration of reconstituted standard = 4 ng/ml). Allow the standard to reconstitute for 10-30 minutes. Mix well prior to making dilutions.

After usage remaining standard cannot be stored and has to be discarded.

Standard dilutions can be prepared directly on the microwell plate (see 10.d) or alternatively in tubes (see 0).

9.5.1 External Standard Dilution

Label 7 tubes, one for each standard point.

S1, S2, S3, S4, S5, S6, S7


Then prepare 1:2 serial dilutions for the standard curve as follows:

Pipette 225 μ l of Assay Buffer (1x) into each tube.

Pipette 225 μ l of reconstituted standard (concentration of standard = 4 ng/ml) into the first tube, labelled S1, and mix (concentration of standard 1 = 2 ng/ml). Pipette 225 μ l of this dilution into the second tube, labelled S2, and mix thoroughly before the next transfer. Repeat serial dilutions 5 more times thus creating the points of the standard curve (see Figure 7).

Assay Buffer (1x) serves as blank.

Figure 7



9.6 Addition of Colour-giving Reagents: Blue-Dye, Green-Dye, Red-Dye

This procedure is optional, does not in any way interfere with the test results, and is designed to help the customer with the performance of the test, but can also be omitted, just following the instruction booklet.

Alternatively, the dye solutions from the stocks provided (**Blue-Dye, Green-Dye, Red-Dye**) can be added to the reagents according to the following guidelines:

1. Diluent: Before standard dilution add the **Blue-Dye** at a dilution of 1:250 (see table below) to the appropriate diluent (1x) according to the test protocol. After addition of **Blue-Dye**, proceed according to the instruction booklet.

5 ml Assay Buffer (1x)	20 µl Blue-Dye
12 ml Assay Buffer (1x)	48 µl Blue-Dye
50 ml Assay Buffer (1x)	200 µl Blue-Dye

2. Biotin-Conjugate: Before dilution of the concentrated Biotin-Conjugate, add the **Green-Dye** at a dilution of 1:100 (see table below) to the Assay Buffer (1x) used for the final conjugate dilution. Proceed after addition of **Green-Dye** according to the instruction booklet: Preparation of Biotin-Conjugate.

3 ml Assay Buffer (1x)	30 µl Green-Dye
6 ml Assay Buffer (1x)	60 µl Green-Dye
12 ml Assay Buffer (1x)	120 µl Green-Dye


3. Streptavidin-HRP: Before dilution of the concentrated Streptavidin-HRP, add the **Red-Dye** at a dilution of 1:250 (see table below) to the Assay Buffer (1x) used for the final Streptavidin-HRP dilution. Proceed after addition of **Red-Dye** according to the instruction booklet: Preparation of Streptavidin-HRP.

6 ml Assay Buffer (1x)	24 µl Red-Dye
12 ml Assay Buffer (1x)	48 µl Red-Dye

10. TEST PROTOCOL

- a. Prepare your samples before starting the test procedure. Dilute serum, plasma and cell culture supernatant samples 1:10 with Assay Buffer (1x) according to the following scheme: 20 µl sample + 180 µl Assay Buffer (1x) Add 20 µl 1N HCl to 200 µl prediluted sample, mix and incubate for 1 hour at room temperature. Neutralize by addition of 20 µl 1N NaOH.
- b. Determine the number of microwell strips required to test the desired number of samples plus appropriate number of wells needed for running blanks and standards. Each sample, standard, blank and optional control sample should be assayed in duplicate. Remove extra microwell strips from holder and store in foil bag with the desiccant provided at 2°-8°C sealed tightly.
- c. Wash the microwell strips twice with approximately 400 µl **Wash Buffer** per well with thorough aspiration of microwell contents between washes. Allow the Wash Buffer to sit in the wells for about **10 – 15 seconds** before aspiration. Take care not to scratch the surface of the microwells. After the last wash step, empty wells and tap microwell strips on absorbent pad or paper towel to remove excess Wash Buffer. Use the microwell strips immediately after washing. Alternatively microwell strips can be placed upside down on a wet absorbent paper for not longer than 15 minutes. **Do not allow wells to dry.**
- d. **Standard dilution on the microwell plate** (Alternatively the standard dilution can be prepared in tubes – see 0): Add 100 µl of Assay Buffer (1x) in duplicate to all **standard wells**. Pipette 100 µl of prepared **standard** (see Preparation of Standard 0, concentration = 4000 pg/ml) in duplicate into well A1 and A2 (see Table 1). Mix the contents of wells A1 and A2 by repeated aspiration and ejection (concentration of standard 1, S1 = 2000 pg/ml), and transfer 100 µl to wells B1 and B2, respectively (see Figure 8). Take care not to scratch the inner surface of the microwells. Continue this procedure 5 times, creating two rows of human TGF- β 1 standard dilutions ranging from 2000 to 31 pg/ml. Discard 100 µl of the contents from the last microwells (G1, G2) used.

Figure 8



In case of an **external standard dilution** (see 0), pipette 100 µl of these standard dilutions (S1 - S7) in the standard wells according to Table 1.

Table 1
Table depicting an example of the arrangement of blanks, standards and samples in the microwell strips:

	1	2	3	4
A	Standard 1 (2000 pg/ml)	Standard 1 (2000 pg/ml)	Sample 1	Sample 1
B	Standard 2 (1000 pg/ml)	Standard 2 (1000 pg/ml)	Sample 2	Sample 2
C	Standard 3 (500 pg/ml)	Standard 3 (500 pg/ml)	Sample 3	Sample 3
D	Standard 4 (250 pg/ml)	Standard 4 (250 pg/ml)	Sample 4	Sample 4
E	Standard 5 (125 pg/ml)	Standard 5 (125 pg/ml)	Sample 5	Sample 5
F	Standard 6 (63 pg/ml)	Standard 6 (63 pg/ml)	Sample 6	Sample 6
G	Standard 7 (31 pg/ml)	Standard 7 (31 pg/ml)	Sample 7	Sample 7
H	Blank	Blank	Sample 8	Sample 8

- e. Add 100 µl of **Assay Buffer (1x)** in duplicate to the **blank wells**.
- f. Add 60 µl of **Assay Buffer (1x)** to the **sample wells**.
- g. Add 40 µl of each pretreated **sample** in duplicate to the **sample wells**.
- h. Cover with an adhesive film and incubate at room temperature (18 to 25°C) for 2 hours, on a microplate shaker set at 100 rpm. (**Shaking is absolutely necessary for an optimal test performance.**)
- i. Prepare **Biotin-Conjugate** (see Preparation of Biotin-Conjugate 0).
- j. Remove adhesive film and empty wells. **Wash** microwell strips 5 times according to point c. of the test protocol. Proceed immediately to the next step.
- k. Add 100 µl of **Biotin-Conjugate** to all wells.
- l. Cover with an adhesive film and incubate at room temperature (18 to 25°C) for 1 hour, on a microplate shaker set at 100 rpm. (**Shaking is absolutely necessary for an optimal test performance.**)
- m. Prepare **Streptavidin-HRP** (refer to Preparation of Streptavidin-HRP 0).
- n. Remove adhesive film and empty wells. **Wash** microwell strips 5 times according to point c. of the test protocol. Proceed immediately to the next step.
- o. Add 100 µl of diluted **Streptavidin-HRP** to all wells, including the blank wells.
- p. Cover with an adhesive film and incubate at room temperature (18° to 25°C) for 1 hour, on a microplate shaker set at 100 rpm. (**Shaking is absolutely necessary for an optimal test performance.**)
- q. Remove adhesive film and empty wells. **Wash** microwell strips 5 times according to point c. of the test protocol. Proceed immediately to the next step.
- r. Pipette 100 µl of **TMB Substrate Solution** to all wells.
- s. Incubate the microwell strips at room temperature (18° to 25°C) for **about 30 min**. Avoid direct exposure to intense light.
 - 1) **The colour development on the plate should be monitored and the substrate reaction stopped (see next point of this protocol) before positive wells are no longer properly recordable. Determination of the ideal time period for colour development has to be done individually for each assay.**
 - 2) It is recommended to add the stop solution when the highest standard has developed a dark blue colour. Alternatively the colour development can be monitored by the ELISA reader at 620 nm. The substrate reaction should be stopped as soon as Standard 1 has reached an OD of 0.90 – 0.95.
- t. Stop the enzyme reaction by quickly pipetting 100 µl of **Stop Solution** into each well. It is important that the Stop Solution is spread quickly and uniformly throughout the microwells to completely inactivate the enzyme. Results must be read immediately after the Stop Solution is added or within one hour if the microwell strips are stored at 2 - 8°C in the dark.
- u. Read absorbance of each microwell on a spectro-photometer using 450 nm as the primary wave length (optionally 620 nm as the reference wave length; 610 nm to 650 nm is acceptable). Blank the plate reader according to the manufacturer's instructions by using the blank wells. Determine the absorbance of both the samples and the standards.

11. CALCULATION OF RESULTS

- Calculate the average absorbance values for each set of duplicate standards and samples. Duplicates should be within 20 per cent of the mean value.
- Create a standard curve by plotting the mean absorbance for each standard concentration on the ordinate against the human TGF- β 1 concentration on the abscissa. Draw a best fit curve through the points of the graph (a 5-parameter curve fit is recommended).
- To determine the concentration of circulating human TGF- β 1 for each sample, first find the mean absorbance value on the ordinate and extend a horizontal line to the standard curve. At the point of intersection, extend a vertical line to the abscissa and read the corresponding human TGF- β 1 concentration.
- **If instructions in this protocol have been followed samples have been diluted 1:30 (20 μ l sample + 180 μ l Assay Buffer (1x) + 20 μ l 1N HCl + 20 μ l 1N NaOH and 40 μ l pretreated sample + 60 μ l Assay Buffer (1x)), the concentration read from the standard curve must be multiplied by the dilution factor (x 30).**
- **Calculation of samples with a concentration exceeding standard 1 may result in incorrect, low human TGF- β 1 levels. Such samples require further external predilution according to expected human TGF- β 1 values with Assay Buffer (1x) in order to precisely quantitate the actual human TGF- β 1 level.**
- It is suggested that each testing facility establishes a control sample of known human TGF- β 1 concentration and runs this additional control with each assay. If the values obtained are not within the expected range of the control, the assay results may be invalid.
- A representative standard curve is shown in Figure 9. This curve cannot be used to derive test results. Each laboratory must prepare a standard curve for each group of microwell strips assayed.

Figure 9
Representative standard curve for human TGF- β 1 ELISA. Human TGF- β 1 was diluted in serial 2-fold steps in Assay Buffer (1x). Do not use this standard curve to derive test results. A standard curve must be run for each group of microwell strips assayed.




Table 2

Typical data using the human TGF- β 1 ELISA

Measuring wavelength: 450 nm

Reference wavelength: 620 nm

Standard	Human TGF- β 1 Concentration (pg/ml)	O.D. at 450 nm	Mean O.D. at 450 nm	C.V. (%)
1	2000	2.005 1.995	2.000	0.2
2	1000	1.151 1.060	1.105	4.1
3	500	0.668 0.663	0.665	0.4
4	250	0.388 0.392	0.390	0.6
5	125	0.243 0.244	0.244	0.2
6	63	0.171 0.166	0.168	1.6
7	31	0.131 0.131	0.131	0.2
Blank	0	0.097 0.098	0.098	0.3

The OD values of the standard curve may vary according to the conditions of assay performance (e.g. operator, pipetting technique, washing technique or temperature effects). Furthermore shelf life of the kit may affect enzymatic activity and thus colour intensity. Values measured are still valid.

12. LIMITATIONS

- Since exact conditions may vary from assay to assay, a standard curve must be established for every run.
- Bacterial or fungal contamination of either screen samples or reagents or cross-contamination between reagents may cause erroneous results.
- Disposable pipette tips, flasks or glassware are preferred, reusable glassware must be washed and thoroughly rinsed of all detergents before use.
- Improper or insufficient washing at any stage of the procedure will result in either false positive or false negative results. Empty wells completely before dispensing fresh wash solution, fill with Wash Buffer as indicated for each wash cycle and do not allow wells to sit uncovered or dry for extended periods.
- The use of radioimmunotherapy has significantly increased the number of patients with human anti-mouse IgG antibodies (HAMA). HAMA may interfere with assays utilizing murine monoclonal antibodies leading to both false positive and false negative results. Serum samples containing antibodies to murine immunoglobulins can still be analysed in such assays when murine immunoglobulins (serum, ascitic fluid, or monoclonal antibodies of irrelevant specificity) are added to the sample.

13. PERFORMANCE CHARACTERISTICS

13.1 Sensitivity

The limit of detection of human TGF- β 1 defined as the analyte concentration resulting in an absorbance significantly higher than that of the dilution medium (mean plus 2 standard deviations) was determined to be 9 pg/ml (mean of 6 independent assays).

13.2 Reproducibility

13.2.1 Intra-assay

Reproducibility within the assay was evaluated in 3 independent experiments. Each assay was carried out with 6 replicates of 8 serum samples containing different concentrations of human TGF- β 1. 2 standard curves were run on each plate. Data below show the mean human TGF- β 1 concentration and the coefficient of variation for each sample (see Table 3). The calculated overall intra-assay coefficient of variation was 5.1%.

Table 3

The mean human TGF- β 1 concentration and the coefficient of variation for each sample

Sample	Experiment	Mean Human TGF- β 1 Concentration (pg/ml)	Coefficient of Variation (%)
1	1	45241	3.3
	2	41522	4.5
	3	42803	3.5
2	1	24734	10.6
	2	22431	1.8
	3	22822	2.0
3	1	10409	6.2
	2	11284	1.9
	3	13140	5.6
4	1	6903	8.2
	2	5919	8.0
	3	5673	4.0
5	1	8313	7.2
	2	9488	2.5
	3	9559	2.6
6	1	12299	6.8
	2	10982	4.7
	3	13200	2.5
7	1	5861	10.6
	2	5427	4.5
	3	5984	2.0
8	1	2953	8.0
	2	3543	6.9
	3	3808	5.6

13.2.2 Inter-assay

Assay to assay reproducibility within one laboratory was evaluated in 3 independent experiments. Each assay was carried out with 6 replicates of 8 serum plasma samples containing different concentrations of human TGF- β 1. 2 standard curves were run on each plate. Data below show the mean human TGF- β 1 concentration and the coefficient of variation calculated on 18 determinations of each sample (see Table 4). The calculated overall inter-assay coefficient of variation was 8.4%.

Table 4

The mean human TGF- β 1 concentration and the coefficient of variation of each sample

Sample	Mean Human TGF- β 1 Concentration (pg/ml)	Coefficient of Variation (%)
1	43189	4.4
2	23329	5.3
3	11611	12.0
4	6165	10.6
5	9120	7.7
6	12160	9.2
7	5757	5.1
8	3435	12.7

13.3 Spiking Recovery

The spiking recovery was evaluated by spiking 2 levels of human TGF- β 1 into serum, plasma and cell culture supernatant. Recoveries were determined with 4 replicates each.

The amount of endogenous human TGF- β 1 in unspiked samples was subtracted from the spike values.

For recovery data see Table 5.

Table 5

Sample matrix*	Spike high (%)	Spike medium (%)
Serum	84	83
Plasma (EDTA)	74	110
Plasma (citrate)	91	97
Plasma (heparin)	97	90
Cell culture supernatant	124	141

*Due to high endogen TGF- β 1 levels data for low spikes are not indicated.

13.4 Dilution Linearity

Serum, plasma and cell culture supernatant samples with different levels of human TGF- β 1 were analysed at serial 2 fold dilutions with 4 replicates each. For recovery data see Table 6.

Table 6

Sample matrix	Recovery of Exp. Val.	
	Range (%)	Mean (%)
Serum	88 - 107	97
Plasma (EDTA)	108 - 150	125
Plasma (citrate)	109 - 151	132
Plasma (heparin)	52 - 81	67
Cell culture supernatant	90 - 115	106

13.5 Sample Stability

13.5.1 Freeze-Thaw Stability

Aliquots of serum samples (spiked or unspiked) were stored at -20°C and thawed 5 times, and the human TGF- β 1 levels determined. There was no significant loss of human TGF- β 1 immunoreactivity detected by freezing and thawing.

13.5.2 Storage Stability

Aliquots of serum samples (spiked or unspiked) were stored at -20°C, 2-8°C and room temperature (RT), and the human TGF- β 1 level determined after 24 h. There was no significant loss of human TGF- β 1 immunoreactivity detected during storage under above conditions.

13.6 Specificity

The assay detects both natural and recombinant human TGF- β 1.

The cross reactivity of TGF- β 2 and TGF- β 3, and of TNF- β , IL-8, IL-6, IL-2, TNF- α , IL-1 β , IL-4, IFN- γ , IL12p70, IL-5 and IL-10 was evaluated by spiking these proteins at physiologically relevant concentrations into serum. There was no cross reactivity detected.

13.7 Expected Values

A panel of samples from randomly selected apparently healthy donors (males and females) was tested for human TGF- β 1.

For detected human TGF- β 1 levels see Table 7.

Table 7

Sample Matrix	Number of Samples Evaluated	Range (pg/ml)	Mean (pg/ml)	Standard Deviation (pg/ml)
Serum	40	4639 - 14757	8225	2218
Plasma (EDTA)	40	2404 - 25558	6479	3778
Plasma (Citrate)	40	6518 - 31737	16942	5762
Plasma (Heparin)	40	3081 - 15879	6580	2882

14. REFERENCES

- 1) Astiz M, Saha D, Lustbader D, Lin R, Rackow E. Monocyte response to bacterial toxins, expression of cell surface receptors, and release of anti-inflammatory cytokines during sepsis. *J Lab Clin Med* 1996 Dec;128(6):594-600.
- 2) Bayer EM, Herr W, Kanzler S, Waldmann C, Meyer Zum Buschenfelde KH, Diense HP; Lohse AW. Transforming growth factor-beta1 in autoimmune hepatitis: correlation of liver tissue expression and serum levels with disease activity. *J Hepatol* 1998 May;28(5):803-811.
- 3) Border WA. Transforming growth factor-beta and the pathogenesis of glomerular diseases. *Curr Opin Nephrol Hypertens* 1994 Jan;3(1):54-58.
- 4) Carrieri PB, Provitera V, Bruno R, Perrella M, Tartaglia G, Busto A, Perrella O. Possible role of transforming growth factor-beta in relapsing-remitting multiple sclerosis. *Neurol Res* 1997 Dec;19(6):599-600.
- 5) Chambaz EM, Souhelitskiy S, Pellerin S, Defaye G, Cochet C, Feige JJ. Transforming growth factors-beta s: a multifunctional cytokine family. Implication in the regulation of adrenocortical cell endocrine functions. *Horm Res* 1996;45(3-5):222-226.
- 6) Chao CC, Janoff EN, Hu SX, Thomas K, Gallagher M, Tsang M, Peterson PK. Altered cytokine release in peripheral blood mononuclear cell cultures from patients with the chronic fatigue syndrome. *Cytokine* 1991 Jul;3(4):292-298.
- 7) Chopra V, Dinh TV, Hannigan EV. Circulating serum levels of cytokines and angiogenic factors in patients with cervical cancer. *Cancer Invest* 1998;16(3):152-159.
- 8) Del Giudice G, Crow MK. Role of transforming growth factor beta (TGF beta) in systemic autoimmunity. *Lupus* 1993 Aug;2(4):
- 9) Eder IE, Stenzl A, Hobisch A, Cronauer MV, Bartsch G, Klocker H. Expression of transforming growth factors beta-1, beta 2 and beta 3 in human bladder carcinomas. *Br J Cancer* 1997;75(12):
- 10) Finch CE, Laping NJ, Morgan TE, Nichols NR, Pasinetti GM. TGF-beta 1 is an organizer of responses to neurodegeneration. *J Cell Biochem* 1993 Dec;53(4):314-322.
- 11) Flanders KC, Ren RF, Lippa CF. Transforming growth factor-betas in neurodegenerative disease. *Prog Neurobiol* 1998 Jan;54(1):
- 12) Kim JS, Yoon SS, Kim YH, Ryu JS. Serial measurement of interleukin-6, transforming growth factor-beta, and S-100 protein in patients with acute stroke. *Stroke* 1996 Sep;27(9):1553-1557.
- 13) Kropf J, JO Schurek, A Wollner, and AM Gressner. Immunological measurement of transforming growth factor-beta I (TGF-b1) in blood; assay development and comparison. *Clinical Chemistry* 1997;43(10):1965-1974.
- 14) Langdahl BL, Knudsen JY, Jensen HK, Gregersen N, Eriksen EF. A sequence variation: 713-8delC in the transforming growth factor-beta 1 gene has higher prevalence in osteoporotic women than in normal women and is associated with very low bone mass in osteoporotic women and increased bone turnover in both osteoporotic and normal women. *Bone* 1997 Mar;20(3):289-294.
- 15) Lawrence DA. Transforming growth factor-beta: a general review. *Eur Cytokine Netw* 1996 Sep;7(3):363-374.

- 16) Martyre MC, Magdelenat H, Bryckaert MC, Laine-Bidron C, Calvo F. Increased intraplatelet of platelet-derived growth factor and transforming growth factor-beta in patients with myelofibrosis with myeloid metaplasia. *Br J Haematol* 1991 Jan;77(1):80-86.
- 17) Matsubara T, Umezawa Y, Tsuru S, Motohashi T, Yabuta K, Furukawa S. Decrease in the concentrations of transforming growth factor-beta 1 in the sera of patients with Kawasaki disease. *Scand J Rheumatol* 1997;26(4):314-317.
- 18) Mattson MP, Barger SW, Furukawa K, Bruce AJ, Wyss-Coray T, Mark RJ, Mucke L. Cellular signaling roles of TGF beta, TNF alpha and beta APP in brain injury responses and Alzheimer's disease. *Brain Res Rev* 1997 Feb;23(1-2):47-61.
- 19) Muller F, Aukrust P, Nilssen DE, Froland SS. Reduced serum level of transforming growth factor-beta in patients with IgA deficiency. *Clin Immunol Immunopathol* 1995 Aug;76(2):203-208.
- 20) Murawaki Y, Ikuta Y, Nishimura Y, Koda M, Kawasaki H. Serum markers for fibrosis and plasma transforming growth factor-beta 1 in patients with hepatocellular carcinoma in comparison with patients with liver cirrhosis. *J Gastroenterol Hepatol* 1996 May;11(5):443-450.
- 21) Ossege LM, Sindern E, Voss B, Malin JP. Corticosteroids induce expression of transforming-growth-factor-beta1 mRNA in peripheral blood mononuclear cells of patients with multiple sclerosis. *J Neuroimmunol* 1998 Apr 1;84(1):1-6.
- 22) Pfeilschifter J, Diel I, Scheppach B, Bretz A, Krempien R, Erdmann J, Schmid G, Reske N, Bismar H, Seck T, Krempien B, Ziegler R. Concentration of transforming growth factor beta in human bone tissue: relationship to age, menopause, bone turnover, and bone volume. *J Bone Miner Res* 1998 Apr;13(4):716-730.
- 23) Shankland SJ, Johnson RJ. TGF-beta in glomerular disease. *Miner Electrolyte Metab* 1998;24(2-3):168-173.
- 24) Sindern E, Schweppe K, Ossege LM, Malin JP. Potential role of transforming growth factor-beta 1 in terminating the immune response in patients with Guillain-Barre syndrome. *J Neurol* 1996 Mar;243(3):264-268.
- 25) Tornquist SJ, Oaks JL, Crawford TB. Elevation of cytokines associated with the thrombocytopenia of equine infectious anaemia. *J Gen Virol* 1997 Oct;78(Pt 10):2541-2548.
- 26) Wenisch C, Parschalk B, Burgmann H, Looareesuwan S, Graninger W. Decreased serum levels of TGF-beta in patients with acute *Plasmodium falciparum* malaria. *J Clin Immunol* 1995 Mar;15(2):69-73.
- 27) Wolff JM, Fandel T, Borchers H, Brehmer B Jr, Jakse G. Transforming growth factor-beta 1 serum concentration in patients with prostatic cancer and benign prostatic hyperplasia. *Br J Urol* 1998 Mar;81(3):403-405.
- 28) Yokoyama H, Deckert T. Central role of TGF-beta in the pathogenesis of diabetic nephropathy and macrovascular complications: a hypothesis. *Diabet Med* 1996 Apr;13(4):313-320.
- 29) Zauli G, Gugliotta L, Catani L, Vianelli N, Borgatti P, Belmonte MM, Tura S. Increased serum levels of transforming growth factor beta-1 in patients affected by thrombotic thrombocytopenic purpura (TTP): its implications on bone marrow haematopoiesis. *Br J Haematol* 1993 Jul;84(3):381-386.

15. REAGENT PREPARATION SUMMARY

15.1 Wash Buffer (1x)

Add **Wash Buffer Concentrate** 20x (50 ml) to 950 ml distilled water.

Number of Strips	Wash Buffer Concentrate (ml)	Distilled Water (ml)
1 - 6	25	475
1 - 12	50	950

15.2 Assay Buffer (1x)

Add **Assay Buffer Concentrate** 20x (5 ml) to 95 ml distilled water.

Number of Strips	Assay Buffer Concentrate (ml)	Distilled Water (ml)
1 - 6	2.5	47.5
1 - 12	5.0	95.0

15.3 Biotin-Conjugate

Make a 1:100 dilution of **Biotin-Conjugate** in Assay Buffer (1x):

Number of Strips	Biotin-Conjugate (ml)	Assay Buffer (1x) (ml)
1 - 6	0.06	5.94
1 - 12	0.12	11.88

15.4 Streptavidin-HRP

Make a 1:100 dilution of **Streptavidin-HRP** in Assay Buffer (1x):

Number of Strips	Streptavidin-HRP (ml)	Assay Buffer (1x) (ml)
1 - 6	0.06	5.94
1 - 12	0.12	11.88

15.5 Human TGF- β 1 Standard

Reconstitute lyophilized **human TGF- β 1 standard** with distilled water. (Reconstitution volume is stated in the Quality Control Sheet.).


16. TEST PROTOCOL SUMMARY

1. Pretreatment: 1:10 predilution (20 µl sample + 180 µl Assay Buffer (1x)), add 20 µl 1N HCl to 200 µl prediluted sample, mix and incubate for 1 hour at room temperature, add 20 µl 1N NaOH;
2. Determine the number of microwell strips required.
3. Wash microwell strips twice with Wash Buffer.
4. Standard dilution on the microwell plate: Add 100 µl Assay Buffer (1x), in duplicate, to all standard wells. Pipette 100 µl prepared standard into the first wells and create standard dilutions by transferring 100 µl from well to well. Discard 100 µl from the last wells. Alternatively external standard dilution in tubes (see 0): Pipette 100 µl of these standard dilutions in the microwell strips.
5. Add 100 µl Assay Buffer (1x), in duplicate, to the blank wells.
6. Add 60 µl Assay Buffer (1x) to sample wells.
7. Add 40 µl sample in duplicate, to designated sample wells.
8. Cover microwell strips and incubate 2 hours at room temperature (Shaking is absolutely necessary for an optimal test performance.)
9. Prepare Biotin-Conjugate.
10. Empty and wash microwell strips 5 times with Wash Buffer.
11. Add 100 µl Biotin-Conjugate to all wells.
12. Cover microwell strips and incubate 1 hours at room temperature. (Shaking is absolutely necessary for an optimal test performance.)
13. Prepare Streptavidin-HRP.
14. Empty and wash microwell strips 5 times with Wash Buffer.
15. Add 100 µl diluted Streptavidin-HRP to all wells.
16. Cover microwell strips and incubate 1 hour at room temperature. (Shaking is absolutely necessary for an optimal test performance.)
17. Empty and wash microwell strips 5 times with Wash Buffer.
18. Add 100 µl of TMB Substrate Solution to all wells.
19. Incubate the microwell strips for **about 30 minutes** at room temperature (18° to 25°C).
20. Add 100 µl Stop Solution to all wells.
21. Blank microwell reader and measure colour intensity at 450 nm.

Note: If instructions in this protocol have been followed samples have been diluted 1:30 (20 µl sample + 180 µl Assay Buffer (1x) + 20µl 1N HCl + 20µl 1N NaOH and 40 µl pretreated sample + 60 µl Assay Buffer (1x)), the concentration read from the standard curve must be multiplied by the dilution factor (x 30).

NOTES





HEADQUARTERS: BioVendor - Laboratorní medicína a.s.	Karasek 1767/1	621 00 Brno CZECH REPUBLIC	Phone: +420-549-124-185 Fax: +420-549-211-460	E-mail: info@biovendor.com Web: www.biovendor.com
EUROPEAN UNION: BioVendor GmbH	Im Neuenheimer Feld 583	D-69120 Heidelberg GERMANY	Phone: +49-6221-433-9100 Fax: +49-6221-433-9111	E-mail: infoEU@biovendor.com
USA, CANADA AND MEXICO: BioVendor LLC	1463 Sand Hill Road Suite 227	Candler, NC 28715 USA	Phone: +1-828-670-7807 +1-800-404-7807 Fax: +1-828-670-7809	E-mail: infoUSA@biovendor.com
CHINA - Hong Kong Office: BioVendor Laboratories Ltd	Room 4008 Hong Kong Plaza, No.188	Connaught Road West Hong Kong, CHINA	Phone: +852-2803-0523 Fax: +852-2803-0525	E-mail: infoHK@biovendor.com
CHINA – Mainland Office: BioVendor Laboratories Ltd	Room 2917, 29/F R & F Ying Feng Plaza, No.2 Huaqiang road	Pearl River New Town Guang Zhou, CHINA	Phone: +86-20-38065519 Fax: +86-20-38065529	E-mail: infoCN@biovendor.com