

Anti PACAP 27 (Human) Serum

Cat. No. Y050

Lot No. 080271013

Description: This antiserum was raised in a rabbit by immunization with a carrier free synthetic PACAP 27 (human) peptide. The product vial contains 50 μ L of the titled antiserum obtained by lyophilizing its 0.001 M phosphate buffer (pH 7.0, 0.5mL) solution. It can be used for immunoassay, immunohistochemistry or any other immunoreaction with PACAP27 (human).

Immunogen: Synthetic PACAP 27 (human) peptide, carrier free **Host:** Rabbit

Amino Acid Sequence of PACAP27 (human)¹⁾:

HSDGIFTDSY SRYRKQMAVK KYLA AVL-NH₂

Product Form: Lyophilized unpurified serum **Size:** 50 μ L

Reconstitution: Reconstitute the product with 0.5mL of 0.01M PBS (pH 7.0) to make a 10 fold diluted stock solution. If it is stored in a refrigerator, add moderate antiseptic to the solution (e.g. NaN₃ 0.1%).

Storage: The product will be stable for over one year if it be stored at -20°C to -80°C until opened. Upon reconstitution, the antiserum solution must be stored at 2°C to 8°C and used within one month. Repeated freezing-thawing should be avoided.

Suggested Working Dilution Range: 1:5,000-40,000 (final dilution ~1:280,000) for radioimmunoassay; 1: 1,000-4,000 for immunohistochemistry (frozen or paraffin sections). Optimal dilution should be determined by each laboratory for each application.

Specificity (based on radioimmunoassay): PACAP27 (human) 100%, PACAP38 (human) 0%, VIP (porcine) 0%, PHI (rat) 0%, PHI (human) 0%, secretin (human) 0%, helodermin 0%

Positive Control (immunohistochemistry): Rat colon

Species Tested: Rat

REFERENCES:

1) A. Miyata, A. Arimura et al., Isolation of a neuropeptide corresponding to the N-terminal 27 residues of the pituitary adenylate cyclase activating polypeptide with 38 residues (PACAP38). Biochemistry Biophysics Research Communications 170:643-648, 1990

FOR RESEARCH USE ONLY

<Distributed by>

SCETI K.K.

DF Kasumigaseki Place, 3-6-7 Kasumigaseki, Chiyoda-ku, Tokyo 100-0013 Japan

URL: <http://www.sceti.co.jp/export/> e-mail: exp-pet@sceti.co.jp

<Manufacturer>

Yanaihara Institute Inc.

2480-1 Awakura, Fujinomiya-shi, Shizuoka 418-0011 JAPAN